

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM043016\
 Data File : BM005185.D
 Acq On : 30 Apr 2016 22:01
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058

Quant Time: May 01 04:16:59 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 01 04:14:40 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	48270	20.00	ng/ul	0.00
18) Naphthalene-d8	10.57	136	227164	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.43	164	140520	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.17	188	311934	20.00	ng/ul	0.00
75) Chrysene-d12	21.36	240	268708	20.00	ng/ul	0.00
83) Perylene-d12	23.64	264	239257	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.29	96	7916	8.33	ng/uL	0.00
5) Phenol-d5	6.96	99	85416	21.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.12	67	50011	22.01	ng/ul	0.00
9) 2-Chlorophenol-d4	7.32	132	67785	21.46	ng/ul	0.00
13) 4-Methylphenol-d8	8.49	113	70879	20.94	ng/ul	0.00
19) Nitrobenzene-d5	8.95	128	33692	21.41	ng/ul	0.00
22) 2-Nitrophenol-d4	9.66	143	37789	21.53	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.20	165	71276	21.37	ng/ul	0.00
29) 4-Chloroaniline-d4	10.72	131	94718	24.23	ng/ul	0.00
43) Dimethylphthalate-d6	13.83	166	228715	20.61	ng/ul	0.00
46) Acenaphthylene-d8	14.12	160	276058	20.89	ng/ul	0.00
51) 4-Nitrophenol-d4	14.63	143	36463	18.80	ng/ul	0.00
57) Fluorene-d10	15.42	176	195671	20.06	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.54	200	33809	19.06	ng/ul	0.00
70) Anthracene-d10	17.27	188	291000	20.80	ng/ul	0.00
76) Pyrene-d10	19.57	212	292211	23.89	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.50	264	218440	20.38	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.32	88	14428	8.19	ng/uL	95
4) Benzaldehyde	6.93	77	51828	26.14	ng/ul	97
6) Phenol	6.98	94	89238	21.44	ng/ul	98
8) Bis(2-Chloroethyl)ether	7.21	93	69318	21.72	ng/ul	98
10) 2-Chlorophenol	7.35	128	68834	21.20	ng/ul	97
11) 2-Methylphenol	8.23	108	69190	21.30	ng/ul	98
12) 2,2'-oxybis(1-Chloropropan	8.32	45	92716	22.46	ng/ul	97
14) Acetophenone	8.60	105	111517	21.89	ng/ul	98
15) N-Nitroso-di-n-propylamine	8.59	70	56424	22.11	ng/ul	98
16) 4-Methylphenol	8.56	108	76803	21.24	ng/ul	95
17) Hexachloroethane	8.86	117	26850	20.77	ng/ul	93
20) Nitrobenzene	8.99	77	82559	21.33	ng/ul	99
21) Isophorone	9.50	82	161422	21.95	ng/ul	98
23) 2-Nitrophenol	9.69	139	40632	21.30	ng/ul	91
24) 2,4-Dimethylphenol	9.75	107	85215	20.74	ng/ul	99
25) Bis(2-Chloroethoxy)methane	9.99	93	98537	21.56	ng/ul	99
27) 2,4-Dichlorophenol	10.23	162	72041	21.02	ng/ul	99
28) Naphthalene	10.63	128	234618	20.48	ng/ul	100
30) 4-Chloroaniline	10.74	127	96405	24.23	ng/ul	100
31) Hexachlorobutadiene	10.90	225	44291	19.58	ng/ul	98
32) Caprolactam	11.50	113	16262	13.72	ng/ul	82
33) 4-Chloro-3-methylphenol	11.86	107	83028	21.18	ng/ul	98
34) 2-Methylnaphthalene	12.24	142	173925	20.51	ng/ul	100

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM043016\
 Data File : BM005185.D
 Acq On : 30 Apr 2016 22:01
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058

Quant Time: May 01 04:16:59 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 01 04:14:40 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	12.61	216	90111	20.53	ng/ul	99
37) Hexachlorocyclopentadiene	12.59	237	46141	17.71	ng/ul	95
38) 2,4,6-Trichlorophenol	12.85	196	59599	21.34	ng/ul	100
39) 2,4,5-Trichlorophenol	12.93	196	66301	21.20	ng/ul	98
40) 1,1'-Biphenyl	13.26	154	233276	20.99	ng/ul	99
41) 2-Chloronaphthalene	13.30	162	177390	20.98	ng/ul	100
42) 2-Nitroaniline	13.51	65	52394	22.85	ng/ul	98
44) Dimethylphthalate	13.88	163	226885	20.38	ng/ul	100
45) 2,6-Dinitrotoluene	14.01	165	46994	21.86	ng/ul	97
47) Acenaphthylene	14.15	152	291636	20.86	ng/ul	99
48) 3-Nitroaniline	14.34	138	49258	22.53	ng/ul	99
49) Acenaphthene	14.49	153	189469	20.59	ng/ul	99
50) 2,4-Dinitrophenol	14.55	184	18024	14.37	ng/ul	99
52) 4-Nitrophenol	14.65	109	30822	17.77	ng/ul	96
53) Dibenzofuran	14.83	168	275117	20.28	ng/ul	100
54) 2,4-Dinitrotoluene	14.80	165	67819	20.77	ng/ul	97
55) 2,3,4,6-Tetrachlorophenol	15.06	232	54147	19.94	ng/ul	99
56) Diethylphthalate	15.25	149	230298	20.50	ng/ul	100
58) Fluorene	15.48	166	213927	20.31	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.47	204	106879	20.11	ng/ul	98
60) 4-Nitroaniline	15.50	138	48016	20.36	ng/ul	98
63) 4,6-Dinitro-2-methylphenol	15.56	198	35857	19.19	ng/ul	95
64) N-Nitrosodiphenylamine	15.69	169	189767	22.13	ng/ul	99
65) 4-Bromophenyl-phenylether	16.36	248	66000	21.68	ng/ul	98
66) Hexachlorobenzene	16.48	284	71757	20.69	ng/ul	97
67) Atrazine	16.64	200	68294	21.62	ng/ul	98
68) Pentachlorophenol	16.83	266	37026	18.51	ng/ul	99
69) Phenanthrene	17.22	178	343713	20.70	ng/ul	100
71) Anthracene	17.31	178	346559	20.71	ng/ul	100
72) Carbazole	17.58	167	306717	21.37	ng/ul	100
73) Di-n-butylphthalate	18.14	149	364634	21.70	ng/ul	100
74) Fluoranthene	19.23	202	365119	19.98	ng/ul	100
77) Pyrene	19.60	202	368160	23.72	ng/ul	99
78) Butylbenzylphthalate	20.50	149	142499	23.87	ng/ul	97
79) 3,3'-Dichlorobenzidine	21.29	252	99632	20.73	ng/ul	98
80) Benzo(a)anthracene	21.35	228	317949	20.67	ng/ul	100
81) Bis(2-ethylhexyl)phthalate	21.27	149	192184	22.92	ng/ul	98
82) Chrysene	21.40	228	302773	20.57	ng/ul	100
84) Di-n-octyl phthalate	22.16	149	308962	20.77	ng/ul	98
85) Benzo(b)fluoranthene	22.96	252	293807	20.39	ng/ul	99
86) Benzo(k)fluoranthene	23.00	252	273997	19.55	ng/ul	100
88) Benzo(a)pyrene	23.54	252	272415	20.24	ng/ul	99
89) Indeno(1,2,3-cd)pyrene	25.95	276	290368	22.62	ng/ul	99
90) Dibenzo(a,h)anthracene	25.96	278	244988	22.84	ng/ul	98
91) Benzo(g,h,i)perylene	26.66	276	241460	22.99	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM043016\
Data File : BM005185.D
Acq On : 30 Apr 2016 22:01
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
Client Sampled :
SSTD02058

Quant Time: May 01 04:16:59 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 01 04:14:40 2016
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 8.19 ng/uL

RT: 3.32 min Scan# 108

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion: 88 Resp: 14428

Ion Ratio Lower Upper

88 100

43 26.8 23.0 34.4




58 74.1 55.4 83.2



Abundance Ion 88.00 (87.70 to 88.70): BM005185.D (-74) (-)

Ion 43.00 (42.70 to 43.70): BM005185.D (-74) (-)

Ion 58.00 (57.70 to 58.70): BM005185.D (-74) (-)



#4

Benzaldehyde

Concen: 26.14 ng/uL

RT: 6.93 min Scan# 722

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion: 77 Resp: 51828

Ion Ratio Lower Upper

77 100

105 97.9 81.3 121.9



106 94.9 77.2 115.8




Abundance Ion 77.00 (76.70 to 77.70): BM005185.D (-688) (-)

Ion 105.00 (104.70 to 105.70): BM005185.D (-688) (-)

Ion 106.00 (105.70 to 106.70): BM005185.D (-688) (-)





#6

Phenol

Concen: 21.44 ng/ul

RT: 6.98 min Scan# 730

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion: 94 Resp: 89238

Ion Ratio Lower Upper

94 100

65 27.3 21.6 32.4




66 38.9 29.6 44.4



Abundance Ion 94.00 (93.70 to 94.70): BM005185.D (-736) (-)

Ion 65.00 (64.70 to 65.70): BM005185.D (-736) (-)

Ion 66.00 (65.70 to 66.70): BM005185.D (-736) (-)



#8

Bis(2-Chloroethyl)ether

Concen: 21.72 ng/ul

RT: 7.21 min Scan# 769

Delta R.T. -0.01 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion: 93 Resp: 69318

Ion Ratio Lower Upper

93 100

63 76.0 59.3 88.9



95 32.1 25.4 38.0




Abundance Ion 93.00 (92.70 to 93.70): BM005185.D (-736) (-)

Ion 63.00 (62.70 to 63.70): BM005185.D (-736) (-)

Ion 95.00 (94.70 to 95.70): BM005185.D (-736) (-)









#10
2-Chlorophenol
Concen: 21.20 ng/ul
RT: 7.35 min Scan# 793
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Instrument :
BNA_M
ClientSampleId :
SSTD02058

Tgt Ion:128 Resp: 68834
Ion Ratio Lower Upper
128 100
64 48.0 35.7 53.5
130 32.6 25.9 38.9




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 130.00 (129.70 to 130.70): E






#11
2-Methylphenol
Concen: 21.30 ng/ul
RT: 8.23 min Scan# 942
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:108 Resp: 69190
Ion Ratio Lower Upper
108 100
107 88.4 69.4 104.0



Abundance Ion 108.00 (107.70 to 108.70): E
Ion 107.00 (106.70 to 107.70): E









#12
 2,2'-oxybis(1-chloropropane)
 Concen: 22.46 ng/ul
 RT: 8.32 min Scan# 957
 Delta R.T. -0.01 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion: 45 Resp: 92716
 Ion Ratio Lower Upper
 45 100
 77 14.4 12.6 18.8
 79 11.6 9.9 14.9



#14
 Acetophenone
 Concen: 21.89 ng/ul
 RT: 8.60 min Scan# 1006
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion: 105 Resp: 111517
 Ion Ratio Lower Upper
 105 100
 77 77.7 61.4 92.0
 51 26.4 20.0 30.0









#15
N-Nitroso-di-n-propylamine
Concen: 22.11 ng/ul
RT: 8.59 min Scan# 1004
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion: 70 Resp: 56424
Ion Ratio Lower Upper
70 100
42 51.8 40.6 60.8
101 10.2 8.2 12.4
130 21.6 18.5 27.7



#16
4-Methylphenol
Concen: 21.24 ng/ul
RT: 8.56 min Scan# 998
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion: 108 Resp: 76803
Ion Ratio Lower Upper
108 100
107 108.0 91.0 136.6





#17

Hexachloroethane

Concen: 20.77 ng/ul

RT: 8.86 min Scan# 1049

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:117 Resp: 26850

Ion Ratio Lower Upper




117 100

201 103.1 88.5 132.7

199 64.4 55.4 83.2



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 201.00 (200.70 to 201.70): E
Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 21.33 ng/ul

RT: 8.99 min Scan# 1071

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion: 77 Resp: 82559

Ion Ratio Lower Upper



77 100


123 46.0 37.7 56.5

65 13.7 10.9 16.3



Abundance Ion 77.00 (76.70 to 77.70): BM
Ion 123.00 (122.70 to 123.70): E
Ion 65.00 (64.70 to 65.70): BM





#21

Isophorone

Concen: 21.95 ng/ul

RT: 9.50 min Scan# 1159

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampled :

SSTD02058


Tgt Ion: 82 Resp: 161422

Ion Ratio Lower Upper

82 100

95 7.0 6.0 9.0




138 17.4 14.6 22.0



Abundance Ion 82.00 (81.70 to 82.70): BM005185.D (-1125) (-)

Ion 95.00 (94.70 to 95.70): BM005185.D (-1125) (-)

Ion 138.00 (137.70 to 138.70): BM005185.D (-1125) (-)



#23

2-Nitrophenol

Concen: 21.30 ng/ul

RT: 9.69 min Scan# 1191

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion: 139 Resp: 40632

Ion Ratio Lower Upper

139 100

65 53.8 36.2 54.2



109 32.8 24.8 37.2




Abundance Ion 139.00 (138.70 to 139.70): BM005185.D (-1157) (-)

Ion 65.00 (64.70 to 65.70): BM005185.D (-1157) (-)

Ion 109.00 (108.70 to 109.70): BM005185.D (-1157) (-)









#24
 2,4-Dimethylphenol
 Concen: 20.74 ng/ul
 RT: 9.75 min Scan# 1201
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion: 107 Resp: 85215
 Ion Ratio Lower Upper
 107 100
 121 49.1 39.8 59.8
 122 84.0 68.5 102.7



#25
 Bis(2-Chloroethoxy)methane
 Concen: 21.56 ng/ul
 RT: 9.99 min Scan# 1241
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion: 93 Resp: 98537
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.2 39.2
 123 11.0 9.4 14.0









#27
 2,4-Dichlorophenol
 Concen: 21.02 ng/ul
 RT: 10.23 min Scan# 1282
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion:162 Resp: 72041
 Ion Ratio Lower Upper
 162 100
 164 65.0 52.9 79.3
 98 36.3 28.3 42.5



#28
 Naphthalene
 Concen: 20.48 ng/ul
 RT: 10.63 min Scan# 1350
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion:128 Resp: 234618
 Ion Ratio Lower Upper
 128 100
 129 10.9 8.7 13.1
 127 12.9 10.6 15.8









#30
4-Chloroaniline
Concen: 24.23 ng/ul
RT: 10.74 min Scan# 1369
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion:127 Resp: 96405
Ion Ratio Lower Upper
127 100
129 32.6 26.0 39.0



#31
Hexachlorobutadiene
Concen: 19.58 ng/ul
RT: 10.90 min Scan# 1397
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:225 Resp: 44291
Ion Ratio Lower Upper
225 100
223 62.4 51.9 77.9
227 62.8 51.0 76.6









#32
 Caprolactam
 Concen: 13.72 ng/ul
 RT: 11.50 min Scan# 1498
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion:113 Resp: 16262
 Ion Ratio Lower Upper
 113 100
 55 168.1 117.0 175.4
 56 129.9 87.7 131.5



#33
 4-Chloro-3-methylphenol
 Concen: 21.18 ng/ul
 RT: 11.86 min Scan# 1560
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion:107 Resp: 83028
 Ion Ratio Lower Upper
 107 100
 144 24.7 19.7 29.5
 142 76.0 62.5 93.7









#34
 2-Methylnaphthalene
 Concen: 20.51 ng/ul
 RT: 12.24 min Scan# 1624
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion:142 Resp: 173925
 Ion Ratio Lower Upper
 142 100
 141 88.1 70.3 105.5



#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 20.53 ng/ul
 RT: 12.61 min Scan# 1687
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion:216 Resp: 90111
 Ion Ratio Lower Upper
 216 100
 214 78.1 61.9 92.9
 179 21.4 16.9 25.3
 108 18.0 13.2 19.8





#37

Hexachlorocyclopentadiene

Concen: 17.71 ng/ul

RT: 12.59 min Scan# 1683

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:237 Resp: 46141

Ion Ratio Lower Upper

237 100

235 61.8 52.7 79.1




272 12.2 10.2 15.4



Abundance Ion 237.00 (236.70 to 237.70): E

Ion 235.00 (234.70 to 235.70): E

Ion 272.00 (271.70 to 272.70): E



#38

2,4,6-Trichlorophenol

Concen: 21.34 ng/ul

RT: 12.85 min Scan# 1728

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:196 Resp: 59599

Ion Ratio Lower Upper

196 100

198 95.5 76.1 114.1



200 30.8 24.3 36.5




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#39
 2,4,5-Trichlorophenol
 Concen: 21.20 ng/ul
 RT: 12.93 min Scan# 1741
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion:196 Resp: 66301
 Ion Ratio Lower Upper
 196 100
 198 94.0 76.7 115.1
 200 30.4 25.4 38.0



#40
 1,1'-Biphenyl
 Concen: 20.99 ng/ul
 RT: 13.26 min Scan# 1797
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion:154 Resp: 233276
 Ion Ratio Lower Upper
 154 100
 153 40.2 31.5 47.3
 76 12.6 10.1 15.1









#41
2-Chloronaphthalene
Concen: 20.98 ng/ul
RT: 13.30 min Scan# 1804
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Instrument :
BNA_M
ClientSampleId :
SSTD02058

Tgt Ion: 162 Resp: 177390
Ion Ratio Lower Upper
162 100
164 32.5 25.8 38.8
127 40.4 32.3 48.5




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 127.00 (126.70 to 127.70): E






#42
2-Nitroaniline
Concen: 22.85 ng/ul
RT: 13.51 min Scan# 1840
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion: 65 Resp: 52394
Ion Ratio Lower Upper
65 100
92 71.1 56.9 85.3
138 108.0 89.3 133.9



Abundance Ion 65.00 (64.70 to 65.70): BM0
Ion 92.00 (91.70 to 92.70): BM0
Ion 138.00 (137.70 to 138.70): E









#44
Dimethylphthalate
Concen: 20.38 ng/ul
RT: 13.88 min Scan# 1903
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion:163 Resp: 226885
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.2 12.2



#45
2,6-Dinitrotoluene
Concen: 21.86 ng/ul
RT: 14.01 min Scan# 1925
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:165 Resp: 46994
Ion Ratio Lower Upper
165 100
89 56.4 43.1 64.7
121 20.5 16.0 24.0





#47

Acenaphthylene

Concen: 20.86 ng/ul

RT: 14.15 min Scan# 1948

Delta R.T. -0.01 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01





Instrument :

BNA_M

ClientSampleId :

SSTD02058

Tgt Ion:	152	Resp:	291636
Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.8	23.8
153	13.5	10.3	15.5



#48

3-Nitroaniline

Concen: 22.53 ng/ul




RT: 14.34 min Scan# 1981


Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Tgt Ion:	138	Resp:	49258
Ion	Ratio	Lower	Upper
138	100		
108	11.2	8.8	13.2
92	116.9	94.9	142.3





#49

Acenaphthene

Concen: 20.59 ng/ul

RT: 14.49 min Scan# 2007

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:153 Resp: 189469

Ion Ratio Lower Upper

153 100

152 47.6 37.7 56.5




154 88.9 71.4 107.2



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#50

2,4-Dinitrophenol

Concen: 14.37 ng/ul

RT: 14.55 min Scan# 2016

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:184 Resp: 18024

Ion Ratio Lower Upper

184 100

63 66.1 51.8 77.6



154 64.8 51.7 77.5




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 154.00 (153.70 to 154.70): E





#52

4-Nitrophenol

Concen: 17.77 ng/ul

RT: 14.65 min Scan# 2034

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:109 Resp: 30822

Ion Ratio Lower Upper

109 100

139 128.4 104.5 156.7




65 126.7 96.5 144.7



Abundance Ion 109.00 (108.70 to 109.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 65.00 (64.70 to 65.70): BM



#53

Dibenzofuran

Concen: 20.28 ng/ul

RT: 14.83 min Scan# 2064

Delta R.T. 0.00 min

Lab File: BM005185.D


Acq: 30 Apr 2016 22:01

Tgt Ion:168 Resp: 275117

Ion Ratio Lower Upper



168 100


139 37.7 29.9 44.9



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E









#54
 2,4-Dinitrotoluene
 Concen: 20.77 ng/ul
 RT: 14.80 min Scan# 2059
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02058


Tgt Ion	Ratio	Lower	Upper
165	100		
63	42.2	32.1	48.1
182	2.8	2.2	3.4



#55
 2,3,4,6-Tetrachlorophenol
 Concen: 19.94 ng/ul
 RT: 15.06 min Scan# 2103
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion	Ratio	Lower	Upper
232	100		
131	44.7	35.4	53.2
130	2.4	2.1	3.1
166	30.8	25.6	38.4





#56

Diethylphthalate

Concen: 20.50 ng/ul

RT: 15.25 min Scan# 2136

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:149 Resp: 230298

Ion Ratio Lower Upper

149 100

177 22.3 17.9 26.9




150 12.3 9.8 14.6



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#58

Fluorene

Concen: 20.31 ng/ul

RT: 15.48 min Scan# 2175

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:166 Resp: 213927

Ion Ratio Lower Upper

166 100

165 98.3 77.7 116.5



167 13.3 10.9 16.3




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
4-Chlorophenyl-phenylether
Concen: 20.11 ng/ul
RT: 15.47 min Scan# 2173
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion:204 Resp: 106879
Ion Ratio Lower Upper
204 100
206 33.2 26.2 39.4
141 61.4 47.2 70.8



#60
4-Nitroaniline
Concen: 20.36 ng/ul
RT: 15.50 min Scan# 2179
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:138 Resp: 48016
Ion Ratio Lower Upper
138 100
92 53.8 41.0 61.6
108 84.3 66.8 100.2









#63
 4,6-Dinitro-2-methylphenol
 Concen: 19.19 ng/ul
 RT: 15.56 min Scan# 2189
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Instrument :
 BNA_M
 ClientSampleId :
 SST02058

Tgt Ion	Ratio	Lower	Upper
198	100		
182	3.9	3.9	5.9
77	24.0	21.4	32.2




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): BM






#64
 N-Nitrosodiphenylamine
 Concen: 22.13 ng/ul
 RT: 15.69 min Scan# 2210
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion	Ratio	Lower	Upper
169	100		
168	65.6	51.8	77.6
167	35.7	28.2	42.2



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E









#65
4-Bromophenyl-phenylether
Concen: 21.68 ng/ul
RT: 16.36 min Scan# 2325
Delta R.T. -0.01 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion:248 Resp: 66000
Ion Ratio Lower Upper
248 100
250 97.0 77.8 116.6
141 82.6 62.7 94.1



#66
Hexachlorobenzene
Concen: 20.69 ng/ul
RT: 16.48 min Scan# 2345
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:284 Resp: 71757
Ion Ratio Lower Upper
284 100
142 39.2 29.8 44.8
249 33.6 25.9 38.9









#67
 Atrazine
 Concen: 21.62 ng/ul
 RT: 16.64 min Scan# 2372
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SST02058


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.8	21.0	31.4
215	47.6	39.4	59.2



#68
 Pentachlorophenol
 Concen: 18.51 ng/ul
 RT: 16.83 min Scan# 2404
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.6	49.8	74.6
268	63.1	49.6	74.4





#69

Phenanthrene

Concen: 20.70 ng/ul

RT: 17.22 min Scan# 2470

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:178 Resp: 343713

Ion Ratio Lower Upper

178 100

179 15.2 11.8 17.8




176 19.0 15.2 22.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 20.71 ng/ul

RT: 17.31 min Scan# 2486

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:178 Resp: 346559

Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1



176 18.4 14.6 22.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 21.37 ng/ul

RT: 17.58 min Scan# 2532

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01





Instrument :

BNA_M

ClientSampleId :

SSTD02058

Tgt Ion:	167	Resp:	306717
Ion Ratio	Lower	Upper	
167	100		
166	20.9	16.6	25.0
139	12.9	10.2	15.4



#73

Di-n-butylphthalate

Concen: 21.70 ng/ul




RT: 18.14 min Scan# 2627


Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Tgt Ion:	149	Resp:	364634
Ion Ratio	Lower	Upper	
149	100		
150	9.1	7.3	10.9
104	4.9	3.9	5.9









#74
 Fluoranthene
 Concen: 19.98 ng/ul
 RT: 19.23 min Scan# 2813
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01




Instrument :
 BNA_M
 ClientSampleId :
 SST02058


Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.3	9.0	13.4
100	8.5	6.6	10.0



#77
 Pyrene
 Concen: 23.72 ng/ul
 RT: 19.60 min Scan# 2875
 Delta R.T. 0.00 min
 Lab File: BM005185.D
 Acq: 30 Apr 2016 22:01

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.4	10.5	15.7
100	10.8	8.2	12.2









#78
Butylbenzylphthalate
Concen: 23.87 ng/ul
RT: 20.50 min Scan# 3028
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion:149 Resp: 142499
Ion Ratio Lower Upper
149 100
91 69.5 57.0 85.6
206 22.5 16.6 24.8



#79
3,3'-Dichlorobenzidine
Concen: 20.73 ng/ul
RT: 21.29 min Scan# 3162
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion:252 Resp: 99632
Ion Ratio Lower Upper
252 100
254 64.8 50.4 75.6
126 10.9 8.7 13.1





#80

Benzo(a)anthracene

Concen: 20.67 ng/ul

RT: 21.35 min Scan# 3173

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:228 Resp: 317949

Ion Ratio Lower Upper

228 100

229 19.6 15.8 23.6




226 26.9 21.5 32.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#81

Bis(2-ethylhexyl)phthalate

Concen: 22.92 ng/ul

RT: 21.27 min Scan# 3159

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:149 Resp: 192184

Ion Ratio Lower Upper

149 100

167 27.5 22.7 34.1



279 3.8 3.4 5.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 20.57 ng/ul

RT: 21.40 min Scan# 3182

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Instrument :

BNA_M

ClientSampleId :

SSTD02058


Tgt Ion:228 Resp: 302773

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.2




229 19.6 15.5 23.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 20.77 ng/ul

RT: 22.16 min Scan# 3310

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01


Tgt Ion:149 Resp: 308962

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9



43 8.8 6.4 9.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BM









#85
Benzo(b)fluoranthene
Concen: 20.39 ng/ul
RT: 22.96 min Scan# 3446
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01




Instrument :
BNA_M
ClientSampleId :
SSTD02058


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.4	26.0
125	10.2	7.8	11.6



#86
Benzo(k)fluoranthene
Concen: 19.55 ng/ul
RT: 23.00 min Scan# 3454
Delta R.T. 0.00 min
Lab File: BM005185.D
Acq: 30 Apr 2016 22:01

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.5	26.3
125	10.0	7.9	11.9





#88

Benzo(a)pyrene

Concen: 20.24 ng/ul

RT: 23.54 min Scan# 3546

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01





Instrument :

BNA_M

ClientSampleId :

SSTD02058

Tgt Ion:	252	Resp:	272415
Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.5	26.3
125	10.9	8.3	12.5



#89

Indeno(1,2,3-cd)pyrene

Concen: 22.62 ng/ul

RT: 25.95 min Scan# 3955

Delta R.T. 0.00 min

Lab File: BM005185.D

Acq: 30 Apr 2016 22:01

Tgt Ion:	276	Resp:	290368
Ion	Ratio	Lower	Upper
276	100		
138	26.5	20.2	30.4
277	25.5	20.4	30.6

