

Data Path : Z:\HPCHEM1\BNA M\DATA\BM043016\
 Data File : BM005157.D
 Acq On : 30 Apr 2016 00:33
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Apr 30 04:37:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 30 03:07:27 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.730	0.776	-6.3	126	0.00
3 S	1,4-Dioxane-d8	0.394	0.416	-5.6	124	0.00
4	Benzaldehyde	0.821	1.082	-31.8#	126	0.00
5 S	Phenol-d5	1.654	1.786	-8.0	126	0.00
6	Phenol	1.725	1.885	-9.3	128	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.941	1.047	-11.3	129	0.00
8	Bis(2-Chloroethyl)ether	1.322	1.450	-9.7	127	0.00
9 S	2-Chlorophenol-d4	1.309	1.390	-6.2	124	0.00
10	2-Chlorophenol	1.345	1.447	-7.6	127	0.00
11	2-Methylphenol	1.346	1.436	-6.7	126	0.00
12	2,2'-oxybis(1-Chloropropane	1.710	1.918	-12.2	130	0.00
13 S	4-Methylphenol-d8	1.403	1.481	-5.6	126	0.00
14	Acetophenone	2.111	2.330	-10.4	124	0.00
15 P	N-Nitroso-di-n-propylamine	1.057	1.196	-13.2	129	0.00
16	4-Methylphenol	1.498	1.622	-8.3	126	0.00
17	Hexachloroethane	0.536	0.561	-4.7	124	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
19 S	Nitrobenzene-d5	0.139	0.149	-7.2	125	0.00
20	Nitrobenzene	0.341	0.365	-7.0	125	0.00
21	Isophorone	0.647	0.706	-9.1	126	0.00
22 S	2-Nitrophenol-d4	0.155	0.167	-7.7	126	0.00
23 C	2-Nitrophenol	0.168	0.182	-8.3	124	0.00
24	2,4-Dimethylphenol	0.362	0.377	-4.1	122	0.00
25	Bis(2-Chloroethoxy)methane	0.402	0.437	-8.7	128	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.311	-5.8	123	0.00
27 C	2,4-Dichlorophenol	0.302	0.316	-4.6	121	0.00
28	Naphthalene	1.009	1.033	-2.4	121	0.00
29 S	4-Chloroaniline-d4	0.344	0.417	-21.2#	122	0.00
30	4-Chloroaniline	0.350	0.424	-21.1#	122	0.00
31 C	Hexachlorobutadiene	0.199	0.191	4.0	115	0.00
32	Caprolactam	0.104	0.105	-1.0	121	0.00
33 C	4-Chloro-3-methylphenol	0.345	0.369	-7.0	123	0.00
34	2-Methylnaphthalene	0.747	0.766	-2.5	120	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
36	1,2,4,5-Tetrachlorobenzene	0.625	0.640	-2.4	114	0.00
37	Hexachlorocyclopentadiene	0.371	0.346	6.7	111	0.00
38 C	2,4,6-Trichlorophenol	0.397	0.424	-6.8	118	0.00
39	2,4,5-Trichlorophenol	0.445	0.471	-5.8	116	0.00
40	1,1'-Biphenyl	1.582	1.662	-5.1	116	0.00
41	2-Chloronaphthalene	1.203	1.263	-5.0	117	0.00
42	2-Nitroaniline	0.326	0.374	-14.7	121	0.00
43 S	Dimethylphthalate-d6	1.580	1.637	-3.6	114	0.00
44	Dimethylphthalate	1.584	1.636	-3.3	114	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM043016\
 Data File : BM005157.D
 Acq On : 30 Apr 2016 00:33
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Apr 30 04:37:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 30 03:07:27 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.306	0.332	-8.5	118	0.00
46 S	Acenaphthylene-d8	1.881	1.969	-4.7	114	0.00
47	Acenaphthylene	1.990	2.075	-4.3	114	0.00
48	3-Nitroaniline	0.311	0.351	-12.9	117	0.00
49 C	Acenaphthene	1.309	1.360	-3.9	114	0.00
50	2,4-Dinitrophenol	0.179	0.139	22.3#	104	0.00
51 S	4-Nitrophenol-d4	0.276	0.276	0.0	112	0.00
52	4-Nitrophenol	0.247	0.231	6.5	107	0.00
53	Dibenzofuran	1.931	1.981	-2.6	112	0.00
54	2,4-Dinitrotoluene	0.465	0.488	-4.9	113	0.00
55	2,3,4,6-Tetrachlorophenol	0.387	0.395	-2.1	111	0.00
56	Diethylphthalate	1.599	1.649	-3.1	113	0.00
57 S	Fluorene-d10	1.388	1.408	-1.4	112	0.00
58	Fluorene	1.499	1.544	-3.0	111	0.00
59	4-Chlorophenyl-phenylether	0.756	0.773	-2.2	111	0.00
60	4-Nitroaniline	0.336	0.353	-5.1	114	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.114	0.112	1.8	105	0.00
63	4,6-Dinitro-2-methylphenol	0.120	0.117	2.5	102	0.00
64	N-Nitrosodiphenylamine	0.550	0.600	-9.1	112	0.00
65	4-Bromophenyl-phenylether	0.195	0.206	-5.6	110	0.00
66	Hexachlorobenzene	0.222	0.228	-2.7	107	0.00
67	Atrazine	0.203	0.223	-9.9	111	0.00
68 C	Pentachlorophenol	0.128	0.122	4.7	105	0.00
69	Phenanthrene	1.064	1.110	-4.3	109	0.00
70 S	Anthracene-d10	0.897	0.939	-4.7	107	0.00
71	Anthracene	1.073	1.124	-4.8	107	0.00
72	Carbazole	0.920	1.019	-10.8	108	0.00
73	Di-n-butylphthalate	1.077	1.217	-13.0	114	0.00
74 C	Fluoranthene	1.171	1.265	-8.0	103	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
76 S	Pyrene-d10	0.910	1.043	-14.6	101	0.00
77	Pyrene	1.155	1.318	-14.1	100	0.00
78	Butylbenzylphthalate	0.444	0.548	-23.4#	106	0.00
79	3,3'-Dichlorobenzidine	0.358	0.387	-8.1	86	0.00
80	Benzo(a)anthracene	1.145	1.207	-5.4	91	0.00
81	Bis(2-ethylhexyl)phthalate	0.624	0.752	-20.5#	102	0.00
82	Chrysene	1.095	1.137	-3.8	91	0.00
83 I	Perylene-d12	1.000	1.000	0.0	78	0.00
84	Di-n-octyl phthalate	1.243	1.454	-17.0	91	0.00
85	Benzo(b)fluoranthene	1.205	1.261	-4.6	81	0.00
86	Benzo(k)fluoranthene	1.171	1.202	-2.6	78	0.00
87 S	Benzo(a)pyrene-d12	0.896	0.926	-3.3	78	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM043016\
Data File : BM005157.D
Acq On : 30 Apr 2016 00:33
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02055

Quant Time: Apr 30 04:37:58 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 30 03:07:27 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.125	1.169	-3.9	79	0.00
89	Indeno(1,2,3-cd)pyrene	1.073	1.173	-9.3	80	0.00
90	Dibenzo(a,h)anthracene	0.897	0.984	-9.7	80	0.00
91	Benzo(a,h,i)perylene	0.878	0.966	-10.0	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0