

Data File : BG026798.D
Acq On : 1 May 2017 7:23
Operator : SJ/MA
Sample : I2846-12
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
JHSR5

Quant Time: May 01 08:11:41 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon May 01 06:49:04 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	203818	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	909019	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	481519	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.37	188	841612	20.00	ng/ul	0.01
75) Chrysene-d12	21.61	240	749040	20.00	ng/ul	0.01
83) Perylene-d12	24.79	264	809644	20.00	ng/ul	0.03

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.43	96	15847	3.50	ng/uL	0.00
5) Phenol-d5	7.18	99	351443	19.32	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.33	67	226860	21.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	304997	19.75	ng/ul	0.00
13) 4-Methylphenol-d8	8.72	113	290775	19.26	ng/ul	0.00
19) Nitrobenzene-d5	9.17	128	185150	24.37	ng/ul	0.00
22) 2-Nitrophenol-d4	9.89	143	185453	22.44	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.44	165	278755	21.23	ng/ul	0.00
29) 4-Chloroaniline-d4	10.93	131	736156	41.88	ng/ul	-0.02
43) Dimethylphthalate-d6	14.02	166	827674	21.36	ng/ul	0.00
46) Acenaphthylene-d8	14.32	160	1122428	23.06	ng/ul	0.00
51) 4-Nitrophenol-d4	14.85	143	162372	23.24	ng/ul	0.01
57) Fluorene-d10	15.61	176	721787	21.34	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.73	200	119543	23.39	ng/ul	0.00
70) Anthracene-d10	17.46	188	968182	23.70	ng/ul	0.00
76) Pyrene-d10	19.75	212	889170	22.49	ng/ul	0.01
87) Benzo(a)pyrene-d12	24.56	264	882315	23.05	ng/ul	0.02

Target Compounds

					Qvalue
4) Benzaldehyde	7.15	77	73398	8.35 ng/ul#	1
6) Phenol	7.21	94	22578	1.22 ng/ul#	53
14) Acetophenone	8.81	105	495814	22.36 ng/ul#	22
15) N-Nitroso-di-n-propylamine	8.78	70	28197	2.58 ng/ul#	89
16) 4-Methylphenol	8.79	108	28347	1.78 ng/ul	96
17) Hexachloroethane	9.12	117	243059	42.53 ng/ul#	1
20) Nitrobenzene	9.20	77	33352	2.27 ng/ul#	18
21) Isophorone	9.73	82	87771	3.12 ng/ul#	1
23) 2-Nitrophenol	9.93	139	10252	1.18 ng/ul#	1
24) 2,4-Dimethylphenol	10.02	107	24242	1.53 ng/ul#	47
28) Naphthalene	10.86	128	226923	4.80 ng/ul#	79
32) Caprolactam	11.72	113	26876	5.49 ng/ul#	1
34) 2-Methylnaphthalene	12.46	142	668460	19.04 ng/ul	94
40) 1,1'-Biphenyl	13.46	154	136591	3.37 ng/ul#	97
44) Dimethylphthalate	14.07	163	215754	5.69 ng/ul#	92
45) 2,6-Dinitrotoluene	14.17	165	14974	1.87 ng/ul#	32
48) 3-Nitroaniline	14.53	138	11180	1.33 ng/ul#	27
49) Acenaphthene	14.70	153	57625	1.67 ng/ul#	48
53) Dibenzofuran	15.02	168	81555	1.78 ng/ul#	56
58) Fluorene	15.67	166	120250	3.21 ng/ul#	70
64) N-Nitrosodiphenylamine	15.87	169	231409	8.37 ng/ul#	48
67) Atrazine	16.83	200	12163	1.41 ng/ul#	9
69) Phenanthrene	17.41	178	343586	7.25 ng/ul#	87

Data File : BG026798.D
Acq On : 1 May 2017 7:23
Operator : SJ/MA
Sample : I2846-12
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
JHSR5


Quant Time: May 01 08:11:41 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon May 01 06:49:04 2017
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
73) Di-n-butylphthalate	18.31	149	69696	1.30	ng/ul#	68
74) Fluoranthene	19.41	202	91862	2.09	ng/ul#	68
77) Pyrene	19.77	202	182479	3.63	ng/ul#	73
78) Butylbenzylphthalate	20.64	149	59272	2.54	ng/ul#	88
81) Bis(2-ethylhexyl)phthalate	21.48	149	798015	23.96	ng/ul#	97
82) Chrysene	21.65	228	56573	1.39	ng/ul#	72
85) Benzo(b)fluoranthene	23.78	252	50053	1.08	ng/ul#	1
86) Benzo(k)fluoranthene	23.78	252	50053	1.10	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ALS Vial : 35 Sample Multiplier: 1

Response via : Initial Calibration





#4

Benzaldehyde

Concen: 8.35 ng/ul

RT: 7.15 min Scan# 692

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampled :

JHSR5


Tgt Ion: 77 Resp: 73398

Ion Ratio Lower Upper

77 100

105 954.2 62.0 93.0#




106 89.3 60.2 90.4



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#6

Phenol

Concen: 1.22 ng/ul

RT: 7.21 min Scan# 702

Delta R.T. 0.00 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion: 94 Resp: 22578

Ion Ratio Lower Upper

94 100

65 87.8 26.8 40.2#



66 42.1 44.4 66.6#




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC





#14

Acetophenone

Concen: 22.36 ng/ul

RT: 8.81 min Scan# 974

Delta R.T. -0.02 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion: 105 Resp: 495814

Ion Ratio Lower Upper

105 100

77 12.0 76.0 114.0#




51 3.7 34.8 52.2#



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#15

N-Nitroso-di-n-propylamine

Concen: 2.58 ng/ul

RT: 8.78 min Scan# 968

Delta R.T. -0.03 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Tgt Ion: 70 Resp: 28197


Ion Ratio Lower Upper

70 100

42 75.6 59.0 88.6

101 2.0 6.6 9.8#

130 0.0 15.6 23.4#






Abundance Ion 70.00 (69.70 to 70.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E









#16
4-Methylphenol
Concen: 1.78 ng/ul
RT: 8.79 min Scan# 970
Delta R.T. 0.00 min
Lab File: BG026798.D
Acq: 1 May 2017 7:23




Instrument :
BNA_M
ClientSampleId :
JHSR5


Tgt Ion:108 Resp: 28347
Ion Ratio Lower Upper
108 100
107 110.7 91.7 137.5



#17
Hexachloroethane
Concen: 42.53 ng/ul
RT: 9.12 min Scan# 1026
Delta R.T. 0.02 min
Lab File: BG026798.D
Acq: 1 May 2017 7:23

Tgt Ion:117 Resp: 243059
Ion Ratio Lower Upper
117 100
201 0.0 91.9 137.9#
199 0.0 64.3 96.5#









#20
 Nitrobenzene
 Concen: 2.27 ng/ul
 RT: 9.20 min Scan# 1040
 Delta R.T. -0.01 min
 Lab File: BG026798.D
 Acq: 1 May 2017 7:23




Instrument :
 BNA_M
 ClientSampleId :
 JHSR5


Tgt Ion: 77 Resp: 33352
 Ion Ratio Lower Upper
 77 100
 123 16.9 27.4 41.0#
 65 98.3 13.3 19.9#



#21
 Isophorone
 Concen: 3.12 ng/ul
 RT: 9.73 min Scan# 1131
 Delta R.T. 0.01 min
 Lab File: BG026798.D
 Acq: 1 May 2017 7:23

Tgt Ion: 82 Resp: 87771
 Ion Ratio Lower Upper
 82 100
 95 174.7 7.8 11.6#
 138 11.4 12.2 18.2#





#23

2-Nitrophenol

Concen: 1.18 ng/ul

RT: 9.93 min Scan# 1164

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5





Tgt Ion:139 Resp: 10252

Ion Ratio Lower Upper

139 100

65 119.1 67.0 100.6#

109 239.5 39.6 59.4#



#24

2,4-Dimethylphenol

Concen: 1.53 ng/ul

RT: 10.02 min Scan# 1179

Delta R.T. 0.03 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23




Tgt Ion:107 Resp: 24242


Ion Ratio Lower Upper

107 100

121 52.4 32.3 48.5#

122 15.2 61.2 91.8#









#28
Naphthalene
Concen: 4.80 ng/ul
RT: 10.86 min Scan# 1323
Delta R.T. 0.01 min
Lab File: BG026798.D
Acq: 1 May 2017 7:23




Instrument :
BNA_M
ClientSampleId :
JHSR5


Tgt Ion:128 Resp: 226923
Ion Ratio Lower Upper
128 100
129 23.6 9.4 14.2#
127 17.4 10.3 15.5#



#32
Caprolactam
Concen: 5.49 ng/ul
RT: 11.72 min Scan# 1469
Delta R.T. 0.01 min
Lab File: BG026798.D
Acq: 1 May 2017 7:23

Tgt Ion:113 Resp: 26876
Ion Ratio Lower Upper
113 100
55 852.1 168.6 252.8#
56 429.6 128.5 192.7#









#34
 2-Methylnaphthalene
 Concen: 19.04 ng/ul
 RT: 12.46 min Scan# 1595
 Delta R.T. 0.00 min
 Lab File: BG026798.D
 Acq: 1 May 2017 7:23




Instrument :
 BNA_M
 ClientSampleId :
 JHSR5


Tgt Ion:142 Resp: 668460
 Ion Ratio Lower Upper
 142 100
 141 91.4 68.6 102.8



#40
 1,1'-Biphenyl
 Concen: 3.37 ng/ul
 RT: 13.46 min Scan# 1765
 Delta R.T. 0.00 min
 Lab File: BG026798.D
 Acq: 1 May 2017 7:23

Tgt Ion:154 Resp: 136591
 Ion Ratio Lower Upper
 154 100
 153 38.7 32.2 48.2
 76 14.7 9.6 14.4#





#44

Dimethylphthalate

Concen: 5.69 ng/ul

RT: 14.07 min Scan# 1869

Delta R.T. 0.00 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion:163 Resp: 215754

Ion Ratio Lower Upper

163 100

194 11.5 3.4 5.0#




164 12.2 9.0 13.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#45

2,6-Dinitrotoluene

Concen: 1.87 ng/ul

RT: 14.17 min Scan# 1886

Delta R.T. -0.02 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion:165 Resp: 14974

Ion Ratio Lower Upper

165 100

89 20.0 52.9 79.3#



121 75.8 22.2 33.4#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BG

Ion 121.00 (120.70 to 121.70): E





#48

3-Nitroaniline

Concen: 1.33 ng/ul

RT: 14.53 min Scan# 1948

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5





Tgt Ion:138 Resp: 11180

Ion Ratio Lower Upper

138 100

108 102.0 6.3 9.5#

92 60.7 106.4 159.6#



#49

Acenaphthene

Concen: 1.67 ng/ul

RT: 14.70 min Scan# 1976

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23




Tgt Ion:153 Resp: 57625


Ion Ratio Lower Upper

153 100

152 94.4 36.5 54.7#

154 51.7 72.3 108.5#





#53

Dibenzenofuran

Concen: 1.78 ng/ul

RT: 15.02 min Scan# 2031

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :





JHSR5

Tgt Ion:168 Resp: 81555

Ion Ratio Lower Upper

168 100

139 70.5 33.8 50.8#



#58

Fluorene

Concen: 3.21 ng/ul

RT: 15.67 min Scan# 2141

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23




Tgt Ion:166 Resp: 120250


Ion Ratio Lower Upper

166 100

165 125.2 79.7 119.5#

167 37.6 11.4 17.2#





#64

N-Nitrosodiphenylamine

Concen: 8.37 ng/ul

RT: 15.87 min Scan# 2176

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion:169 Resp: 231409

Ion Ratio Lower Upper

169 100

168 27.9 59.1 88.7#




167 9.4 29.7 44.5#



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#67

Atrazine

Concen: 1.41 ng/ul

RT: 16.83 min Scan# 2338

Delta R.T. 0.02 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion:200 Resp: 12163

Ion Ratio Lower Upper

200 100

173 133.8 19.2 28.8#



215 70.2 40.3 60.5#




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Phenanthrene

Concen: 7.25 ng/ul

RT: 17.41 min Scan# 2437

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion:178 Resp: 343586

Ion Ratio Lower Upper

178 100

179 25.4 12.4 18.6#




176 22.9 16.5 24.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#73

Di-n-butylphthalate

Concen: 1.30 ng/ul

RT: 18.31 min Scan# 2591

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion:149 Resp: 69696

Ion Ratio Lower Upper

149 100

150 27.3 7.9 11.9#



104 11.1 5.8 8.8#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 2.09 ng/ul

RT: 19.41 min Scan# 2778

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion:202 Resp: 91862

Ion Ratio Lower Upper

202 100

101 20.4 6.2 9.2#




100 15.6 5.2 7.8#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 3.63 ng/ul

RT: 19.77 min Scan# 2840

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion:202 Resp: 182479

Ion Ratio Lower Upper

202 100

101 20.3 6.9 10.3#



100 16.1 6.6 10.0#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#78

Butylbenzylphthalate

Concen: 2.54 ng/ul

RT: 20.64 min Scan# 2988

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion:149 Resp: 59272

Ion Ratio Lower Upper

149 100

91 92.5 65.8 98.6

206 18.5 20.1 30.1#




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E

Time-->



Time-->



#81

Bis(2-ethylhexyl)phthalate

Concen: 23.96 ng/ul

RT: 21.48 min Scan# 3131

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23


Tgt Ion:149 Resp: 798015

Ion Ratio Lower Upper

149 100

167 28.8 21.8 32.8

279 3.5 4.5 6.7#




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E

Time-->



Time-->



#82

Chrysene

Concen: 1.39 ng/ul

RT: 21.65 min Scan# 3160

Delta R.T. 0.01 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5





Tgt Ion:228 Resp: 56573

Ion Ratio Lower Upper

228 100

226 37.1 24.0 36.0#

229 44.7 16.9 25.3#



#85

Benzo(b)fluoranthene

Concen: 1.08 ng/ul

RT: 23.78 min Scan# 3521

Delta R.T. 0.02 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23




Tgt Ion:252 Resp: 50053


Ion Ratio Lower Upper

252 100

253 65.5 17.7 26.5#

125 391.9 4.0 6.0#





#86

Benzo(k)fluoranthene

Concen: 1.10 ng/ul

RT: 23.78 min Scan# 3521

Delta R.T. -0.05 min

Lab File: BG026798.D

Acq: 1 May 2017 7:23

Instrument :

BNA_M

ClientSampleId :

JHSR5


Tgt Ion: 252 Resp: 50053

Ion Ratio Lower Upper

252 100

253 65.5 18.5 27.7#

125 391.9 4.1 6.1#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

