

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM043021\
 Data File : BM029724.D
 Acq On : 30 Apr 2021 11:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 12:46:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM043021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 17:34:54 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	121	0.00
2	1,4-Dioxane	40.000	38.890	2.8	114	0.00
3	Pyridine	40.000	43.428	-8.6	126	0.00
4	n-Nitrosodimethylamine	40.000	37.097	7.3	114	0.00
5 S	2-Fluorophenol	80.000	77.732	2.8	118	0.00
6	Aniline	40.000	39.805	0.5	120	0.00
7 S	Phenol-d6	80.000	80.267	-0.3	121	0.00
8	2-Chlorophenol	40.000	39.638	0.9	120	0.00
9	Benzaldehyde	40.000	38.176	4.6	117	0.00
10 C	Phenol	40.000	39.933	0.2	120	0.00
11	bis(2-Chloroethyl)ether	40.000	39.491	1.3	122	0.00
12	1,3-Dichlorobenzene	40.000	39.495	1.3	120	0.00
13 C	1,4-Dichlorobenzene	40.000	39.348	1.6	119	0.00
14	1,2-Dichlorobenzene	40.000	38.997	2.5	121	0.00
15	Benzyl Alcohol	40.000	40.576	-1.4	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.172	12.1	111	0.00
17	2-Methylphenol	40.000	40.011	-0.0	122	0.00
18	Hexachloroethane	40.000	38.585	3.5	116	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.114	-0.3	120	0.00
20	3+4-Methylphenols	40.000	41.546	-3.9	123	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	122	0.00
22	Acetophenone	40.000	40.095	-0.2	121	0.00
23 S	Nitrobenzene-d5	80.000	79.996	0.0	120	0.00
24	Nitrobenzene	40.000	39.585	1.0	120	0.00
25	Isophorone	40.000	40.100	-0.3	119	0.00
26 C	2-Nitrophenol	40.000	43.606	-9.0	130	0.00
27	2,4-Dimethylphenol	40.000	40.691	-1.7	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.631	0.9	118	0.00
29 C	2,4-Dichlorophenol	40.000	42.428	-6.1	126	0.00
30	1,2,4-Trichlorobenzene	40.000	40.973	-2.4	125	0.00
31	Naphthalene	40.000	39.658	0.9	122	0.00
32	Benzoic acid	40.000	42.319	-5.8	131	0.02
33	4-Chloroaniline	40.000	41.127	-2.8	122	0.00
34 C	Hexachlorobutadiene	40.000	42.434	-6.1	128	0.00
35	Caprolactam	40.000	42.883	-7.2	123	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.177	-2.9	122	0.00
37	2-Methylnaphthalene	40.000	40.858	-2.1	123	0.00
38	1-Methylnaphthalene	40.000	40.874	-2.2	125	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.694	-6.7	129	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.367	1.6	125	0.00
42 S	2,4,6-Tribromophenol	80.000	83.943	-4.9	126	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.968	-7.4	127	0.00
44	2,4,5-Trichlorophenol	40.000	42.834	-7.1	128	0.00
45 S	2-Fluorobiphenyl	80.000	82.473	-3.1	126	0.00
46	1,1'-Biphenyl	40.000	40.609	-1.5	124	0.00
47	2-Chloronaphthalene	40.000	40.491	-1.2	124	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM043021\
 Data File : BM029724.D
 Acq On : 30 Apr 2021 11:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 12:46:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM043021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 17:34:54 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.620	-4.0	123	0.00
49	Acenaphthylene	40.000	40.203	-0.5	123	0.00
50	Dimethylphthalate	40.000	39.268	1.8	121	0.00
51	2,6-Dinitrotoluene	40.000	40.673	-1.7	122	0.00
52 C	Acenaphthene	40.000	40.038	-0.1	123	0.00
53	3-Nitroaniline	40.000	41.800	-4.5	123	0.00
54 P	2,4-Dinitrophenol	40.000	39.700	0.7	124	0.00
55	Dibenzofuran	40.000	40.516	-1.3	123	0.00
56 P	4-Nitrophenol	40.000	40.675	-1.7	119	0.00
57	2,4-Dinitrotoluene	40.000	41.312	-3.3	120	0.00
58	Fluorene	40.000	40.063	-0.2	123	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.236	-5.6	128	0.00
60	Diethylphthalate	40.000	38.629	3.4	118	0.00
61	4-Chlorophenyl-phenylether	40.000	41.009	-2.5	126	0.00
62	4-Nitroaniline	40.000	41.657	-4.1	123	0.00
63	Azobenzene	40.000	38.283	4.3	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.150	-7.9	125	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.724	-4.3	122	0.00
67	4-Bromophenyl-phenylether	40.000	42.613	-6.5	124	0.00
68	Hexachlorobenzene	40.000	42.177	-5.4	126	0.00
69	Atrazine	40.000	40.680	-1.7	118	0.00
70 C	Pentachlorophenol	40.000	42.824	-7.1	121	0.00
71	Phenanthrene	40.000	39.643	0.9	122	0.00
72	Anthracene	40.000	40.077	-0.2	122	0.00
73	Carbazole	40.000	39.764	0.6	121	0.00
74	Di-n-butylphthalate	40.000	41.659	-4.1	125	0.00
75 C	Fluoranthene	40.000	41.318	-3.3	125	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	110	0.00
77	Benzydine	40.000	37.996	5.0	106	0.00
78	Pyrene	40.000	41.940	-4.8	124	0.00
79 S	Terphenyl-d14	80.000	84.851	-6.1	124	0.00
80	Butylbenzylphthalate	40.000	41.500	-3.8	119	0.00
81	Benzo(a)anthracene	40.000	40.009	-0.0	110	0.00
82	3,3'-Dichlorobenzidine	40.000	40.755	-1.9	112	0.00
83	Chrysene	40.000	39.187	2.0	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.378	-0.9	108	0.00
85 c	Di-n-octyl phthalate	40.000	40.372	-0.9	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.826	2.9	102	-0.01
88	Benzo(b)fluoranthene	40.000	41.034	-2.6	112	0.00
89	Benzo(k)fluoranthene	40.000	39.256	1.9	110	0.00
90 C	Benzo(a)pyrene	40.000	40.008	-0.0	109	0.00
91	Dibenzo(a,h)anthracene	40.000	38.402	4.0	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.560	3.6	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM043021\
Data File : BM029724.D
Acq On : 30 Apr 2021 11:12
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Apr 30 12:46:54 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM043021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 29 17:34:54 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0