

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050115\
 Data File : BM001152.D
 Acq On : 02 May 2015 03:55
 Operator : TP/IZ
 Sample : PB83102BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB83102BL

Quant Time: May 02 05:57:46 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 02 02:22:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	16119	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	61055	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	33860	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	86562	5.00	ng	0.00
24) Chrysene-d12	21.15	240	69248	5.00	ng	0.00
30) Perylene-d12	23.29	264	55592	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.11	112	44545	14.21	ng/uL	-0.01
5) Phenol-d6	6.71	99	55816	14.45	ng/uL	0.00
7) Nitrobenzene-d5	8.69	82	34475	9.97	ng	0.00
13) 2,4,6-Tribromophenol	15.69	330	19211	11.54	ng/uL	0.00
14) 2-Fluorobiphenyl	12.81	172	110056	9.83	ng	0.00
26) Terphenyl-d14	19.59	244	96463	10.10	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) n-Nitrosodimethylamine	3.55	42	7	0.00	ng	# 28
8) Nitrobenzene	8.69	77	133	0.18	ng/uL	# 37
9) Naphthalene	10.36	128	11	0.00	ng	# 1
10) Hexachlorobutadiene	10.46	225	21	0.01	ng/uL	# 21
11) 2-Methylnaphthalene	11.98	142	2	0.00	ng	# 45
15) Acenaphthylene	13.56	152	328	0.02	ng	# 67
16) Acenaphthene	14.18	154	153	0.01	ng	# 4
19) Hexachlorobenzene	15.95	284	475	0.10	ng	# 42
20) Pentachlorophenol	16.63	266	2	0.22	ng	# 58
21) Phenanthrene	16.94	178	59	0.00	ng	# 1
22) Anthracene	17.08	178	21	0.00	ng	# 61
23) Fluoranthene	19.01	202	17	0.00	ng	# 80
25) Pyrene	19.38	202	25	0.00	ng	# 51
27) Benzo(a)anthracene	21.14	228	275	0.01	ng	# 78
28) Chrysene	21.14	228	275	0.01	ng	# 81
29) Indeno(1,2,3-cd)pyrene	25.43	276	71	0.00	ng	# 86
31) Benzo(b)fluoranthene	22.66	252	59	0.00	ng	# 1
32) Benzo(k)fluoranthene	22.70	252	32	0.00	ng	# 1
33) Benzo(a)pyrene	23.20	252	53	0.00	ng	# 1
34) Dibenzo(a,h)anthracene	25.45	278	89	0.01	ng	# 1
35) Benzo(g,h,i)perylene	26.08	276	56	0.00	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050115\
Data File : BM001152.D
Acq On : 02 May 2015 03:55
Operator : TP/IZ
Sample : PB83102BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB83102BL

Quant Time: May 02 05:57:46 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 02 02:22:19 2015
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.52 min Scan# 314
Delta R.T. -0.00 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Instrument :
BNA_M
ClientSampleId :
PB83102BL

Tgt Ion:152 Resp: 16119
Ion Ratio Lower Upper
152 100
150 172.3 140.1 210.1
115 69.8 57.8 86.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#3


n-Nitrosodimethylamine
Concen: 0.00 ng
RT: 3.55 min Scan# 57
Delta R.T. 0.22 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion: 42 Resp: 7
Ion Ratio Lower Upper
42 100
74 57.1 72.0 108.0#
44 271.4 3.7 5.5#



Abundance Ion 42.00 (41.70 to 42.70): BM0
Ion 74.00 (73.70 to 74.70): BM0
Ion 44.00 (43.70 to 44.70): BM0





#4

2-Fluorophenol

Concen: 14.21 ng/uL

RT: 5.11 min Scan# 158

Delta R.T. -0.01 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion: 112 Resp: 44545

Ion Ratio Lower Upper

112 100

64 65.1 52.6 79.0




63 35.8 29.5 44.3



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 63.00 (62.70 to 63.70): BM0



#5

Phenol-d6

Concen: 14.45 ng/uL

RT: 6.71 min Scan# 261

Delta R.T. 0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion: 99 Resp: 55816

Ion Ratio Lower Upper

99 100

42 24.4 20.9 31.3



71 35.2 28.5 42.7




Abundance Ion 99.00 (98.70 to 99.70): BM0

Ion 42.00 (41.70 to 42.70): BM0

Ion 71.00 (70.70 to 71.70): BM0





#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.31 min Scan# 532

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion:136 Resp: 61055

Ion Ratio Lower Upper

136 100

137 10.5 8.2 12.4




54 11.5 9.0 13.4



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BM



#7

Nitrobenzene-d5

Concen: 9.97 ng

RT: 8.69 min Scan# 402

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion: 82 Resp: 34475

Ion Ratio Lower Upper

82 100

128 40.9 31.9 47.9



54 53.1 46.2 69.2




Abundance Ion 82.00 (81.70 to 82.70): BM

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BM





#8

Nitrobenzene

Concen: 0.18 ng/uL

RT: 8.69 min Scan# 402

Delta R.T. -0.04 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion: 77 Resp: 133

Ion Ratio Lower Upper

77 100

123 0.0 31.0 46.6#




65 36.8 10.6 15.8#



Abundance Ion 77.00 (76.70 to 77.70): BM001142.D (-400) (-)

Ion 123.00 (122.70 to 123.70): BM001142.D (-400) (-)

Ion 65.00 (64.70 to 65.70): BM001142.D (-400) (-)



#9

Naphthalene

Concen: 0.00 ng

RT: 10.36 min Scan# 535

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion: 128 Resp: 11

Ion Ratio Lower Upper

128 100

129 92.5 9.0 13.6#



127 90.6 10.0 15.0#




Abundance Ion 128.00 (127.70 to 128.70): BM001142.D (-529) (-)

Ion 129.00 (128.70 to 129.70): BM001142.D (-529) (-)

Ion 127.00 (126.70 to 127.70): BM001142.D (-529) (-)









#10
Hexachlorobutadiene
Concen: 0.01 ng/uL
RT: 10.46 min Scan# 541
Delta R.T. -0.19 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55




Instrument :
BNA_M
ClientSampleId :
PB83102BL


Tgt Ion: 225 Resp: 21
Ion Ratio Lower Upper
225 100
223 0.0 50.8 76.2#
227 4.8 51.0 76.4#



#11
2-Methylnaphthalene
Concen: 0.00 ng
RT: 11.98 min Scan# 660
Delta R.T. -0.01 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion: 142 Resp: 2
Ion Ratio Lower Upper
142 100
141 102.1 65.1 97.7#
115 108.3 25.1 37.7#





#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.20 min Scan# 838

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion:164 Resp: 33860

Ion Ratio Lower Upper

164 100

162 90.2 72.7 109.1




160 35.1 28.7 43.1



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#13

2,4,6-Tribromophenol

Concen: 11.54 ng/uL

RT: 15.69 min Scan# 905

Delta R.T. 0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion:330 Resp: 19211

Ion Ratio Lower Upper

330 100

332 92.5 74.5 111.7



141 43.2 21.0 31.4#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#14
2-Fluorobiphenyl
Concen: 9.83 ng
RT: 12.81 min Scan# 740
Delta R.T. -0.00 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55




Instrument :
BNA_M
ClientSampleId :
PB83102BL


Tgt Ion:172 Resp: 110056
Ion Ratio Lower Upper
172 100
171 31.5 25.8 38.6
170 18.9 16.1 24.1



#15
Acenaphthylene
Concen: 0.02 ng
RT: 13.56 min Scan# 809
Delta R.T. -0.36 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion:152 Resp: 328
Ion Ratio Lower Upper
152 100
151 0.0 0.0 0.0
153 0.0 10.3 15.5#





#16

Acenaphthene

Concen: 0.01 ng

RT: 14.18 min Scan# 837

Delta R.T. -0.07 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion:154 Resp: 153

Ion Ratio Lower Upper

154 100

153 0.0 90.6 136.0#




152 0.0 45.0 67.4#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.94 min Scan# 965

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion:188 Resp: 86562

Ion Ratio Lower Upper

188 100

94 11.3 9.3 13.9



80 11.7 9.8 14.6




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0









#19
Hexachlorobenzene
Concen: 0.10 ng
RT: 15.95 min Scan# 917
Delta R.T. -0.31 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55




Instrument :
BNA_M
ClientSampleId :
PB83102BL


Tgt Ion: 284 Resp: 475
Ion Ratio Lower Upper
284 100
142 0.0 0.0 0.0
249 0.0 26.5 39.7#



#20
Pentachlorophenol
Concen: 0.22 ng
RT: 16.63 min Scan# 950
Delta R.T. 0.02 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion: 266 Resp: 2
Ion Ratio Lower Upper
266 100
268 100.0 52.6 79.0#
264 95.9 51.0 76.4#









#21
Phenanthrene
Concen: 0.00 ng
RT: 16.94 min Scan# 965
Delta R.T. -0.04 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55




Instrument :
BNA_M
ClientSampleId :
PB83102BL


Tgt Ion:178 Resp: 59
Ion Ratio Lower Upper
178 100
176 0.0 15.5 23.3#
179 100.0 12.0 18.0#



#22
Anthracene
Concen: 0.00 ng
RT: 17.08 min Scan# 972
Delta R.T. -0.00 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion:178 Resp: 21
Ion Ratio Lower Upper
178 100
176 0.0 14.3 21.5#
179 0.0 12.2 18.2#





#23

Fluoranthene

Concen: 0.00 ng

RT: 19.01 min Scan# 1095

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion:202 Resp: 17

Ion Ratio Lower Upper

202 100

101 0.0 9.3 13.9#




203 11.8 13.7 20.5#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.15 min Scan# 1279

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion:240 Resp: 69248

Ion Ratio Lower Upper

240 100

120 8.0 9.2 13.8#



236 26.2 22.4 33.6




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25
 Pyrene
 Concen: 0.00 ng
 RT: 19.38 min Scan# 1125
 Delta R.T. -0.00 min
 Lab File: BM001152.D
 Acq: 02 May 2015 03:55




Instrument :
 BNA_M
 ClientSampleId :
 PB83102BL


Tgt Ion: 202 Resp: 25
 Ion Ratio Lower Upper
 202 100
 200 28.0 16.4 24.6#
 203 56.0 13.8 20.6#



#26
 Terphenyl-d14
 Concen: 10.10 ng
 RT: 19.59 min Scan# 1143
 Delta R.T. -0.00 min
 Lab File: BM001152.D
 Acq: 02 May 2015 03:55

Tgt Ion: 244 Resp: 96463
 Ion Ratio Lower Upper
 244 100
 212 8.2 7.1 10.7
 122 5.9 6.2 9.4#





#27

Benzo(a)anthracene

Concen: 0.01 ng

RT: 21.14 min Scan# 1278

Delta R.T. 0.01 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL


Tgt Ion:228 Resp: 275

Ion Ratio Lower Upper

228 100

226 31.0 22.8 34.2




229 39.9 14.4 21.6#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 0.01 ng

RT: 21.14 min Scan# 1278

Delta R.T. -0.04 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55


Tgt Ion:228 Resp: 275

Ion Ratio Lower Upper

228 100

226 31.0 24.4 36.6



229 39.9 14.3 21.5#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#29

Indeno(1,2,3-cd)pyrene

Concen: 0.00 ng

RT: 25.43 min Scan# 1690

Delta R.T. -0.00 min

Lab File: BM001152.D


Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL






Tgt Ion: 276 Resp: 71

Ion Ratio Lower Upper

276 100

138 39.4 24.2 36.4#

277 29.6 19.8 29.6



#30

Perylene-d12

Concen: 5.00 ng

RT: 23.29 min Scan# 1485

Delta R.T. -0.01 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55




Tgt Ion: 264 Resp: 55592


Ion Ratio Lower Upper

264 100

260 25.5 18.9 28.3

265 21.1 18.1 27.1









#31
Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 22.66 min Scan# 1424
Delta R.T. -0.00 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55




Instrument :
BNA_M
ClientSampleId :
PB83102BL


Tgt Ion: 252 Resp: 59
Ion Ratio Lower Upper
252 100
253 146.4 19.0 28.4#
125 82.5 9.1 13.7#



#32
Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 22.70 min Scan# 1428
Delta R.T. -0.00 min
Lab File: BM001152.D
Acq: 02 May 2015 03:55

Tgt Ion: 252 Resp: 32
Ion Ratio Lower Upper
252 100
253 150.0 19.1 28.7#
125 87.0 9.3 13.9#





#33

Benzo(a)pyrene

Concen: 0.00 ng

RT: 23.20 min Scan# 1476

Delta R.T. -0.00 min

Lab File: BM001152.D

Acq: 02 May 2015 03:55

Instrument :





BNA_M

ClientSampleId :

PB83102BL

Tgt Ion: 252 Resp: 53

Ion	Ratio	Lower	Upper
252	100		
253	162.0	18.2	27.2#
125	90.2	10.6	16.0#



#34

Dibenzo(a,h)anthracene

Concen: 0.01 ng

RT: 25.45 min Scan# 1692




Delta R.T. -0.00 min


Lab File: BM001152.D

Acq: 02 May 2015 03:55

Tgt Ion: 278 Resp: 89

Ion	Ratio	Lower	Upper
278	100		
139	91.7	15.5	23.3#
279	120.0	19.7	29.5#





#35

Benzo(g,h,i)perylene

Concen: 0.00 ng

RT: 26.08 min Scan# 1752

Delta R.T. -0.01 min

Lab File: BM001152.D


Acq: 02 May 2015 03:55

Instrument :

BNA_M

ClientSampleId :

PB83102BL




Tgt Ion: 276 Resp: 56

Ion Ratio Lower Upper

276 100

277 90.0 19.0 28.4#

138 93.3 21.3 31.9#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

