

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050115\
 Data File : BM001183.D
 Acq On : 03 May 2015 00:47
 Operator : TP/IZ
 Sample : G2058-11
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EB2015-04-28

Quant Time: May 04 01:54:46 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 01:38:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	15422	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	56305	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	31612	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	82771	5.00	ng	0.00
24) Chrysene-d12	21.15	240	75337	5.00	ng	0.00
30) Perylene-d12	23.29	264	68434	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.13	112	203898	67.98	ng/uL	0.02
5) Phenol-d6	6.72	99	165344	44.74	ng/uL	0.02
7) Nitrobenzene-d5	8.69	82	387877	121.61	ng	0.00
13) 2,4,6-Tribromophenol	15.69	330	254149	161.49	ng/uL	0.00
14) 2-Fluorobiphenyl	12.81	172	1042189	99.69	ng	0.00
26) Terphenyl-d14	19.59	244	994002	95.69	ng	0.00


Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.61	42	70	0.04	ng	# 52
8) Nitrobenzene	8.69	77	1355	0.47	ng/uL#	37
9) Naphthalene	10.36	128	156	0.01	ng	# 17
10) Hexachlorobutadiene	10.65	225	1	0.00	ng/uL#	18
11) 2-Methylnaphthalene	11.99	142	31	0.00	ng	# 51
15) Acenaphthylene	13.51	152	476	0.03	ng	# 1
16) Acenaphthene	14.18	154	162	0.02	ng	# 4
19) Hexachlorobenzene	15.97	284	305	0.07	ng	# 42
20) Pentachlorophenol	16.57	266	315	0.43	ng	# 6
21) Phenanthrene	16.98	178	207	0.01	ng	# 84
22) Anthracene	17.08	178	64	0.00	ng	# 61
23) Fluoranthene	19.01	202	104	0.00	ng	# 85
25) Pyrene	19.59	202	4754	0.20	ng	# 65
27) Benzo(a)anthracene	21.14	228	365	0.02	ng	# 80
28) Chrysene	21.14	228	365	0.02	ng	# 82
29) Indeno(1,2,3-cd)pyrene	25.43	276	123	0.01	ng	# 1
31) Benzo(b)fluoranthene	22.65	252	103	0.00	ng	# 1
32) Benzo(k)fluoranthene	22.69	252	71	0.00	ng	# 1
33) Benzo(a)pyrene	23.19	252	81	0.00	ng	# 1
34) Dibenzo(a,h)anthracene	25.44	278	124	0.01	ng	# 1
35) Benzo(g,h,i)perylene	26.08	276	89	0.01	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050115\
Data File : BM001183.D
Acq On : 03 May 2015 00:47
Operator : TP/IZ
Sample : G2058-11
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EB2015-04-28

Quant Time: May 04 01:54:46 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 04 01:38:30 2015
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.52 min Scan# 314
Delta R.T. -0.00 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Instrument :
BNA_M
ClientSampleId :
EB2015-04-28

Tgt Ion: 152 Resp: 15422
Ion Ratio Lower Upper
152 100
150 167.3 140.1 210.1
115 65.5 57.8 86.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#3


n-Nitrosodimethylamine
Concen: 0.04 ng
RT: 3.61 min Scan# 61
Delta R.T. 0.28 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Tgt Ion: 42 Resp: 70
Ion Ratio Lower Upper
42 100
74 45.7 72.0 108.0#
44 30.0 3.7 5.5#



Abundance Ion 42.00 (41.70 to 42.70): BM0
Ion 74.00 (73.70 to 74.70): BM0
Ion 44.00 (43.70 to 44.70): BM0





#4

2-Fluorophenol

Concen: 67.98 ng/uL

RT: 5.13 min Scan# 159

Delta R.T. 0.02 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion: 112 Resp: 203898

Ion Ratio Lower Upper

112 100

64 64.0 52.6 79.0




63 35.3 29.5 44.3



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#5

Phenol-d6

Concen: 44.74 ng/uL

RT: 6.72 min Scan# 262

Delta R.T. 0.02 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion: 99 Resp: 165344

Ion Ratio Lower Upper

99 100

42 23.7 20.9 31.3



71 34.8 28.5 42.7




Abundance Ion 99.00 (98.70 to 99.70): B

Ion 42.00 (41.70 to 42.70): B

Ion 71.00 (70.70 to 71.70): B





#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.31 min Scan# 532

Delta R.T. -0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:136 Resp: 56305

Ion Ratio Lower Upper

136 100

137 10.5 8.2 12.4




54 9.8 9.0 13.4



Abundance Ion 136.00 (135.70 to 136.70): B

Ion 137.00 (136.70 to 137.70): B

Ion 54.00 (53.70 to 54.70): B



#7

Nitrobenzene-d5

Concen: 121.61 ng

RT: 8.69 min Scan# 402

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion: 82 Resp: 387877

Ion Ratio Lower Upper

82 100

128 43.4 31.9 47.9



54 51.9 46.2 69.2




Abundance Ion 82.00 (81.70 to 82.70): B

Ion 128.00 (127.70 to 128.70): B

Ion 54.00 (53.70 to 54.70): B





#8

Nitrobenzene

Concen: 0.47 ng/uL

RT: 8.69 min Scan# 402

Delta R.T. -0.04 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion: 77 Resp: 1355

Ion Ratio Lower Upper

77 100

123 0.0 31.0 46.6#




65 37.8 10.6 15.8#



Abundance Ion 77.00 (76.70 to 77.70): BM001142.D (-400) (-)

Ion 123.00 (122.70 to 123.70): BM001142.D (-400) (-)

Ion 65.00 (64.70 to 65.70): BM001142.D (-400) (-)



#9

Naphthalene

Concen: 0.01 ng

RT: 10.36 min Scan# 535

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion: 128 Resp: 156

Ion Ratio Lower Upper

128 100

129 44.4 9.0 13.6#



127 43.7 10.0 15.0#




Abundance Ion 128.00 (127.70 to 128.70): BM001142.D (-529) (-)

Ion 129.00 (128.70 to 129.70): BM001142.D (-529) (-)

Ion 127.00 (126.70 to 127.70): BM001142.D (-529) (-)









#10
Hexachlorobutadiene
Concen: 0.00 ng/uL
RT: 10.65 min Scan# 552
Delta R.T. 0.00 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47




Instrument :
BNA_M
ClientSampleId :
EB2015-04-28


Tgt Ion: 225 Resp: 1
Ion Ratio Lower Upper
225 100
223 0.0 50.8 76.2#
227 0.0 51.0 76.4#



#11
2-Methylnaphthalene
Concen: 0.00 ng
RT: 11.99 min Scan# 661
Delta R.T. 0.01 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Tgt Ion: 142 Resp: 31
Ion Ratio Lower Upper
142 100
141 98.5 65.1 97.7#
115 101.5 25.1 37.7#





#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.20 min Scan# 838

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:164 Resp: 31612

Ion Ratio Lower Upper

164 100

162 89.5 72.7 109.1




160 35.1 28.7 43.1



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#13

2,4,6-Tribromophenol

Concen: 161.49 ng/uL

RT: 15.69 min Scan# 905

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion:330 Resp: 254149

Ion Ratio Lower Upper

330 100

332 96.6 74.5 111.7



141 43.7 21.0 31.4#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#14
2-Fluorobiphenyl
Concen: 99.69 ng
RT: 12.81 min Scan# 740
Delta R.T. 0.00 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47




Instrument :
BNA_M
ClientSampleId :
EB2015-04-28


Tgt Ion:172 Resp: 1042189
Ion Ratio Lower Upper
172 100
171 32.3 25.8 38.6
170 19.7 16.1 24.1



#15
Acenaphthylene
Concen: 0.03 ng
RT: 13.51 min Scan# 807
Delta R.T. -0.38 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Tgt Ion:152 Resp: 476
Ion Ratio Lower Upper
152 100
151 0.0 0.0 0.0
153 192.9 10.3 15.5#





#16

Acenaphthene

Concen: 0.02 ng

RT: 14.18 min Scan# 837

Delta R.T. -0.07 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:154 Resp: 162

Ion Ratio Lower Upper

154 100

153 0.0 90.6 136.0#




152 0.0 45.0 67.4#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.94 min Scan# 965

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion:188 Resp: 82771

Ion Ratio Lower Upper

188 100

94 10.1 9.3 13.9



80 10.4 9.8 14.6




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0









#19
Hexachlorobenzene
Concen: 0.07 ng
RT: 15.97 min Scan# 918
Delta R.T. -0.29 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Instrument :
BNA_M
ClientSampleId :
EB2015-04-28

Tgt Ion: 284 Resp: 305
Ion Ratio Lower Upper
284 100
142 0.0 0.0 0.0
249 0.0 26.5 39.7#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#20
Pentachlorophenol
Concen: 0.43 ng
RT: 16.57 min Scan# 947
Delta R.T. -0.04 min
Lab File: BM001183.D
Acq: 03 May 2015 00:47

Tgt Ion: 266 Resp: 315
Ion Ratio Lower Upper
266 100
268 174.9 52.6 79.0#
264 24.6 51.0 76.4#



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#21

Phenanthrene

Concen: 0.01 ng

RT: 16.98 min Scan# 967

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:178 Resp: 207

Ion Ratio Lower Upper

178 100

176 19.8 15.5 23.3

179 0.0 12.0 18.0#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E

16.98

150



100

50

0

16.80 16.90 17.00

Time-->



#22

Anthracene

Concen: 0.00 ng

RT: 17.08 min Scan# 972

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion:178 Resp: 64

Ion Ratio Lower Upper

178 100

176 0.0 14.3 21.5#

179 0.0 12.2 18.2#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E

17.08

150


100


50

0

17.00 17.20

Time-->





#23

Fluoranthene

Concen: 0.00 ng

RT: 19.01 min Scan# 1095

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :


BNA_M

ClientSampleId :

EB2015-04-28

Tgt Ion:202 Resp: 104




Ion	Ratio	Lower	Upper
202	100		
101	0.0	9.3	13.9#
203	19.2	13.7	20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.15 min Scan# 1279


Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Tgt Ion:240 Resp: 75337



Ion	Ratio	Lower	Upper
240	100		
120	7.5	9.2	13.8#
236	26.0	22.4	33.6




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25
 Pyrene
 Concen: 0.20 ng
 RT: 19.59 min Scan# 1143
 Delta R.T. 0.22 min
 Lab File: BM001183.D
 Acq: 03 May 2015 00:47




Instrument :
 BNA_M
 ClientSampleId :
 EB2015-04-28


Tgt Ion:202 Resp: 4754
 Ion Ratio Lower Upper
 202 100
 200 50.0 16.4 24.6#
 203 17.5 13.8 20.6



#26
 Terphenyl-d14
 Concen: 95.69 ng
 RT: 19.59 min Scan# 1143
 Delta R.T. 0.00 min
 Lab File: BM001183.D
 Acq: 03 May 2015 00:47

Tgt Ion:244 Resp: 994002
 Ion Ratio Lower Upper
 244 100
 212 9.8 7.1 10.7
 122 10.7 6.2 9.4#





#27

Benzo(a)anthracene

Concen: 0.02 ng

RT: 21.14 min Scan# 1278

Delta R.T. 0.01 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:228 Resp: 365

Ion Ratio Lower Upper

228 100

226 29.9 22.8 34.2




229 39.0 14.4 21.6#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 0.02 ng

RT: 21.14 min Scan# 1278

Delta R.T. -0.04 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion:228 Resp: 365

Ion Ratio Lower Upper

228 100

226 29.9 24.4 36.6



229 39.0 14.3 21.5#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#29

Indeno(1,2,3-cd)pyrene

Concen: 0.01 ng

RT: 25.43 min Scan# 1690

Delta R.T. 0.00 min

Lab File: BM001183.D


Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28






Tgt Ion: 276 Resp: 123

Ion Ratio Lower Upper

276 100

138 165.0 24.2 36.4#

277 31.7 19.8 29.6#



#30

Perylene-d12

Concen: 5.00 ng

RT: 23.29 min Scan# 1485

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47




Tgt Ion: 264 Resp: 68434


Ion Ratio Lower Upper

264 100

260 25.0 18.9 28.3

265 21.6 18.1 27.1





#31

Benzo(b)fluoranthene

Concen: 0.00 ng

RT: 22.65 min Scan# 1423

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion:252 Resp: 103

Ion Ratio Lower Upper

252 100

253 131.2 19.0 28.4#




125 89.4 9.1 13.7#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#32

Benzo(k)fluoranthene

Concen: 0.00 ng

RT: 22.69 min Scan# 1427

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion:252 Resp: 71

Ion Ratio Lower Upper

252 100

253 135.8 19.1 28.7#



125 82.8 9.3 13.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#33

Benzo(a)pyrene

Concen: 0.00 ng

RT: 23.19 min Scan# 1475

Delta R.T. -0.01 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28


Tgt Ion: 252 Resp: 81

Ion Ratio Lower Upper

252 100

253 144.6 18.2 27.2#




125 90.8 10.6 16.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#34

Dibenzo(a,h)anthracene

Concen: 0.01 ng

RT: 25.44 min Scan# 1691

Delta R.T. 0.00 min

Lab File: BM001183.D

Acq: 03 May 2015 00:47


Tgt Ion: 278 Resp: 124

Ion Ratio Lower Upper

278 100

139 109.3 15.5 23.3#



279 121.3 19.7 29.5#




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#35

Benzo(g,h,i)perylene

Concen: 0.01 ng

RT: 26.08 min Scan# 1752

Delta R.T. -0.01 min

Lab File: BM001183.D


Acq: 03 May 2015 00:47

Instrument :

BNA_M

ClientSampleId :

EB2015-04-28




Tgt Ion: 276 Resp: 89

Ion Ratio Lower Upper

276 100

277 82.4 19.0 28.4#

138 93.2 21.3 31.9#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

26.08

Time--> 25.90 26.00 26.10 26.20