


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050115\
 Data File : BM001138.D
 Acq On : 01 May 2015 19:33
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 30 19:07:47 2015



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2495

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.3	88453	PASS
68	69	0.00	2	1.1	869	PASS
69	198	0.00	100	46.7	76144	PASS
70	69	0.00	2	0.4	309	PASS
127	198	10	80	53.5	87293	PASS
197	198	0.00	2	0.5	844	PASS
198	198	100	100	100.0	163018	PASS
199	198	5	9	6.5	10645	PASS
275	198	10	60	23.6	38469	PASS
365	198	1	100	3.0	4813	PASS
441	442	0.01	24	14.1	18866	PASS
442	198	50	100	81.9	133493	PASS
443	442	15	24	19.4	25868	PASS