

Data Path : Z:\HPCHEM1\BNA M\Data\BM050115\
 Data File : BM001153.D
 Acq On : 02 May 2015 04:31
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 04 17:09:06 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	0.00
2	1,4-Dioxane	1.000	1.071	-7.1	94	0.02
3	n-Nitrosodimethylamine	1.000	0.919	8.1	85	0.00
4 S	2-Fluorophenol	1.000	1.009	-0.9	97	0.00
5 S	Phenol-d6	1.000	0.996	0.4	97	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	97	0.00
7 S	Nitrobenzene-d5	1.000	0.988	1.2	97	0.00
8	Nitrobenzene	1.000	0.981	1.9	96	0.00
9	Naphthalene	1.000	1.010	-1.0	96	0.00
10	Hexachlorobutadiene	1.000	1.011	-1.1	99	0.00
11	2-Methylnaphthalene	1.000	1.024	-2.4	100	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	1.000	1.267	-26.7#	157	0.00
14 S	2-Fluorobiphenyl	1.000	1.032	-3.2	101	0.00
15	Acenaphthylene	1.000	1.026	-2.6	101	-0.02
16	Acenaphthene	1.000	1.025	-2.5	103	0.00
17	Fluorene	1.000	1.039	-3.9	100	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.00
19	Hexachlorobenzene	1.000	0.990	1.0	102	0.00
20	Pentachlorophenol	1.000	0.858	14.2	95	0.00
21	Phenanthrene	1.000	0.986	1.4	101	0.00
22	Anthracene	1.000	1.004	-0.4	100	0.00
23	Fluoranthene	1.000	1.001	-0.1	102	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	101	0.00
25	Pyrene	1.000	0.995	0.5	101	0.00
26 S	Terphenyl-d14	1.000	1.015	-1.5	101	0.00
27	Benzo(a)anthracene	1.000	0.999	0.1	102	0.00
28	Chrysene	1.000	1.001	-0.1	98	0.00
29	Indeno(1,2,3-cd)pyrene	1.000	1.023	-2.3	100	0.00
30 I	Perylene-d12	5.000	5.000	0.0	100	-0.01
31	Benzo(b)fluoranthene	1.000	0.989	1.1	92	-0.01
32	Benzo(k)fluoranthene	1.000	1.046	-4.6	112	-0.01
33 C	Benzo(a)pyrene	1.000	1.013	-1.3	101	0.00
34	Dibenzo(a,h)anthracene	1.000	1.020	-2.0	101	-0.01
35	Benzo(q,h,i)perylene	1.000	1.010	-1.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0