

Data Path : Z:\HPCHEM1\BNA M\Data\BM050115\
 Data File : BM001170.D
 Acq On : 02 May 2015 15:50
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 04 17:35:54 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.612	0.451	26.3#	73	0.00
3	n-Nitrosodimethylamine	0.612	0.602	1.6	94	0.00
4 S	2-Fluorophenol	0.972	0.987	-1.5	101	-0.01
5 S	Phenol-d6	1.198	1.215	-1.4	101	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
7 S	Nitrobenzene-d5	0.283	0.279	1.4	99	0.00
8	Nitrobenzene	0.303	0.298	1.7	93	0.00
9	Naphthalene	1.099	1.109	-0.9	99	0.00
10	Hexachlorobutadiene	0.212	0.216	-1.9	102	0.00
11	2-Methylnaphthalene	0.754	0.786	-4.2	104	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
13 S	2,4,6-Tribromophenol	0.188	0.279	-48.4#	164#	0.00
14 S	2-Fluorobiphenyl	1.654	1.682	-1.7	103	0.00
15	Acenaphthylene	2.185	2.247	-2.8	104	-0.02
16	Acenaphthene	1.510	1.534	-1.6	106	0.00
17	Fluorene	1.751	1.848	-5.5	104	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
19	Hexachlorobenzene	0.269	0.274	-1.9	107	0.00
20	Pentachlorophenol	0.067	0.062	7.5	101	0.00
21	Phenanthrene	1.287	1.337	-3.9	108	0.00
22	Anthracene	1.064	1.048	1.5	100	0.00
23	Fluoranthene	1.303	1.311	-0.6	105	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
25	Pyrene	1.562	1.615	-3.4	104	0.00
26 S	Terphenyl-d14	0.689	0.716	-3.9	103	0.00
27	Benzo(a)anthracene	1.410	1.412	-0.1	102	0.00
28	Chrysene	1.384	1.392	-0.6	99	0.00
29	Indeno(1,2,3-cd)pyrene	1.262	1.123	11.0	87	0.00
30 I	Perylene-d12	1.000	1.000	0.0	93	-0.01
31	Benzo(b)fluoranthene	1.540	1.536	0.3	86	-0.01
32	Benzo(k)fluoranthene	1.442	1.544	-7.1	106	-0.01
33 C	Benzo(a)pyrene	1.299	1.325	-2.0	95	0.00
34	Dibenzo(a,h)anthracene	1.139	1.085	4.7	87	-0.01
35	Benzo(q,h,i)perylene	1.202	1.126	6.3	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0