

Data Path : Z:\HPCHEM1\BNA M\DATA\BM050115\
 Data File : BM001153.D
 Acq On : 02 May 2015 04:31
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Manual Integrations
 APPROVED

apatel
 5/5/2015 8:04:01 AM

Quant Time: May 04 17:09:06 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	15852	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	61797	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	35131	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	87097	5.00	ng	0.00
24) Chrysene-d12	21.15	240	76633	5.00	ng	0.00
30) Perylene-d12	23.29	264	66266	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.13	112	3110	1.01	ng	0.00
5) Phenol-d6	6.71	99	3783	1.00	ng	0.00
7) Nitrobenzene-d5	8.69	82	3458	0.99	ng	0.00
13) 2,4,6-Tribromophenol	15.69	330	1941	1.27	ng	0.00
14) 2-Fluorobiphenyl	12.81	172	11995	1.03	ng	0.00
26) Terphenyl-d14	19.59	244	10725	1.02	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.03	88	1893	1.07	ng	98
3) n-Nitrosodimethylamine	3.34	42	1784	0.92	ng	# 91
8) Nitrobenzene	8.73	77	3908	0.98	ng	100
9) Naphthalene	10.36	128	13720	1.01	ng	99
10) Hexachlorobutadiene	10.65	225	2649	1.01	ng	100
11) 2-Methylnaphthalene	11.99	142	9548	1.02	ng	99
15) Acenaphthylene	13.89	152	15750	1.03	ng	# 67
16) Acenaphthene	14.24	154	10880	1.03	ng	100
17) Fluorene	15.24	166	12788	1.04	ng	99
19) Hexachlorobenzene	16.26	284	4646	0.99	ng	# 95
20) Pentachlorophenol	16.61	266	1033	0.86	ng	100
21) Phenanthrene	16.98	178	22106	0.99	ng	99
22) Anthracene	17.08	178	18611	1.00	ng	99
23) Fluoranthene	19.01	202	22711	1.00	ng	99
25) Pyrene	19.38	202	23830	1.00	ng	100
27) Benzo(a)anthracene	21.13	228	21584	1.00	ng	97
28) Chrysene	21.18	228	21222	1.00	ng	97
29) Indeno(1,2,3-cd)pyrene	25.43	276	19789	1.02	ng	100
31) Benzo(b)fluoranthene	22.65	252	20194	0.99	ng	96
32) Benzo(k)fluoranthene	22.69	252	19985	1.05	ng	97
33) Benzo(a)pyrene	23.20	252	17439	1.01	ng	99
34) Dibenzo(a,h)anthracene	25.44	278	15404	1.02	ng	97
35) Benzo(g,h,i)perylene	26.09	276	16092	1.01	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA M\DATA\BM050115\
Data File : BM001153.D
Acq On : 02 May 2015 04:31
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM

Quant Time: May 04 17:09:06 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM050115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 04 16:48:07 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.53 min Scan# 314
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
152	100		
150	172.1	140.1	210.1
115	69.9	57.8	86.6

Manual Integrations
APPROVED




apatel
5/5/2015 8:04:01 AM




#2

1,4-Dioxane
Concen: 1.07 ng
RT: 3.03 min Scan# 23
Delta R.T. 0.02 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
88	100		
58	60.8	48.2	72.2
43	28.0	24.8	37.2





#3

n-Nitrosodimethylamine

Concen: 0.92 ng

RT: 3.34 min Scan# 43

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tot Ion:	42	Resp:	1784
Ion	Ratio	Lower	Upper
42	100		
74	98.2	72.0	108.0
44	7.1	3.7	5.5#

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#4

2-Fluorophenol

Concen: 1.01 ng




RT: 5.13 min Scan# 159


Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	112	Resp:	3110
Ion	Ratio	Lower	Upper
112	100		
64	64.7	52.6	79.0
63	36.4	29.5	44.3





#5

Phenol-d6

Concen: 1.00 ng

RT: 6.71 min Scan# 261

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M


ClientSampleId :

SSTDCCC001EC

Tot Ion:	99	Resp:	3783
Ion	Ratio	Lower	Upper
99	100		
42	25.5	20.9	31.3
71	35.7	28.5	42.7

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



Abundance Ion 99.00 (98.70 to 99.70): BM001142.D (-242) (-)

Ion 42.00 (41.70 to 42.70): BM001142.D (-242) (-)

Ion 71.00 (70.70 to 71.70): BM001142.D (-242) (-)

6.71

2500

2000



1500

1000

500

0

Time-->



#6

Naphthalene-d8

Concen: 5.00 ng


RT: 10.31 min Scan# 532

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	136	Resp:	61797
Ion	Ratio	Lower	Upper
136	100		
137	10.4	8.2	12.4
54	11.2	9.0	13.4



Abundance Ion 136.00 (135.70 to 136.70): BM001142.D (-528) (-)

Ion 137.00 (136.70 to 137.70): BM001142.D (-528) (-)

Ion 54.00 (53.70 to 54.70): BM001142.D (-528) (-)

10.31

40000


30000


20000

10000

0

Time-->





#7

Nitrobenzene-d5

Concen: 0.99 ng

RT: 8.69 min Scan# 402

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M


ClientSampleId :

SSTDCCC001EC

Tot Ion:	82	Resp:	3458
Ion	Ratio	Lower	Upper
82	100		
128	40.6	31.9	47.9
54	55.8	46.2	69.2

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM






Abundance Ion 82.00 (81.70 to 82.70): BM001142.D (-398) (-)

Ion 128.00 (127.70 to 128.70): BM001142.D (-398) (-)

Ion 54.00 (53.70 to 54.70): BM001142.D (-398) (-)

Time-->



#8

Nitrobenzene

Concen: 0.98 ng


RT: 8.73 min Scan# 406

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	77	Resp:	3908
Ion	Ratio	Lower	Upper
77	100		
123	39.1	31.0	46.6
65	13.1	10.6	15.8






Abundance Ion 77.00 (76.70 to 77.70): BM001142.D (-400) (-)

Ion 123.00 (122.70 to 123.70): BM001142.D (-400) (-)

Ion 65.00 (64.70 to 65.70): BM001142.D (-400) (-)

Time-->





#9

Naphthalene

Concen: 1.01 ng

RT: 10.36 min Scan# 535

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tot Ion:	128	Resp:	13720
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.0	13.6
127	13.0	10.0	15.0

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#10

Hexachlorobutadiene

Concen: 1.01 ng




RT: 10.65 min Scan# 552


Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	225	Resp:	2649
Ion Ratio	Lower	Upper	
225	100		
223	63.5	50.8	76.2
227	64.0	51.0	76.4









#11
2-Methylnaphthalene
Concen: 1.02 ng
RT: 11.99 min Scan# 661
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
142	100		
141	81.8	65.1	97.7
115	30.6	25.1	37.7




Manual Integrations
APPROVED


apatel
5/5/2015 8:04:01 AM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.20 min Scan# 838
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
164	100		
162	90.3	72.7	109.1
160	35.7	28.7	43.1









#13
 2,4,6-Tribromophenol
 Concen: 1.27 ng
 RT: 15.69 min Scan# 905
 Delta R.T. 0.00 min
 Lab File: BM001153.D
 Acq: 02 May 2015 04:31

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.6	74.5	111.7
141	0.0	21.0	31.4




Manual Integrations
 APPROVED


apatel
 5/5/2015 8:04:01 AM



#14
 2-Fluorobiphenyl
 Concen: 1.03 ng
 RT: 12.81 min Scan# 740
 Delta R.T. 0.00 min
 Lab File: BM001153.D
 Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
172	100		
171	32.1	25.8	38.6
170	19.6	16.1	24.1









#15
Acenaphthylene
Concen: 1.03 ng
RT: 13.89 min Scan# 824
Delta R.T. -0.02 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
152	100		
151	18.8	0.0	0.0#
153	0.0	10.3	15.5#




Manual Integrations
APPROVED


apatel
5/5/2015 8:04:01 AM



#16
Acenaphthene
Concen: 1.03 ng
RT: 14.24 min Scan# 840
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.3	90.6	136.0
152	56.0	45.0	67.4





#17

Fluorene

Concen: 1.04 ng

RT: 15.24 min Scan# 885

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tot Ion:	166	Resp:	12788
Ion	Ratio	Lower	Upper
166	100		
165	102.6	81.0	121.6
167	13.1	10.4	15.6

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 16.94 min Scan# 965


Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	188	Resp:	87097
Ion	Ratio	Lower	Upper
188	100		
94	11.0	9.3	13.9
80	11.7	9.8	14.6









#19
Hexachlorobenzene
Concen: 0.99 ng
RT: 16.26 min Scan# 932
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	0.0	0.0
249	35.9	26.5	39.7




Manual Integrations
APPROVED


apatel
5/5/2015 8:04:01 AM



#20
Pentachlorophenol
Concen: 0.86 ng
RT: 16.61 min Scan# 949
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.8	52.6	79.0
264	64.0	51.0	76.4





#21

Phenanthrene

Concen: 0.99 ng

RT: 16.98 min Scan# 967

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tot Ion:	178	Resp:	22106
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.5	23.3
179	15.3	12.0	18.0

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#22

Anthracene

Concen: 1.00 ng




RT: 17.08 min Scan# 972


Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	178	Resp:	18611
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.3	21.5
179	15.4	12.2	18.2





#23

Fluoranthene

Concen: 1.00 ng

RT: 19.01 min Scan# 1095

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M


ClientSampleId :

SSTDCCC001EC




Tot Ion:	202	Resp:	22711
Ion Ratio	Lower	Upper	
202	100		
101	11.5	9.3	13.9
203	17.4	13.7	20.5

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng


RT: 21.15 min Scan# 1279

Delta R.T. 0.00 min



Lab File: BM001153.D


Acq: 02 May 2015 04:31

Tgt Ion:	240	Resp:	76633
Ion Ratio	Lower	Upper	
240	100		
120	8.3	9.2	13.8#
236	26.6	22.4	33.6



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#25
Pvrene
Concen: 1.00 ng
RT: 19.38 min Scan# 1125
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC




Tot Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.4	24.6
203	17.4	13.8	20.6

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM






Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E




#26
Terphenyl-d14
Concen: 1.02 ng
RT: 19.59 min Scan# 1143
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.8	7.1	10.7
122	6.6	6.2	9.4

Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#27

Benzo(a)anthracene

Concen: 1.00 ng

RT: 21.13 min Scan# 1277

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tot Ion:	228	Resp:	21584
Ion Ratio	Lower	Upper	
228	100		
226	26.9	22.8	34.2
229	18.9	14.4	21.6

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#28

Chrysene

Concen: 1.00 ng




RT: 21.18 min Scan# 1282


Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Tgt Ion:	228	Resp:	21222
Ion Ratio	Lower	Upper	
228	100		
226	29.1	24.4	36.6
229	19.4	14.3	21.5





#29

Indeno(1,2,3-cd)pyrene

Concen: 1.02 ng

RT: 25.43 min Scan# 1690

Delta R.T. 0.00 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31

Instrument :

BNA_M

ClientSampleId :

SSTDCCC001EC

Tot Ion:276 Resp: 19789

Ion Ratio Lower Upper


276 100

138 30.5 24.2 36.4




277 24.7 19.8 29.6

Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E



#30

Perylene-d12

Concen: 5.00 ng

RT: 23.29 min Scan# 1485

Delta R.T. -0.01 min

Lab File: BM001153.D

Acq: 02 May 2015 04:31


Tgt Ion:264 Resp: 66266

Ion Ratio Lower Upper



264 100


260 25.7 18.9 28.3

265 20.8 18.1 27.1



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#31
Benzo(b)fluoranthene
Concen: 0.99 ng
RT: 22.65 min Scan# 1423
Delta R.T. -0.01 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.6	19.0	28.4
125	12.7	9.1	13.7




Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#32
Benzo(k)fluoranthene
Concen: 1.05 ng
RT: 22.69 min Scan# 1427
Delta R.T. -0.01 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	19.1	28.7
125	12.6	9.3	13.9









#33
Benzo(a)pyrene
Concen: 1.01 ng
RT: 23.20 min Scan# 1476
Delta R.T. 0.00 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.4	18.2	27.2
125	13.2	10.6	16.0




Manual Integrations
APPROVED

apatel
5/5/2015 8:04:01 AM



#34
Dibenzo(a,h)anthracene
Concen: 1.02 ng
RT: 25.44 min Scan# 1691
Delta R.T. -0.01 min
Lab File: BM001153.D
Acq: 02 May 2015 04:31

Tgt Ion	Ratio	Lower	Upper
278	100		
139	21.6	15.5	23.3
279	24.0	19.7	29.5





#35
 Benzo(a,h,i)perylene
 Concen: 1.01 ng
 RT: 26.09 min Scan# 1753
 Delta R.T. 0.00 min
 Lab File: BM001153.D
 Acq: 02 May 2015 04:31

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.0	28.4
138	25.8	21.3	31.9

Manual Integrations
 APPROVED

apatel
 5/5/2015 8:04:01 AM

