

Data File : BM009834.D
Acq On : 01 May 2017 17:32
Operator : SJ/MA
Sample : PB98594BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK94

Quant Time: May 02 00:53:05 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 02 00:52:56 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	54127	20.00	ng/ul	0.00
18) Naphthalene-d8	10.61	136	243510	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	174380	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.22	188	461175	20.00	ng/ul	0.00
75) Chrysene-d12	21.40	240	744999	20.00	ng/ul	0.00
83) Perylene-d12	23.72	264	830749	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.29	96	7551	6.42	ng/uL	0.00
5) Phenol-d5	6.99	99	169412	28.81	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.15	67	127310	31.75	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	122395	32.72	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	132730	28.96	ng/ul	0.00
19) Nitrobenzene-d5	8.98	128	63810	32.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.70	143	72078	34.57	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.23	165	114213	27.39	ng/ul	0.00
29) 4-Chloroaniline-d4	10.76	131	179995	35.82	ng/ul	0.00
43) Dimethylphthalate-d6	13.87	166	499190	32.91	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	558665	30.54	ng/ul	0.00
51) 4-Nitrophenol-d4	14.69	143	74450	25.44	ng/ul	0.00
57) Fluorene-d10	15.46	176	426174	31.12	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.60	200	77274	23.75	ng/ul	0.00
70) Anthracene-d10	17.32	188	743276	31.39	ng/ul	0.00
76) Pyrene-d10	19.61	212	977635	29.73	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.56	264	1245703	30.83	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM009834.D

Acq On : 01 May 2017 17:32

Operator : SJ/MA

Sample : PB98594BL

Misc :

ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

SBLK94

Quant Time: May 02 00:53:05 2017

Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M

Quant Title : SVOA CALIBRATION

QLast Update : Tue May 02 00:52:56 2017

Response via : Initial Calibration

