


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050117\
 Data File : BM009831.D
 Acq On : 01 May 2017 14:43
 Operator : SJ/MA
 Sample : DFTPP10
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP10

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-TEST.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 30 00:32:18 2016



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.4	66554	PASS
68	69	0.00	2	1.8	1205	PASS
69	198	0.00	100	58.1	68606	PASS
70	69	0.00	2	0.7	497	PASS
127	198	10	80	50.7	59818	PASS
197	198	0.00	2	0.4	464	PASS
198	198	100	100	100.0	118037	PASS
199	198	5	9	6.7	7858	PASS
275	198	10	60	27.0	31816	PASS
365	198	1	100	5.0	5845	PASS
441	443	0.01	100	76.5	13011	PASS
442	198	50	100	74.2	87610	PASS
443	442	15	24	19.4	17012	PASS

m/z	Abundance
37.10	524.0
37.90	1291.0
39.10	5496.0
39.90	1087.0
41.00	618.0
43.10	524.0
44.00	1262.0
49.00	673.0
50.00	21320.0
51.10	77672.0
52.00	4752.0
54.90	1355.0
56.00	2503.0
57.00	5083.0
60.90	812.0
62.00	1300.0
63.10	3741.0
63.90	504.0
65.00	1417.0
66.00	334.0
67.00	605.0
67.90	1462.0
69.00	74056.0
70.00	645.0
73.10	1783.0
74.00	9226.0
75.00	11140.0
76.10	4443.0
77.10	83488.0
78.10	5695.0
79.00	5596.0
80.00	4119.0
81.00	6602.0
82.10	1746.0
82.90	1551.0
83.90	440.0
85.10	1580.0
86.00	1984.0
86.90	727.0
88.10	444.0
91.00	1261.0
91.90	1252.0
93.00	9762.0
94.00	1029.0
95.00	856.0
97.10	413.0
98.00	7166.0
99.00	5756.0
99.90	357.0
100.90	3116.0
102.90	874.0
104.00	2276.0
105.00	1872.0
107.00	21544.0
108.10	4460.0
108.90	576.0
109.20	632.0
110.00	32976.0
111.00	6254.0
112.00	917.0
116.00	1000.0
117.00	21400.0
118.00	1747.0
119.00	336.0
120.10	472.0
121.10	561.0
121.80	941.0
122.20	1105.0
123.00	2889.0
124.00	1184.0
124.90	802.0
126.20	307.0
127.00	62696.0
127.90	5517.0
129.00	25768.0
130.00	1586.0
131.10	502.0
131.90	637.0
134.00	602.0
134.90	2542.0
135.80	1145.0
136.80	859.0
137.10	1095.0
137.70	581.0
138.90	311.0
141.00	4403.0
141.90	1407.0
142.70	782.0
143.00	870.0
143.80	354.0
144.90	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

146.10	579.0
147.10	2507.0
147.90	5149.0
148.90	1162.0
151.10	641.0
153.00	1251.0
154.10	820.0
155.10	2467.0
156.00	2151.0
157.30	729.0
157.90	563.0
159.00	535.0
159.90	1046.0
161.00	1203.0
162.10	583.0
162.90	647.0
164.80	1016.0
166.00	789.0
167.00	7721.0
167.90	4886.0
168.90	459.0
171.20	396.0
171.80	682.0
173.00	1117.0
174.00	1747.0
174.90	2548.0
176.00	458.0
177.00	1460.0
178.00	646.0
178.90	6443.0
180.00	3948.0
181.10	1818.0
182.10	349.0
183.70	364.0
185.00	2117.0
186.00	16456.0
187.00	5746.0
188.10	860.0
188.80	1431.0
190.90	1268.0
191.90	1863.0
193.10	2969.0
194.00	397.0
195.00	356.0
195.90	3262.0
197.00	608.0
197.90	112640.0
198.90	6108.0
200.10	353.0
201.40	542.0
202.80	631.0
204.00	4638.0
205.00	8351.0
206.00	31952.0
207.10	6453.0
207.90	1848.0
209.10	776.0
209.90	654.0
210.30	666.0
211.10	780.0
214.80	709.0
215.90	735.0
216.90	10030.0
217.80	943.0
221.00	4914.0
222.10	1541.0
223.00	2037.0
224.00	18704.0
225.00	5704.0
226.00	517.0
227.00	9320.0
228.10	1095.0
229.00	1422.0
231.00	963.0
233.90	930.0
235.00	608.0
236.10	693.0
237.00	844.0
238.90	467.0
240.00	498.0
241.10	500.0
242.00	1597.0
243.00	1438.0
244.00	14605.0
245.10	1727.0
246.00	3436.0
247.10	635.0
249.10	830.0
253.00	652.0
254.10	454.0
255.00	79800.0
256.00	10511.0
256.80	1079.0
258.00	5056.0
259.00	687.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

265.00	1201.0
266.00	436.0
272.90	2297.0
274.00	4755.0
275.00	28488.0
275.90	3920.0
276.90	3622.0
278.00	716.0
279.10	375.0
280.90	395.0
283.10	588.0
285.00	469.0
292.90	1166.0
295.00	497.0
296.00	8686.0
297.00	1284.0
303.00	894.0
304.00	316.0
313.80	864.0
314.90	1088.0
316.00	513.0
320.90	360.0
323.00	2273.0
324.00	438.0
324.90	363.0
327.00	961.0
327.80	713.0
332.90	393.0
334.00	1917.0
334.90	527.0
346.00	399.0
352.00	544.0
352.90	336.0
354.00	1322.0
364.90	4371.0
365.90	790.0
372.10	1250.0
401.90	574.0
403.00	678.0
421.00	765.0
422.00	537.0
423.00	3678.0
424.00	773.0
441.00	10412.0
442.10	68728.0
443.10	12436.0
444.00	1558.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
36.90	312.0
38.00	929.0
39.00	6963.0
40.10	1443.0
41.00	731.0
43.00	576.0
44.00	730.0
44.90	384.0
49.10	566.0
50.00	22056.0
51.00	77248.0
52.00	4348.0
55.10	1459.0
56.00	3004.0
57.10	5439.0
58.00	438.0
60.90	973.0
62.00	1670.0
63.10	2797.0
64.10	401.0
65.10	1864.0
66.90	393.0
67.20	395.0
68.00	1327.0
69.00	83104.0
69.90	355.0
71.10	352.0
73.00	2227.0
74.00	10646.0
75.00	12280.0
76.00	3533.0
77.10	92624.0
78.10	5736.0
79.00	6540.0
80.00	5124.0
81.00	6550.0
82.00	1695.0
83.10	1873.0
84.00	493.0
85.00	1345.0
86.00	1401.0
86.90	1637.0
88.10	562.0
89.00	379.0
91.00	1779.0
91.90	1362.0
93.00	11920.0
94.10	796.0
95.10	930.0
96.10	439.0
97.90	8617.0
99.00	5746.0
100.10	560.0
101.00	3178.0
102.90	1420.0
103.90	2534.0
105.00	2238.0
106.10	757.0
107.00	25640.0
108.00	3663.0
109.00	614.0
110.00	39368.0
111.00	5965.0
111.90	607.0
115.80	1034.0
117.00	28968.0
118.00	2291.0
120.00	312.0
121.80	1655.0
123.00	2687.0
123.90	1041.0
125.10	1451.0
126.10	381.0
127.00	71920.0
128.00	5408.0
128.90	29776.0
129.90	2655.0
131.00	707.0
133.00	708.0
134.00	1858.0
135.00	2927.0
136.00	1495.0
136.90	1516.0
137.90	731.0
139.90	597.0
141.00	5170.0
142.00	2024.0
142.90	1176.0
144.00	480.0
146.10	679.0
147.00	3484.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

148.00	6366.0
148.90	1750.0
150.00	761.0
151.00	485.0
151.70	421.0
153.10	1278.0
153.90	1321.0
155.00	2682.0
156.00	2707.0
157.00	1079.0
157.90	942.0
158.70	654.0
159.80	2115.0
160.90	2466.0
162.00	686.0
164.90	2387.0
166.00	1389.0
167.00	10975.0
167.90	5392.0
168.90	940.0
171.10	729.0
171.80	720.0
173.10	1361.0
173.90	1314.0
175.00	3712.0
176.10	1294.0
176.90	1898.0
177.90	662.0
178.90	6030.0
179.90	5089.0
181.00	2785.0
181.80	306.0
183.80	823.0
185.00	2964.0
186.00	22320.0
187.00	5029.0
188.10	1144.0
189.00	1653.0
190.10	327.0
190.90	1481.0
192.00	3028.0
193.10	3271.0
193.90	778.0
195.00	362.0
196.00	4003.0
196.90	786.0
197.90	140544.0
199.00	10294.0
200.00	668.0
201.20	563.0
202.70	837.0
204.00	6368.0
205.00	13004.0
206.10	40328.0
207.00	10706.0
208.00	2123.0
209.10	929.0
210.20	1017.0
211.10	1671.0
215.20	590.0
215.80	810.0
216.90	13636.0
218.00	1534.0
221.00	6113.0
221.90	2028.0
223.00	2588.0
224.00	23936.0
225.00	5265.0
227.00	11120.0
227.90	1304.0
229.00	2440.0
229.90	548.0
231.10	675.0
233.00	436.0
233.90	1126.0
234.90	1106.0
236.00	744.0
237.00	785.0
238.80	477.0
240.00	473.0
241.00	921.0
242.00	2033.0
243.00	1489.0
244.00	23904.0
245.10	3009.0
246.00	4261.0
246.80	951.0
248.70	632.0
252.80	707.0
253.80	323.0
255.00	106080.0
256.00	11659.0
256.80	1005.0
257.90	6382.0
259.00	516.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

265.00	2286.0
265.90	447.0
273.00	3069.0
274.00	8422.0
275.00	38968.0
276.00	6580.0
276.90	3410.0
277.90	612.0
281.10	760.0
283.00	608.0
283.90	327.0
284.90	714.0
290.00	328.0
292.90	771.0
293.80	327.0
295.10	354.0
295.90	11726.0
297.00	1998.0
303.10	1787.0
313.90	677.0
314.80	2325.0
316.10	1042.0
320.80	524.0
321.90	352.0
323.10	4035.0
324.10	949.0
326.80	791.0
332.10	368.0
334.00	2145.0
335.00	959.0
341.00	550.0
345.90	1501.0
346.90	444.0
352.00	1468.0
353.10	719.0
354.10	800.0
365.00	7476.0
366.00	668.0
371.90	1568.0
373.00	599.0
383.00	392.0
402.00	467.0
403.10	1144.0
404.10	334.0
421.00	1044.0
422.00	923.0
423.00	5200.0
424.00	1313.0
441.10	16232.0
442.00	102112.0
443.10	21040.0
444.00	1928.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
38.20	807.0
39.00	4518.0
40.00	1071.0
41.00	1036.0
43.10	539.0
44.00	1430.0
45.00	740.0
50.00	12158.0
51.10	44744.0
52.00	2762.0
55.00	777.0
56.00	1313.0
57.00	4173.0
60.90	773.0
62.00	544.0
63.00	3131.0
64.00	433.0
65.20	1057.0
67.00	524.0
67.90	826.0
69.00	50704.0
70.00	491.0
71.10	343.0
73.00	2654.0
74.10	4433.0
75.00	7523.0
76.00	3092.0
77.10	54848.0
78.00	3891.0
79.00	4676.0
80.00	2935.0
81.00	4725.0
81.90	1314.0
83.10	1440.0
85.00	919.0
86.00	1342.0
88.80	338.0
91.00	1064.0
92.00	2101.0
93.00	8718.0
94.10	799.0
95.10	616.0
96.00	726.0
98.00	5019.0
99.00	4165.0
100.00	540.0
101.00	2369.0
103.00	485.0
104.00	2220.0
105.00	1153.0
107.00	18960.0
107.90	1937.0
108.90	636.0
110.00	27512.0
111.00	3916.0
111.80	343.0
116.00	1531.0
117.00	18032.0
118.00	1210.0
119.90	460.0
122.00	1432.0
123.10	1731.0
124.10	1194.0
125.00	840.0
127.00	44840.0
127.90	2995.0
129.00	20592.0
130.10	2209.0
130.80	422.0
132.80	575.0
134.00	580.0
135.00	2093.0
135.90	664.0
136.80	1083.0
137.10	1138.0
139.00	505.0
140.00	377.0
141.00	3209.0
142.00	839.0
143.00	1124.0
145.70	535.0
146.10	392.0
147.00	2182.0
148.10	4919.0
148.90	1252.0
150.80	621.0
151.90	783.0
153.10	1126.0
153.90	714.0
154.90	1507.0
156.00	3715.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

156.90	619.0
158.00	481.0
158.80	866.0
159.90	1234.0
160.90	1382.0
162.00	974.0
162.90	466.0
165.00	1143.0
165.80	881.0
166.20	546.0
166.90	6981.0
168.00	3888.0
169.00	647.0
171.90	524.0
173.00	397.0
174.10	1167.0
175.00	2359.0
176.00	931.0
176.90	1762.0
177.60	451.0
178.00	491.0
179.00	5223.0
180.00	2926.0
180.90	2044.0
181.80	309.0
182.70	512.0
183.80	457.0
185.00	3056.0
186.00	14185.0
186.90	4733.0
188.00	322.0
188.90	1446.0
189.90	377.0
190.90	896.0
192.10	1418.0
193.10	2518.0
194.00	664.0
195.00	701.0
196.00	3272.0
196.70	309.0
197.90	100928.0
199.00	7172.0
200.00	751.0
201.20	360.0
202.80	1425.0
204.00	4949.0
205.00	7012.0
206.00	28200.0
207.00	6591.0
207.90	2327.0
208.90	854.0
210.10	738.0
210.80	1377.0
211.80	303.0
215.00	395.0
216.00	918.0
216.90	10201.0
217.90	748.0
221.10	3414.0
221.70	2196.0
223.00	1888.0
224.00	19184.0
225.00	4094.0
226.00	453.0
227.00	7195.0
227.90	792.0
228.90	1661.0
230.80	553.0
235.00	1199.0
235.90	689.0
236.80	592.0
239.10	341.0
239.70	329.0
241.00	909.0
242.10	1477.0
243.00	1763.0
244.00	14276.0
244.90	1440.0
246.00	3767.0
247.00	941.0
247.70	561.0
249.00	761.0
252.00	355.0
252.80	385.0
254.10	501.0
255.00	70504.0
256.00	10440.0
257.00	431.0
257.90	6784.0
259.20	487.0
265.00	1659.0
265.70	449.0
271.10	351.0
272.90	2430.0
274.00	5909.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

275.00	27992.0
276.00	4173.0
277.00	2760.0
277.80	649.0
281.00	1227.0
282.00	387.0
283.90	359.0
285.10	698.0
293.00	901.0
293.80	354.0
295.20	380.0
296.00	10452.0
296.90	1896.0
303.10	1794.0
314.00	602.0
314.90	1479.0
315.90	435.0
320.80	624.0
323.00	3733.0
324.00	675.0
327.00	413.0
334.00	2192.0
341.10	644.0
346.00	1029.0
351.80	810.0
353.10	398.0
354.10	1007.0
355.00	646.0
365.00	5690.0
365.80	930.0
371.90	1866.0
372.90	454.0
390.00	345.0
401.90	358.0
403.10	735.0
421.00	388.0
422.10	878.0
423.00	4461.0
424.10	2157.0
425.10	365.0
441.10	12390.0
442.10	91992.0
443.00	17560.0
444.10	831.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
39.90	719.0
41.00	838.0
43.00	690.0
44.00	726.0
55.00	572.0
56.10	324.0
57.00	791.0
67.10	512.0
69.10	682.0
73.10	1117.0
74.00	322.0
75.10	525.0
77.10	313.0
81.20	409.0
83.10	363.0
85.10	404.0
95.00	475.0
107.00	501.0
118.90	308.0
133.00	423.0
147.00	362.0
163.00	388.0
167.00	344.0
191.20	306.0
193.00	412.0
207.00	2377.0
208.00	478.0
209.00	523.0
221.00	377.0
280.90	388.0
282.00	417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10