




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050125\
Data File : BM050061.D
Acq On : 01 May 2025 12:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 01 13:23:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration



TIC: BM050061.D\data.ms

(70) Pentachlorophenol (C)




16.804min (-0.017) 298828.62 ng

response 984778

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.00	64.86
264.00	62.00	62.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050125\
Data File : BM050061.D
Acq On : 01 May 2025 12:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 01 13:23:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration



TIC: BM050061.D\data.ms

(77) Benzidine

19.392min (-0.012) 77464.20 ng

response 3260623

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.41
183.00	11.10	11.16
0.00	0.00	0.00