


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM050219\
 Data File : BM020100.D
 Acq On : 02 May 2019 09:01
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM043019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 01 03:58:37 2019



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.9	39418	PASS
68	69	0.00	2	1.4	927	PASS
69	198	0.00	100	51.3	65309	PASS
70	69	0.00	2	0.7	468	PASS
127	198	10	80	43.9	55933	PASS
197	198	0.00	2	0.7	908	PASS
198	198	100	100	100.0	127381	PASS
199	198	5	9	6.9	8790	PASS
275	198	10	60	30.3	38581	PASS
365	198	1	100	4.7	6050	PASS
441	198	0.01	100	13.2	16802	PASS
442	198	50	100	86.8	110530	PASS
443	442	15	24	19.1	21138	PASS