

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050223\
 Data File : BM039800.D
 Acq On : 02 May 2023 13:53
 Operator : CG/JU
 Sample : PB152414BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK414

Quant Time: May 03 04:15:35 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 02 18:41:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	3731	0.400	ng/u1	-0.02
4) Naphthalene-d8	10.584	136	13281	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.446	164	7062	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.200	188	14063	0.400	ng/u1	0.00
17) Chrysene-d12	21.411	240	9798	0.400	ng/u1	# 0.00
23) Perylene-d12	23.752	264	10284	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.180	96	31349	5.701	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.206	152	5783	0.339	ng/u1	0.00
18) Fluoranthene-d10	19.239	212	10261	0.389	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050223\
Data File : BM039800.D
Acq On : 02 May 2023 13:53
Operator : CG/JU
Sample : PB152414BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK414

Quant Time: May 03 04:15:35 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 02 18:41:11 2023
Response via : Initial Calibration

