

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050322\
 Data File : BM034910.D
 Acq On : 04 May 2022 03:19
 Operator : CG/JU
 Sample : N2562-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW469

Quant Time: May 04 04:11:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 02:59:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	5404	0.400	ng/ul	0.00
4) Naphthalene-d8	10.732	136	14636	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.556	164	8016	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.296	188	15001	0.400	ng/ul #	0.00
17) Chrysene-d12	21.473	240	10291	0.400	ng/ul #	0.00
23) Perylene-d12	23.856	264	10462	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	16970	2.839	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.316	152	4028	0.196	ng/ul	0.00
18) Fluoranthene-d10	19.322	212	7376	0.227	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.781	128	3436	0.082	ng/ul	94
7) 2-Methylnaphthalene	12.387	142	2708	0.097	ng/ul	99
8) 1-Methylnaphthalene	12.607	142	2606	0.095	ng/ul	99
15) Phenanthrene	17.338	178	1692	0.034	ng/ul#	89
19) Fluoranthene	19.354	202	1010	0.020	ng/ul#	70
20) Pyrene	19.712	202	2214	0.045	ng/ul#	78
22) Chrysene	21.508	228	1751	0.041	ng/ul#	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050322\
Data File : BM034910.D
Acq On : 04 May 2022 03:19
Operator : CG/JU
Sample : N2562-08
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW469

Quant Time: May 04 04:11:58 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 02:59:36 2022
Response via : Initial Calibration

