


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050322\
 Data File : BM034885.D
 Acq On : 03 May 2022 11:24
 Operator : CG/JU
 Sample : DFTPP053
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP053

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 04 02:59:36 2022



AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	206368	PASS
68	69	0.00	2	1.8	3780	PASS
69	198	0.00	100	36.7	207596	PASS
70	69	0.00	2	0.5	1087	PASS
127	198	10	80	46.8	264789	PASS
197	198	0.00	2	0.5	2671	PASS
198	198	100	100	100.0	565717	PASS
199	198	5	9	6.8	38597	PASS
275	198	10	60	28.8	163141	PASS
365	198	1	100	4.0	22899	PASS
441	198	0.01	100	14.6	82469	PASS
442	198	50	100	93.3	527680	PASS
443	442	15	24	19.3	102048	PASS

m/z	Abundance
36.10	188.0
39.10	636.0
40.00	7578.0
41.10	665.0
42.00	410.0
43.00	1482.0
44.00	2898.0
45.00	690.0
47.00	161.0
49.00	383.0
51.00	275.0
52.90	180.0
55.10	888.0
56.00	417.0
57.10	350.0
59.90	385.0
67.20	241.0
67.90	175.0
69.10	513.0
70.10	234.0
71.10	376.0
73.10	879.0
74.10	162.0
74.90	282.0
77.00	408.0
79.00	252.0
80.80	177.0
81.10	168.0
82.00	199.0
82.90	351.0
84.00	449.0
85.10	167.0
85.90	333.0
91.00	218.0
93.00	192.0
95.10	171.0
95.90	392.0
96.90	245.0
97.90	183.0
104.80	255.0
106.90	167.0
114.90	150.0
119.00	192.0
132.90	245.0
134.90	152.0
139.10	156.0
147.00	243.0
148.90	228.0
160.90	151.0
166.00	294.0
167.00	907.0
168.00	208.0
177.00	170.0
190.90	374.0
192.90	333.0
207.00	4045.0
208.00	783.0
209.00	597.0
266.90	161.0
281.00	680.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

m/z	Abundance
35.90	299.0
37.10	1039.0
38.10	3649.0
39.10	20672.0
40.00	8287.0
41.10	1393.0
41.90	337.0
43.00	1360.0
44.00	2439.0
45.00	938.0
46.90	242.0
47.90	247.0
49.00	2237.0
50.10	68240.0
51.00	262016.0
52.00	13353.0
53.10	796.0
55.00	1577.0
56.00	7752.0
57.00	17432.0
58.00	860.0
58.80	193.0
60.10	319.0
61.00	3275.0
62.00	3604.0
63.10	10453.0
64.10	1462.0
65.00	5072.0
66.00	397.0
67.00	489.0
68.00	4717.0
69.00	260864.0
70.00	1676.0
70.90	254.0
72.10	343.0
73.10	3345.0
74.00	28856.0
75.00	41104.0
76.00	14581.0
77.10	283776.0
78.10	20048.0
79.00	20736.0
80.00	15431.0
81.00	21992.0
82.00	5687.0
83.00	4172.0
84.00	882.0
85.00	3855.0
86.00	6328.0
87.00	2534.0
88.00	1166.0
89.00	643.0
91.00	5153.0
92.00	5530.0
93.00	37096.0
94.00	2543.0
95.00	438.0
96.10	1650.0
97.00	435.0
98.00	28144.0
99.00	21256.0
100.00	1634.0
101.00	11919.0
102.00	721.0
103.00	3922.0
104.00	7442.0
105.00	7483.0
106.00	2584.0
107.00	93104.0
108.00	13301.0
109.10	2406.0
110.00	159616.0
111.00	25264.0
112.00	3411.0
113.00	1081.0
114.00	221.0
115.00	586.0
116.00	4973.0
117.00	83096.0
118.00	5924.0
118.90	651.0
120.00	965.0
120.90	456.0
122.00	6129.0
123.00	9468.0
124.00	4665.0
125.00	4005.0
126.10	1644.0
127.00	319040.0
128.00	23576.0
129.00	127640.0
130.00	11068.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

131.10	2006.0
132.00	1008.0
133.00	961.0
134.00	3762.0
135.00	9924.0
136.00	4053.0
137.00	6312.0
137.90	1138.0
139.00	905.0
140.00	1513.0
141.00	16872.0
142.00	5720.0
143.00	4263.0
144.00	1072.0
144.90	955.0
146.00	3227.0
147.00	9415.0
148.00	21096.0
149.00	4579.0
150.00	1120.0
151.10	2445.0
152.00	1135.0
153.00	5369.0
154.00	3618.0
155.00	9228.0
156.00	14694.0
157.00	2336.0
157.90	3251.0
159.00	2378.0
160.00	5477.0
161.00	8710.0
162.00	2456.0
162.90	854.0
163.90	747.0
165.00	7092.0
166.00	5158.0
167.00	35248.0
168.00	19968.0
169.00	2884.0
170.00	1275.0
171.00	1475.0
172.00	3066.0
173.00	4070.0
174.00	7616.0
175.00	14077.0
176.00	3606.0
177.00	6857.0
178.00	2131.0
179.00	25456.0
180.00	18584.0
181.00	8718.0
182.00	1542.0
182.80	443.0
183.10	433.0
184.00	2158.0
185.00	12360.0
186.00	98392.0
187.00	28720.0
188.00	3515.0
189.00	6680.0
190.00	847.0
191.00	3135.0
192.00	8511.0
193.00	9681.0
194.00	1919.0
195.00	691.0
196.00	18984.0
197.00	3115.0
198.00	632064.0
199.00	43952.0
200.00	3614.0
201.40	2724.0
203.00	4836.0
204.00	25792.0
205.00	42600.0
206.00	176832.0
207.00	27152.0
208.00	6741.0
209.00	2451.0
210.00	2365.0
211.00	6727.0
213.00	532.0
214.10	247.0
215.00	2284.0
216.00	3759.0
217.00	48376.0
218.00	6145.0
219.00	684.0
221.00	24392.0
221.90	6598.0
223.00	11160.0
224.00	98192.0
225.00	24672.0
226.00	2885.0
227.00	44768.0
228.00	6649.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

229.00	8662.0
230.00	1356.0
231.00	3578.0
232.00	489.0
233.00	960.0
234.00	2584.0
235.00	2791.0
236.00	1897.0
237.00	4414.0
238.00	679.0
239.00	2017.0
240.00	1452.0
241.00	2493.0
242.00	5419.0
243.00	5254.0
244.00	74624.0
245.00	8964.0
246.00	15803.0
247.00	2918.0
248.00	939.0
249.00	2902.0
249.90	470.0
251.00	639.0
252.00	717.0
253.00	1582.0
254.10	3009.0
255.00	392960.0
256.00	57568.0
257.00	5674.0
258.00	24136.0
259.00	3603.0
259.90	855.0
260.90	702.0
263.00	275.0
264.10	524.0
265.00	10023.0
266.10	1412.0
266.90	364.0
267.80	162.0
269.00	226.0
269.90	434.0
271.00	1113.0
272.10	1110.0
273.00	12410.0
274.00	30376.0
275.00	170688.0
276.00	22768.0
277.00	15210.0
278.00	2754.0
279.10	510.0
279.90	151.0
281.00	1100.0
282.00	713.0
283.00	1810.0
284.00	1001.0
285.00	2510.0
286.00	570.0
289.00	521.0
289.90	315.0
291.00	329.0
292.00	876.0
293.00	3550.0
294.00	982.0
295.00	1460.0
296.00	53264.0
297.00	6778.0
298.00	440.0
300.90	789.0
301.90	1061.0
303.00	5471.0
304.00	1371.0
305.00	338.0
308.00	700.0
308.90	453.0
310.10	575.0
311.00	201.0
313.00	443.0
314.00	2575.0
315.00	5392.0
316.00	3230.0
317.00	518.0
320.00	199.0
321.10	1442.0
322.00	731.0
323.10	14843.0
324.00	2971.0
324.90	381.0
325.90	350.0
327.00	3176.0
328.00	1764.0
329.00	339.0
332.00	1097.0
333.00	1237.0
334.00	10185.0
335.00	2490.0
336.00	324.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

339.00	299.0
340.00	268.0
341.00	1951.0
342.10	406.0
346.00	3374.0
347.10	669.0
350.10	152.0
350.90	302.0
352.00	4498.0
353.00	3004.0
354.10	4828.0
355.00	1181.0
356.00	213.0
359.00	247.0
364.20	237.0
365.00	21728.0
366.00	2616.0
367.10	273.0
370.00	575.0
371.00	993.0
372.00	7423.0
373.10	1682.0
374.00	257.0
377.00	209.0
383.00	1995.0
384.00	419.0
385.00	191.0
390.00	820.0
391.00	749.0
392.00	428.0
401.00	415.0
402.00	2897.0
403.00	4140.0
404.10	1475.0
404.90	392.0
410.00	189.0
415.00	246.0
420.10	155.0
421.00	3476.0
422.00	3486.0
423.10	22840.0
424.10	6058.0
425.00	1002.0
440.10	224.0
441.10	71144.0
442.10	466944.0
443.10	91528.0
444.10	8676.0
445.00	689.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

m/z	Abundance
36.00	360.0
37.00	1261.0
38.10	3228.0
39.10	19472.0
40.00	8763.0
41.00	1123.0
42.10	290.0
43.00	1307.0
44.00	2625.0
45.00	1118.0
47.20	274.0
49.00	1919.0
50.10	61960.0
51.10	238336.0
52.00	12688.0
53.00	599.0
53.90	158.0
55.00	1782.0
56.00	7413.0
57.00	15700.0
58.00	797.0
59.00	213.0
60.00	496.0
61.00	2966.0
62.00	3691.0
63.10	8878.0
64.10	1580.0
65.10	4489.0
66.10	440.0
68.10	4572.0
69.00	242816.0
70.10	1406.0
71.00	264.0
73.00	3070.0
74.00	25000.0
75.00	39992.0
76.10	13127.0
77.10	272448.0
78.00	17536.0
79.00	20072.0
80.00	13513.0
81.00	19760.0
82.00	4748.0
83.00	4446.0
84.00	966.0
85.00	3844.0
86.00	6328.0
87.10	2275.0
88.00	1024.0
89.00	639.0
91.00	5202.0
92.00	4969.0
93.00	33616.0
94.00	1840.0
94.90	533.0
96.00	1842.0
96.90	608.0
98.00	26832.0
99.00	19768.0
100.00	2060.0
101.00	11096.0
101.90	483.0
103.00	3504.0
104.00	7180.0
105.00	7308.0
106.10	2103.0
107.00	85832.0
108.00	12849.0
109.00	2392.0
110.00	152640.0
111.00	24048.0
112.00	3254.0
113.00	1204.0
113.80	253.0
115.00	538.0
116.00	4574.0
117.00	81760.0
118.00	5382.0
119.00	789.0
120.00	1255.0
120.90	560.0
122.00	6010.0
123.00	9535.0
124.00	4447.0
125.00	3834.0
126.10	1135.0
127.00	307072.0
128.00	23104.0
129.00	124072.0
130.00	9776.0
131.00	1727.0
132.00	1080.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

133.00	741.0
134.00	3686.0
135.00	10534.0
136.00	3788.0
137.00	5465.0
138.00	1200.0
138.90	662.0
140.00	1994.0
141.00	17184.0
142.00	5070.0
143.00	3371.0
144.00	1220.0
145.00	1067.0
146.00	2427.0
147.00	8802.0
148.00	20520.0
149.00	4477.0
150.00	1325.0
151.00	2386.0
151.90	1491.0
153.00	6015.0
154.00	4182.0
155.00	10268.0
156.00	14017.0
157.00	2557.0
158.00	3831.0
159.00	2921.0
160.00	6266.0
161.00	7659.0
162.00	2598.0
162.90	840.0
164.00	795.0
165.00	6704.0
166.00	4912.0
167.00	34584.0
168.00	18872.0
169.00	3135.0
170.00	1150.0
170.90	1342.0
171.90	3748.0
173.00	4360.0
174.00	6962.0
175.00	14258.0
176.00	3740.0
177.00	6446.0
178.00	2393.0
179.00	27600.0
180.00	18840.0
181.00	8565.0
182.00	1293.0
183.00	851.0
184.00	2433.0
185.00	13227.0
186.00	100992.0
187.00	29288.0
188.00	3046.0
189.00	6837.0
189.90	1192.0
191.00	3512.0
192.00	8181.0
193.00	10179.0
194.00	2264.0
195.00	1105.0
196.00	18776.0
197.00	3062.0
198.00	680960.0
199.00	46112.0
200.00	3678.0
201.40	3219.0
203.00	5342.0
204.00	27968.0
205.00	45088.0
206.00	193152.0
207.00	27024.0
208.00	7671.0
209.00	2999.0
210.10	3121.0
211.00	7552.0
213.00	592.0
215.00	2289.0
216.00	4427.0
217.00	53160.0
218.00	6678.0
219.00	685.0
221.00	27512.0
222.00	6080.0
223.00	12037.0
224.00	108440.0
225.00	28728.0
226.00	2806.0
227.00	48304.0
228.00	6591.0
229.00	9665.0
230.00	1332.0
231.00	3643.0
232.00	603.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

233.00	1032.0
234.00	3549.0
235.00	3517.0
235.90	2302.0
237.00	4087.0
238.00	561.0
239.00	2399.0
240.00	1507.0
241.00	2773.0
242.10	5963.0
243.10	5948.0
244.00	82064.0
245.10	9965.0
246.00	18064.0
247.00	3874.0
248.00	739.0
249.00	3071.0
250.00	613.0
251.00	731.0
251.90	779.0
253.00	1653.0
254.10	2821.0
255.00	443968.0
256.00	63296.0
257.00	5087.0
258.00	27304.0
259.00	4333.0
260.00	823.0
261.00	602.0
262.90	268.0
264.00	630.0
265.00	11350.0
265.90	1811.0
266.90	450.0
267.90	196.0
269.00	200.0
269.90	379.0
271.10	949.0
272.00	981.0
273.00	14223.0
274.00	32792.0
275.00	198784.0
276.00	24848.0
277.00	17464.0
278.00	3181.0
279.00	651.0
281.00	1128.0
282.10	519.0
283.00	2170.0
284.00	1605.0
285.00	3310.0
286.00	656.0
288.00	274.0
288.90	825.0
289.90	595.0
290.90	453.0
292.00	898.0
293.00	3922.0
294.00	1036.0
295.00	1186.0
296.00	61576.0
297.00	9018.0
297.90	729.0
301.00	597.0
302.00	986.0
303.00	6563.0
304.00	2004.0
305.10	400.0
308.00	898.0
308.90	411.0
309.90	664.0
311.00	221.0
311.90	234.0
313.00	501.0
314.00	2660.0
315.00	7105.0
316.00	3929.0
317.00	550.0
319.90	201.0
321.00	2083.0
321.90	1092.0
323.10	17288.0
324.00	3196.0
325.00	326.0
325.90	463.0
327.00	4075.0
328.00	1816.0
329.00	290.0
332.00	1301.0
333.10	1580.0
334.10	12004.0
335.10	3336.0
336.00	420.0
339.00	298.0
340.00	291.0
341.00	2275.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

342.00	653.0
346.00	4216.0
347.00	938.0
350.00	205.0
351.10	306.0
352.00	5172.0
353.00	3708.0
354.10	5752.0
355.10	1380.0
356.00	220.0
359.10	390.0
363.90	288.0
365.00	27688.0
366.00	3967.0
367.00	394.0
370.00	417.0
371.00	1320.0
372.10	9161.0
373.10	2362.0
373.90	212.0
377.00	199.0
383.00	2495.0
383.90	875.0
384.90	235.0
390.00	1314.0
391.00	771.0
392.10	711.0
401.00	704.0
402.00	4300.0
403.10	5475.0
404.00	1731.0
404.90	302.0
415.00	383.0
421.00	4662.0
422.00	4193.0
423.10	31032.0
424.10	7398.0
425.10	860.0
440.10	173.0
441.10	98192.0
442.10	633792.0
443.10	123096.0
444.10	10866.0
445.00	684.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

m/z	Abundance
36.00	241.0
37.00	509.0
38.00	1509.0
39.10	9418.0
40.00	8034.0
41.10	1019.0
42.00	466.0
43.00	1279.0
44.00	2976.0
45.00	930.0
47.00	156.0
47.80	170.0
49.00	1390.0
50.00	31640.0
51.00	119576.0
52.00	6011.0
53.00	317.0
55.00	1096.0
56.00	3687.0
57.00	8504.0
58.00	412.0
58.90	161.0
59.90	329.0
61.00	1811.0
62.10	1799.0
63.00	5339.0
64.00	772.0
65.10	2577.0
66.00	256.0
67.00	218.0
68.10	2576.0
69.00	120648.0
70.00	880.0
71.10	195.0
72.00	186.0
73.10	1867.0
74.00	12511.0
75.00	20248.0
76.10	6192.0
77.10	133760.0
78.10	9068.0
79.00	9797.0
80.00	7427.0
81.00	9761.0
82.00	2670.0
83.00	2109.0
83.90	630.0
85.00	2019.0
86.00	3536.0
87.00	1340.0
87.90	410.0
89.10	219.0
91.00	2329.0
92.00	2810.0
93.00	18360.0
94.00	1401.0
95.10	520.0
96.00	1000.0
97.10	490.0
98.00	13295.0
99.00	10455.0
100.00	1005.0
101.00	5318.0
101.90	362.0
103.00	2235.0
104.00	4495.0
105.00	3753.0
106.00	1110.0
107.00	47672.0
108.00	7666.0
109.10	1252.0
110.00	80240.0
111.00	13324.0
112.00	1772.0
113.00	592.0
114.10	193.0
114.80	277.0
116.00	2403.0
117.00	43672.0
118.00	3266.0
118.90	482.0
119.90	553.0
121.00	313.0
122.00	2934.0
123.00	4891.0
124.00	2361.0
125.00	2322.0
126.10	779.0
127.00	168256.0
128.00	11089.0
129.00	66680.0
130.00	5839.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

131.00	1359.0
131.90	492.0
132.80	575.0
134.10	1912.0
135.00	5957.0
136.00	2387.0
137.00	2984.0
137.90	686.0
139.00	351.0
139.80	864.0
141.00	9121.0
142.00	3192.0
143.00	1756.0
144.00	755.0
144.90	538.0
146.00	1870.0
147.00	5674.0
148.00	12465.0
149.00	2831.0
150.00	796.0
151.00	1209.0
152.10	589.0
153.00	3276.0
154.00	2497.0
155.00	5320.0
156.00	8845.0
157.10	1371.0
158.00	1776.0
159.00	1068.0
160.00	2924.0
161.00	3910.0
162.00	1267.0
162.90	370.0
163.90	733.0
165.00	3714.0
166.00	3129.0
167.00	20088.0
168.00	11027.0
169.00	1986.0
170.00	736.0
171.00	986.0
172.00	2067.0
173.00	2219.0
174.00	4286.0
175.00	7452.0
175.90	2612.0
177.00	4104.0
178.00	1333.0
179.00	15896.0
180.00	10334.0
181.00	5021.0
182.00	899.0
183.10	364.0
183.90	1325.0
185.00	7595.0
186.00	56328.0
187.00	16992.0
188.00	1685.0
189.00	3603.0
190.00	768.0
191.00	2523.0
192.00	5248.0
193.00	5548.0
194.00	1247.0
194.90	578.0
196.00	10911.0
197.00	1836.0
198.00	384128.0
199.00	25728.0
200.00	2003.0
201.50	1632.0
203.00	3411.0
204.00	15851.0
205.00	26368.0
206.00	109936.0
207.00	17472.0
208.00	5056.0
209.00	1808.0
210.20	1859.0
211.00	4275.0
211.80	796.0
213.00	456.0
215.00	1419.0
216.00	2328.0
217.00	29792.0
218.00	3927.0
219.00	380.0
221.00	14978.0
221.80	4099.0
223.00	7017.0
224.00	63128.0
225.00	15812.0
226.00	1622.0
227.00	29264.0
228.00	4164.0
229.00	5374.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

230.00	796.0
231.00	2497.0
232.00	484.0
233.00	553.0
234.00	1751.0
235.00	2184.0
236.00	1192.0
237.00	2448.0
237.90	281.0
239.00	1208.0
240.00	789.0
241.00	1479.0
242.00	3250.0
243.00	3506.0
244.00	48496.0
245.00	5993.0
246.00	10983.0
247.00	2155.0
248.00	535.0
249.00	1924.0
250.00	349.0
250.90	658.0
252.00	416.0
253.00	1129.0
254.10	1734.0
255.00	272832.0
256.00	38208.0
257.00	3288.0
258.00	16056.0
259.00	2789.0
260.00	473.0
261.00	467.0
262.90	182.0
264.00	460.0
265.00	7530.0
266.00	834.0
267.00	349.0
270.00	332.0
270.90	708.0
272.00	620.0
273.00	8235.0
274.00	21352.0
275.00	119952.0
276.00	15312.0
277.00	12045.0
278.00	1950.0
279.00	362.0
281.00	949.0
282.00	627.0
283.00	1321.0
284.00	1019.0
285.00	1755.0
286.00	411.0
288.00	184.0
288.90	516.0
290.00	365.0
290.90	173.0
292.00	525.0
293.00	2769.0
293.90	729.0
295.00	618.0
296.00	39008.0
297.00	5532.0
298.00	330.0
300.80	393.0
302.00	690.0
303.00	4437.0
304.00	1077.0
308.00	571.0
309.00	378.0
310.10	432.0
312.90	270.0
314.10	2007.0
315.00	4449.0
316.10	2236.0
316.90	339.0
321.00	1301.0
322.10	680.0
323.10	12153.0
324.00	2556.0
325.00	260.0
325.80	292.0
327.00	2324.0
327.90	1116.0
329.00	281.0
332.00	955.0
333.00	1285.0
334.00	8741.0
335.00	2104.0
335.90	265.0
339.00	191.0
340.00	215.0
341.00	1545.0
342.10	306.0
346.00	3031.0
347.00	559.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053

349.90	219.0
350.90	160.0
352.00	4015.0
353.00	2853.0
354.10	3643.0
355.00	830.0
358.90	275.0
365.00	19280.0
366.00	2712.0
366.90	169.0
370.00	448.0
371.00	1097.0
372.10	6900.0
373.00	1745.0
374.00	221.0
383.00	1868.0
384.00	655.0
385.10	268.0
390.00	956.0
391.10	494.0
392.00	501.0
400.90	561.0
402.00	2758.0
403.00	4434.0
404.00	1304.0
405.20	175.0
415.00	205.0
421.00	3661.0
422.00	3467.0
423.00	23352.0
424.00	5352.0
425.10	521.0
440.10	234.0
441.10	78072.0
442.10	482304.0
443.10	91520.0
444.10	8720.0
444.90	455.0

Instrument :
BNA_M
ClientSampleId :
DFTPP053