

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050323\
 Data File : BM039829.D
 Acq On : 03 May 2023 18:27
 Operator : CG/JU
 Sample : SST02053
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020015

Quant Time: May 04 00:28:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 23:57:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.772	152	93095	20.000	ng/u1	0.00
20) Naphthalene-d8	10.577	136	391141	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.436	164	236420	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.189	188	483918	20.000	ng/u1	0.00
79) Chrysene-d12	21.394	240	457376	20.000	ng/u1	0.00
88) Perylene-d12	23.747	264	494085	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.172	96	23598	9.730	ng/uL	0.00
4) Pyridine-d5	3.601	84	147580	21.656	ng/u1	0.00
7) Phenol-d5	6.960	99	175287	20.572	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.107	67	119697	21.254	ng/u1	0.00
11) 2-Chlorophenol-d4	7.307	132	146218	21.876	ng/u1	0.00
15) 4-Methylphenol-d8	8.507	113	136036	19.820	ng/u1	0.00
21) Nitrobenzene-d5	8.948	128	69131	21.969	ng/u1	0.00
24) 2-Nitrophenol-d4	9.672	143	79476	22.492	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.219	165	135368	22.264	ng/u1	0.00
31) 4-Chloroaniline-d4	10.736	131	179475	19.764	ng/u1	0.00
46) Dimethylphthalate-d6	13.854	166	406647	21.327	ng/u1	0.00
49) Acenaphthylene-d8	14.130	160	471249	22.000	ng/u1	0.00
54) 4-Nitrophenol-d4	14.689	143	42681	14.791	ng/u1	0.00
60) Fluorene-d10	15.430	176	338486	21.668	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.589	200	65862	20.481	ng/u1	0.00
73) Anthracene-d10	17.289	188	517772	22.248	ng/u1	0.00
81) Pyrene-d10	19.594	212	592886	19.508	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.594	264	577318	21.642	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.207	88	23705	8.931	ng/uL	100
5) Pyridine	3.619	79	154942	21.563	ng/u1	100
6) Benzaldehyde	6.925	77	73404	20.164	ng/u1	100
8) Phenol	6.989	94	181893	20.504	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.201	93	150583	20.359	ng/u1	100
12) 2-Chlorophenol	7.336	128	147134	21.877	ng/u1	100
13) 2-Methylphenol	8.236	108	140329	20.366	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.313	45	230686	22.455	ng/u1	100
16) Acetophenone	8.607	105	225691	20.019	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.589	70	129024	20.678	ng/u1	100
18) 4-Methylphenol	8.572	108	153022	20.570	ng/u1	100
19) Hexachloroethane	8.842	117	68344	22.374	ng/u1	100
22) Nitrobenzene	8.989	77	184469	23.742	ng/u1	100
23) Isophorone	9.513	82	362659	21.888	ng/u1	100
25) 2-Nitrophenol	9.701	139	82634	22.413	ng/u1	100
26) 2,4-Dimethylphenol	9.766	107	170331	22.608	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.001	93	207689	21.317	ng/u1	100
29) 2,4-Dichlorophenol	10.242	162	132515	22.126	ng/u1	100
30) Naphthalene	10.630	128	481609	22.254	ng/u1	100
32) 4-Chloroaniline	10.760	127	184002	20.564	ng/u1	100
33) Hexachlorobutadiene	10.901	225	89674	22.723	ng/u1	100
34) Caprolactam	11.554	113	42874	19.223	ng/u1	100
35) 4-Chloro-3-methylphenol	11.901	107	145641	20.827	ng/u1	100
36) 2-Methylnaphthalene	12.248	142	306290	21.112	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050323\
 Data File : BM039829.D
 Acq On : 03 May 2023 18:27
 Operator : CG/JU
 Sample : SST02053
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020015

Quant Time: May 04 00:28:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 23:57:35 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.466	142	307874	21.117	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.618	216	160556	22.625	ng/ul	100
40) Hexachlorocyclopentadiene	12.583	237	29991	7.093	ng/ul	100
41) 2,4,6-Trichlorophenol	12.877	196	102728	21.669	ng/ul	100
42) 2,4,5-Trichlorophenol	12.960	196	107880	21.894	ng/ul	100
43) 1,1'-Biphenyl	13.265	154	407074	22.280	ng/ul	100
44) 2-Chloronaphthalene	13.313	162	320057	22.331	ng/ul	100
45) 2-Nitroaniline	13.536	65	96640	23.365	ng/ul	100
47) Dimethylphthalate	13.901	163	407161	21.568	ng/ul	100
48) 2,6-Dinitrotoluene	14.036	165	81663	21.389	ng/ul	100
50) Acenaphthylene	14.160	152	511244	22.313	ng/ul	100
51) 3-Nitroaniline	14.371	138	60980	17.040	ng/ul	100
52) Acenaphthene	14.501	153	348104	22.066	ng/ul	100
53) 2,4-Dinitrophenol	14.601	184	32154	15.821	ng/ul	100
55) 4-Nitrophenol	14.707	109	34093	15.554	ng/ul	100
56) Dibenzofuran	14.836	168	478756	22.328	ng/ul	100
57) 2,4-Dinitrotoluene	14.836	165	116020	21.870	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.077	232	92404	20.951	ng/ul	100
59) Diethylphthalate	15.265	149	419034	21.413	ng/ul	100
61) Fluorene	15.489	166	397147	22.391	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.483	204	192092	21.850	ng/ul	100
63) 4-Nitroaniline	15.542	138	47282	17.229	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.601	198	64614	20.899	ng/ul	100
67) N-Nitrosodiphenylamine	15.701	169	325075	22.547	ng/ul	100
68) 4-Bromophenyl-phenylether	16.377	248	113144	21.648	ng/ul	100
69) Hexachlorobenzene	16.495	284	130933	21.806	ng/ul	100
70) Atrazine	16.659	200	116494	21.775	ng/ul	100
71) Pentachlorophenol	16.853	266	64227	18.298	ng/ul	100
72) Phenanthrene	17.236	178	600272	22.343	ng/ul	100
74) Anthracene	17.324	178	619286	22.977	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.230	216	164470	22.821	ng/ul	100
76) Pentachlorobenzene	14.759	250	160342	22.603	ng/ul	100
77) Carbazole	17.606	167	541657	22.616	ng/ul	100
78) Di-n-butylphthalate	18.159	149	730042	21.551	ng/ul	100
80) Fluoranthene	19.259	202	698454	19.607	ng/ul	100
82) Pyrene	19.624	202	735208	19.936	ng/ul	100
83) Butylbenzylphthalate	20.524	149	341343	19.609	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.318	252	238822	23.547	ng/ul	100
85) Benzo(a)anthracene	21.377	228	699823	21.155	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.300	149	507172	19.592	ng/ul	100
87) Chrysene	21.430	228	680226	21.556	ng/ul	100
89) Di-n-octyl phthalate	22.206	149	875703	17.640	ng/ul	100
90) Benzo(b)fluoranthene	23.030	252	703683	20.241	ng/ul	100
91) Benzo(k)fluoranthene	23.077	252	718932	21.057	ng/ul	100
93) Benzo(a)pyrene	23.647	252	626052	21.364	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.188	276	817273	25.356	ng/ul	100
95) Dibenzo(a,h)anthracene	26.194	278	680354	25.458	ng/ul	100
96) Benzo(g,h,i)perylene	26.941	276	664602	25.883	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050323\
 Data File : BM039829.D
 Acq On : 03 May 2023 18:27
 Operator : CG/JU
 Sample : SSTD02053
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020015

Quant Time: May 04 00:28:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 23:57:35 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.931 ng/uL
RT: 3.207 min Scan# 29
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion: 88 Resp: 23705
Ion Ratio Lower Upper
88 100
43 34.7 27.8 41.6
58 74.6 59.7 89.5



#5
Pyridine
Concen: 21.563 ng/uL
RT: 3.619 min Scan# 99
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 79 Resp: 154942
Ion Ratio Lower Upper
79 100
52 77.3 61.8 92.8
51 39.6 31.7 47.5









#6
Benzaldehyde
Concen: 20.164 ng/ul
RT: 6.925 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion: 77 Resp: 73404
Ion Ratio Lower Upper
77 100
105 91.6 73.3 109.9
106 90.1 72.1 108.1



#8
Phenol
Concen: 20.504 ng/ul
RT: 6.989 min Scan# 672
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 94 Resp: 181893
Ion Ratio Lower Upper
94 100
65 30.3 24.2 36.4
66 36.9 29.5 44.3









#10
Bis(2-Chloroethyl)ether
Concen: 20.359 ng/ul
RT: 7.201 min Scan# 708
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion: 93 Resp: 150583
Ion Ratio Lower Upper
93 100
63 81.4 65.1 97.7
95 32.0 25.6 38.4



#12
2-Chlorophenol
Concen: 21.877 ng/ul
RT: 7.336 min Scan# 731
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 128 Resp: 147134
Ion Ratio Lower Upper
128 100
64 52.7 42.2 63.2
130 32.4 25.9 38.9





#13

2-Methylphenol

Concen: 20.366 ng/ul

RT: 8.236 min Scan# 884

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :





SSTD020015

Tgt Ion:108 Resp: 140329

Ion Ratio Lower Upper

108 100

107 87.8 70.2 105.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.455 ng/ul

RT: 8.313 min Scan# 897

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion: 45 Resp: 230686


Ion Ratio Lower Upper

45 100

77 15.0 12.0 18.0

79 12.0 9.6 14.4









#16
Acetophenone
Concen: 20.019 ng/ul
RT: 8.607 min Scan# 947
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:105 Resp: 225691
Ion Ratio Lower Upper
105 100
77 86.0 68.8 103.2
51 31.5 25.2 37.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.678 ng/ul
RT: 8.589 min Scan# 944
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 70 Resp: 129024
Ion Ratio Lower Upper
70 100
42 61.0 48.8 73.2
101 9.5 7.6 11.4
130 19.0 15.2 22.8









#18
4-Methylphenol
Concen: 20.570 ng/ul
RT: 8.572 min Scan# 941
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:108 Resp: 153022
Ion Ratio Lower Upper
108 100
107 111.7 89.4 134.0



#19
Hexachloroethane
Concen: 22.374 ng/ul
RT: 8.842 min Scan# 987
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:117 Resp: 68344
Ion Ratio Lower Upper
117 100
201 95.2 76.2 114.2
199 58.9 47.1 70.7









#22
Nitrobenzene
Concen: 23.742 ng/ul
RT: 8.989 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion: 77 Resp: 184469
Ion Ratio Lower Upper
77 100
123 40.8 32.6 49.0
65 13.6 10.9 16.3



#23
Isophorone
Concen: 21.888 ng/ul
RT: 9.513 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 82 Resp: 362659
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 16.1 12.9 19.3









#25
2-Nitrophenol
Concen: 22.413 ng/ul
RT: 9.701 min Scan# 1133
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:139 Resp: 82634
Ion Ratio Lower Upper
139 100
65 57.0 45.6 68.4
109 31.8 25.4 38.2



#26
2,4-Dimethylphenol
Concen: 22.608 ng/ul
RT: 9.766 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:107 Resp: 170331
Ion Ratio Lower Upper
107 100
121 48.2 38.6 57.8
122 82.4 65.9 98.9





#27

Bis(2-Chloroethoxy)methane

Concen: 21.317 ng/ul

RT: 10.001 min Scan# 1184

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion: 93 Resp: 207689

Ion Ratio Lower Upper

93 100

95 31.5 25.2 37.8

123 11.0 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 22.126 ng/ul

RT: 10.242 min Scan# 1225

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:162 Resp: 132515


Ion Ratio Lower Upper

162 100

164 66.0 52.8 79.2

98 38.2 30.6 45.8









#30
Naphthalene
Concen: 22.254 ng/ul
RT: 10.630 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:128 Resp: 481609
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 20.564 ng/ul
RT: 10.760 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:127 Resp: 184002
Ion Ratio Lower Upper
127 100
129 32.4 25.9 38.9









#33
Hexachlorobutadiene
Concen: 22.723 ng/ul
RT: 10.901 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:225 Resp: 89674
Ion Ratio Lower Upper
225 100
223 63.6 50.9 76.3
227 62.3 49.8 74.8



#34
Caprolactam
Concen: 19.223 ng/ul
RT: 11.554 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:113 Resp: 42874
Ion Ratio Lower Upper
113 100
55 203.7 163.0 244.4
56 143.8 115.0 172.6









#35
4-Chloro-3-methylphenol
Concen: 20.827 ng/ul
RT: 11.901 min Scan# 1507
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:107 Resp: 145641
Ion Ratio Lower Upper
107 100
144 24.2 19.4 29.0
142 75.9 60.7 91.1



#36
2-Methylnaphthalene
Concen: 21.112 ng/ul
RT: 12.248 min Scan# 1566
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:142 Resp: 306290
Ion Ratio Lower Upper
142 100
141 88.3 70.6 106.0









#37
1-Methylnaphthalene
Concen: 21.117 ng/ul
RT: 12.466 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:142 Resp: 307874
Ion Ratio Lower Upper
142 100
141 92.4 73.9 110.9
116 4.1 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.625 ng/ul
RT: 12.618 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:216 Resp: 160556
Ion Ratio Lower Upper
216 100
214 78.1 62.5 93.7
179 20.5 16.4 24.6
108 19.6 15.7 23.5





#40

Hexachlorocyclopentadiene

Concen: 7.093 ng/ul

RT: 12.583 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:237 Resp: 29991

Ion Ratio Lower Upper

237 100

235 67.2 53.8 80.6

272 12.1 9.7 14.5



#41

2,4,6-Trichlorophenol

Concen: 21.669 ng/ul

RT: 12.877 min Scan# 1673

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:196 Resp: 102728


Ion Ratio Lower Upper

196 100

198 96.4 77.1 115.7

200 31.5 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 21.894 ng/ul
RT: 12.960 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:196 Resp: 107880
Ion Ratio Lower Upper
196 100
198 95.8 76.6 115.0
200 31.7 25.4 38.0



#43
1,1'-Biphenyl
Concen: 22.280 ng/ul
RT: 13.265 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:154 Resp: 407074
Ion Ratio Lower Upper
154 100
153 39.0 31.2 46.8
76 14.4 11.5 17.3









#44
2-Chloronaphthalene
Concen: 22.331 ng/ul
RT: 13.313 min Scan# 1747
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:162 Resp: 320057
Ion Ratio Lower Upper
162 100
164 32.8 26.2 39.4
127 39.8 31.8 47.8



#45
2-Nitroaniline
Concen: 23.365 ng/ul
RT: 13.536 min Scan# 1785
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion: 65 Resp: 96640
Ion Ratio Lower Upper
65 100
92 67.9 54.3 81.5
138 96.6 77.3 115.9









#47
Dimethylphthalate
Concen: 21.568 ng/ul
RT: 13.901 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:163 Resp: 407161
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 21.389 ng/ul
RT: 14.036 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:165 Resp: 81663
Ion Ratio Lower Upper
165 100
89 63.7 51.0 76.4
121 22.4 17.9 26.9





#50

Acenaphthylene

Concen: 22.313 ng/ul

RT: 14.160 min Scan# 1891

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:152 Resp: 511244

Ion Ratio Lower Upper

152 100

151 20.0 16.0 24.0

153 13.4 10.7 16.1



#51

3-Nitroaniline

Concen: 17.040 ng/ul

RT: 14.371 min Scan# 1927

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:138 Resp: 60980


Ion Ratio Lower Upper

138 100

108 11.0 8.8 13.2

92 120.8 96.6 145.0









#52
Acenaphthene
Concen: 22.066 ng/ul
RT: 14.501 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:153 Resp: 348104
Ion Ratio Lower Upper
153 100
152 46.9 37.5 56.3
154 88.0 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 15.821 ng/ul
RT: 14.601 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:184 Resp: 32154
Ion Ratio Lower Upper
184 100
63 80.1 64.1 96.1
154 65.3 52.2 78.4









#55
4-Nitrophenol
Concen: 15.554 ng/ul
RT: 14.707 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:109 Resp: 34093
Ion Ratio Lower Upper
109 100
139 123.3 98.6 148.0
65 138.9 111.1 166.7



#56
Dibenzofuran
Concen: 22.328 ng/ul
RT: 14.836 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:168 Resp: 478756
Ion Ratio Lower Upper
168 100
139 39.4 31.5 47.3









#57
2,4-Dinitrotoluene
Concen: 21.870 ng/ul
RT: 14.836 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:165 Resp: 116020
Ion Ratio Lower Upper
165 100
63 76.7 61.4 92.0
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.951 ng/ul
RT: 15.077 min Scan# 2047
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:232 Resp: 92404
Ion Ratio Lower Upper
232 100
131 48.9 39.1 58.7
130 2.8 2.2 3.4
166 31.5 25.2 37.8









#59
Diethylphthalate
Concen: 21.413 ng/ul
RT: 15.265 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:149 Resp: 419034
Ion Ratio Lower Upper
149 100
177 21.1 16.9 25.3
150 11.9 9.5 14.3



#61
Fluorene
Concen: 22.391 ng/ul
RT: 15.489 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:166 Resp: 397147
Ion Ratio Lower Upper
166 100
165 97.5 78.0 117.0
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 21.850 ng/ul
RT: 15.483 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:204 Resp: 192092
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 60.2 48.2 72.2



#63
4-Nitroaniline
Concen: 17.229 ng/ul
RT: 15.542 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:138 Resp: 47282
Ion Ratio Lower Upper
138 100
92 56.4 45.1 67.7
108 89.3 71.4 107.2









#66
4,6-Dinitro-2-methylphenol
Concen: 20.899 ng/ul
RT: 15.601 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:198 Resp: 64614
Ion Ratio Lower Upper
198 100
182 4.5 3.6 5.4
77 33.2 26.6 39.8



#67
N-Nitrosodiphenylamine
Concen: 22.547 ng/ul
RT: 15.701 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:169 Resp: 325075
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 37.3 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 21.648 ng/ul
RT: 16.377 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:248 Resp: 113144
Ion Ratio Lower Upper
248 100
250 96.8 77.4 116.2
141 80.9 64.7 97.1



#69
Hexachlorobenzene
Concen: 21.806 ng/ul
RT: 16.495 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:284 Resp: 130933
Ion Ratio Lower Upper
284 100
142 37.8 30.2 45.4
249 32.2 25.8 38.6





#70

Atrazine

Concen: 21.775 ng/ul

RT: 16.659 min Scan# 2316

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:200 Resp: 116494

Ion Ratio Lower Upper

200 100

173 25.5 20.4 30.6

215 47.1 37.7 56.5



#71

Pentachlorophenol

Concen: 18.298 ng/ul

RT: 16.853 min Scan# 2349

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:266 Resp: 64227


Ion Ratio Lower Upper

266 100

264 62.2 49.8 74.6

268 62.3 49.8 74.8





#72

Phenanthrene

Concen: 22.343 ng/ul

RT: 17.236 min Scan# 2414

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:178 Resp: 600272

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.1 15.3 22.9



#74

Anthracene

Concen: 22.977 ng/ul

RT: 17.324 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:178 Resp: 619286


Ion Ratio Lower Upper

178 100

179 15.1 12.1 18.1

176 18.6 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.821 ng/uL
RT: 13.230 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:216 Resp: 164470
Ion Ratio Lower Upper
216 100
214 77.6 62.1 93.1
218 48.9 39.1 58.7



#76
Pentachlorobenzene
Concen: 22.603 ng/uL
RT: 14.759 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:250 Resp: 160342
Ion Ratio Lower Upper
250 100
248 64.0 51.2 76.8
252 65.5 52.4 78.6





#77

Carbazole

Concen: 22.616 ng/ul

RT: 17.606 min Scan# 2477

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:167 Resp: 541657

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 13.5 10.8 16.2



#78

Di-n-butylphthalate

Concen: 21.551 ng/ul

RT: 18.159 min Scan# 2571

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:149 Resp: 730042


Ion Ratio Lower Upper

149 100

150 8.9 7.1 10.7

104 5.5 4.4 6.6









#80
Fluoranthene
Concen: 19.607 ng/u1
RT: 19.259 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27




Instrument :
BNA_M
ClientSampleId :
SSTD020015


Tgt Ion:202 Resp: 698454
Ion Ratio Lower Upper
202 100
101 12.0 9.6 14.4
100 9.1 7.3 10.9



#82
Pyrene
Concen: 19.936 ng/u1
RT: 19.624 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BM039829.D
Acq: 03 May 2023 18:27

Tgt Ion:202 Resp: 735208
Ion Ratio Lower Upper
202 100
101 14.3 11.4 17.2
100 11.9 9.5 14.3





#83

Butylbenzylphthalate

Concen: 19.609 ng/ul

RT: 20.524 min Scan# 2973

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:149 Resp: 341343

Ion Ratio Lower Upper

149 100

91 76.5 61.2 91.8

206 19.4 15.5 23.3



#84

3,3'-Dichlorobenzidine

Concen: 23.547 ng/ul

RT: 21.318 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:252 Resp: 238822


Ion Ratio Lower Upper

252 100

254 63.7 51.0 76.4

126 11.9 9.5 14.3





#85

Benzo(a)anthracene

Concen: 21.155 ng/ul

RT: 21.377 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:228 Resp: 699823

Ion Ratio Lower Upper

228 100

229 19.9 15.9 23.9

226 27.4 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.592 ng/ul

RT: 21.300 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:149 Resp: 507172


Ion Ratio Lower Upper

149 100

167 27.1 21.7 32.5

279 4.0 3.2 4.8





#87

Chrysene

Concen: 21.556 ng/ul

RT: 21.430 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:228 Resp: 680226

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.8 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 17.640 ng/ul




RT: 22.206 min Scan# 3259


Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Tgt Ion:149 Resp: 875703





#90

Benzo(b)fluoranthene

Concen: 20.241 ng/ul

RT: 23.030 min Scan# 3399

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:252 Resp: 703683

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 11.1 8.9 13.3



#91

Benzo(k)fluoranthene

Concen: 21.057 ng/ul

RT: 23.077 min Scan# 3407

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:252 Resp: 718932


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 11.0 8.8 13.2





#93

Benzo(a)pyrene

Concen: 21.364 ng/ul

RT: 23.647 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:252 Resp: 626052

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 11.2 9.0 13.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 25.356 ng/ul

RT: 26.188 min Scan# 3936

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27




Tgt Ion:276 Resp: 817273


Ion Ratio Lower Upper

276 100

138 26.9 21.5 32.3

277 25.2 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 25.458 ng/ul

RT: 26.194 min Scan# 3937

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Instrument :

BNA_M

ClientSampleId :

SSTD020015





Tgt Ion:278 Resp: 680354

Ion Ratio Lower Upper

278 100

139 18.1 14.5 21.7

279 23.5 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 25.883 ng/ul

RT: 26.941 min Scan# 4064

Delta R.T. 0.000 min

Lab File: BM039829.D

Acq: 03 May 2023 18:27

Tgt Ion:276 Resp: 664602

Ion Ratio Lower Upper

276 100

138 23.0 18.4 27.6

277 23.4 18.7 28.1

