


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050323\
 Data File : BM039826.D
 Acq On : 03 May 2023 13:49
 Operator : CG/JU
 Sample : DFTPP118
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP118

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM050323.MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu May 04 00:51:59 2023



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.8	78946	PASS
68	69	0.00	2	1.5	1226	PASS
69	198	0.00	100	39.6	82587	PASS
70	69	0.00	2	0.6	509	PASS
127	198	10	80	46.7	97427	PASS
197	198	0.00	2	0.4	917	PASS
198	198	100	100	100.0	208811	PASS
199	198	5	9	6.6	13754	PASS
275	198	10	60	26.3	54869	PASS
365	198	1	100	3.7	7670	PASS
441	198	0.01	100	13.6	28437	PASS
442	198	50	100	86.9	181376	PASS
443	442	15	24	19.1	34728	PASS

m/z	Abundance
39.00	279.0
39.90	954.0
41.00	412.0
43.00	447.0
44.00	3279.0
44.80	155.0
48.90	407.0
50.90	193.0
54.90	444.0
56.90	240.0
67.00	229.0
69.00	240.0
71.00	186.0
73.00	740.0
76.90	246.0
78.80	167.0
81.00	238.0
83.10	205.0
83.90	354.0
85.90	229.0
91.10	154.0
94.80	163.0
95.90	289.0
132.90	263.0
135.00	183.0
146.90	183.0
149.00	235.0
190.90	250.0
192.90	266.0
207.00	2390.0
207.90	500.0
209.00	408.0
266.80	202.0
280.90	354.0
282.00	157.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

m/z	Abundance
35.90	208.0
36.90	426.0
38.00	1481.0
39.00	7550.0
39.90	1150.0
41.00	590.0
42.00	171.0
43.00	609.0
44.00	3311.0
45.00	353.0
49.00	958.0
50.00	26832.0
51.00	103616.0
52.00	5148.0
53.00	256.0
55.00	670.0
55.90	2860.0
57.00	6965.0
58.00	306.0
58.90	236.0
61.00	1316.0
62.00	1467.0
63.00	4298.0
63.90	508.0
65.00	2144.0
65.90	275.0
66.90	284.0
68.00	1577.0
69.00	107576.0
70.00	646.0
70.90	176.0
73.00	1569.0
74.00	10740.0
75.00	16672.0
76.00	5692.0
77.00	122296.0
78.00	8878.0
79.00	7231.0
80.00	5622.0
81.00	7879.0
82.00	2296.0
82.90	2108.0
84.00	541.0
85.00	1691.0
86.00	2541.0
86.90	1019.0
87.80	349.0
89.00	273.0
91.00	1923.0
92.00	2152.0
93.00	13490.0
94.00	1035.0
95.10	383.0
95.90	759.0
97.00	468.0
98.00	9782.0
99.00	8256.0
100.00	715.0
101.00	4676.0
102.00	311.0
102.90	1499.0
103.90	2971.0
105.00	2775.0
105.90	745.0
107.00	35520.0
108.00	5395.0
109.00	910.0
110.00	61400.0
111.00	10564.0
112.00	1215.0
112.90	426.0
115.00	257.0
115.90	1736.0
117.00	29984.0
117.90	2298.0
118.90	301.0
120.00	549.0
120.90	189.0
121.90	2193.0
122.90	3623.0
123.90	1638.0
125.00	1591.0
126.00	405.0
127.00	120824.0
128.00	8998.0
129.00	46264.0
130.00	3961.0
130.90	850.0
132.00	394.0
133.00	299.0
134.00	1253.0
134.90	3618.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

136.00	1861.0
136.90	1972.0
137.90	431.0
138.90	272.0
139.90	513.0
140.90	6035.0
141.90	1924.0
142.90	1441.0
143.90	475.0
145.00	337.0
145.90	1037.0
147.00	3258.0
147.90	7292.0
148.90	1787.0
150.00	524.0
151.00	760.0
151.80	340.0
152.90	1860.0
153.90	1501.0
155.00	3648.0
156.00	4947.0
157.00	1083.0
157.90	1039.0
158.90	619.0
159.90	1912.0
160.90	2696.0
161.90	957.0
162.90	231.0
163.80	301.0
164.90	2272.0
165.90	1785.0
166.90	12429.0
167.90	7443.0
168.90	1036.0
169.80	293.0
170.90	587.0
171.90	1133.0
173.00	1425.0
174.00	2671.0
175.00	5160.0
175.90	1164.0
176.90	2148.0
177.90	989.0
178.90	9242.0
179.90	6293.0
180.90	2991.0
181.90	505.0
182.90	179.0
183.90	705.0
185.00	4226.0
186.00	33720.0
187.00	9452.0
187.90	1111.0
188.90	2103.0
189.80	366.0
190.90	1263.0
191.90	2890.0
193.00	3637.0
194.00	722.0
194.90	392.0
195.90	6357.0
196.80	1038.0
197.90	245312.0
198.90	16472.0
199.90	1218.0
201.50	978.0
202.90	1867.0
204.00	8918.0
205.00	15582.0
206.00	63568.0
207.00	10842.0
208.00	2308.0
209.00	1020.0
210.00	1114.0
210.90	2661.0
213.00	226.0
214.90	570.0
215.90	1318.0
216.90	17176.0
217.90	2172.0
219.10	227.0
219.90	190.0
221.00	10099.0
221.80	2596.0
222.90	3522.0
224.00	35416.0
225.00	9289.0
226.00	957.0
226.90	15125.0
227.90	2003.0
228.90	3256.0
229.90	450.0
231.00	1331.0
231.90	171.0
233.00	261.0
233.90	1027.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

235.00	1074.0
235.90	831.0
236.90	1314.0
237.90	180.0
238.80	603.0
239.90	455.0
240.90	913.0
241.90	1909.0
243.00	1980.0
244.00	27472.0
245.00	3753.0
245.90	5441.0
246.90	1097.0
247.80	273.0
249.00	1128.0
250.00	171.0
250.80	231.0
251.90	286.0
252.80	491.0
253.80	651.0
255.00	140480.0
256.00	20600.0
257.00	1474.0
258.00	8410.0
259.00	1369.0
259.90	310.0
261.00	210.0
263.90	178.0
264.90	3598.0
265.90	453.0
266.90	191.0
269.90	191.0
270.90	407.0
272.10	283.0
272.90	4416.0
273.90	10666.0
275.00	61032.0
276.00	8006.0
276.90	5665.0
278.00	954.0
279.00	157.0
281.00	681.0
282.00	272.0
283.00	660.0
284.00	397.0
285.00	1033.0
289.00	179.0
290.00	164.0
290.80	176.0
292.00	255.0
292.90	1371.0
294.00	346.0
294.90	399.0
296.00	17488.0
296.90	2375.0
297.80	151.0
300.90	253.0
302.00	370.0
302.90	1844.0
304.00	442.0
308.00	279.0
308.90	189.0
309.80	267.0
313.00	159.0
314.00	694.0
314.90	2176.0
316.00	1054.0
317.00	168.0
321.00	484.0
322.00	319.0
323.00	5396.0
324.10	1137.0
325.00	196.0
325.90	167.0
327.00	1039.0
327.90	508.0
332.00	350.0
332.80	584.0
334.00	3308.0
335.00	1135.0
341.00	567.0
341.90	237.0
345.90	1339.0
346.90	327.0
352.00	1475.0
353.00	1073.0
354.00	1830.0
355.00	410.0
364.90	8120.0
365.90	1121.0
369.80	157.0
371.00	493.0
372.00	2651.0
373.10	755.0
383.00	855.0
383.90	242.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

389.90	324.0
390.90	296.0
392.00	211.0
400.90	218.0
402.00	1022.0
402.90	1773.0
404.00	546.0
404.80	152.0
421.00	1087.0
422.00	1215.0
423.00	9523.0
424.00	2217.0
425.00	403.0
441.00	26864.0
442.00	168640.0
443.00	33688.0
444.00	3162.0
445.00	220.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

m/z	Abundance
36.90	385.0
38.00	1057.0
39.00	6725.0
40.00	1091.0
41.00	726.0
42.00	161.0
43.00	445.0
44.00	3085.0
45.00	357.0
49.00	928.0
50.00	22288.0
51.00	89608.0
52.00	4554.0
52.90	343.0
55.00	670.0
56.00	2876.0
57.00	6026.0
58.00	354.0
59.00	209.0
60.00	163.0
60.90	1100.0
62.00	1303.0
63.00	3640.0
64.00	652.0
65.00	1651.0
67.10	292.0
68.00	1377.0
69.00	95272.0
70.00	539.0
73.00	1379.0
74.00	9663.0
75.00	15559.0
76.00	5199.0
77.00	108496.0
78.00	7480.0
79.00	6724.0
80.00	4919.0
81.00	7829.0
81.90	1938.0
83.00	2190.0
84.00	563.0
85.00	1368.0
86.00	2290.0
86.90	969.0
88.00	439.0
91.00	1755.0
91.90	1703.0
93.00	12070.0
93.90	878.0
95.00	290.0
95.90	768.0
97.00	368.0
98.00	8688.0
99.00	7283.0
99.90	617.0
101.00	3803.0
101.90	231.0
103.00	1470.0
103.90	2785.0
104.90	2859.0
106.00	942.0
107.00	32304.0
108.00	4502.0
109.00	944.0
110.00	56944.0
111.00	8945.0
112.00	950.0
113.00	410.0
114.80	216.0
115.90	1717.0
116.90	28000.0
117.90	2067.0
119.00	313.0
119.90	348.0
121.90	2015.0
122.90	3457.0
123.90	1539.0
124.90	1470.0
126.00	264.0
127.00	112104.0
128.00	8137.0
128.90	42872.0
129.90	3844.0
131.00	663.0
131.90	554.0
132.90	300.0
134.00	1263.0
135.00	3682.0
135.90	1378.0
137.00	1832.0
137.90	382.0
138.90	283.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

139.80	495.0
140.90	5290.0
141.90	1699.0
142.90	1481.0
143.90	390.0
144.90	271.0
146.00	992.0
146.90	3170.0
148.00	7251.0
149.00	1509.0
150.00	460.0
151.10	790.0
151.80	476.0
152.90	1810.0
153.90	1484.0
154.90	3095.0
156.00	4541.0
156.90	988.0
157.90	1042.0
158.90	838.0
159.90	1760.0
161.00	2529.0
161.90	870.0
163.00	251.0
164.00	262.0
164.90	1905.0
165.90	1884.0
167.00	11725.0
167.90	6353.0
168.90	902.0
169.90	389.0
170.90	491.0
171.80	924.0
173.00	1441.0
173.90	2753.0
175.00	4661.0
176.00	1316.0
176.90	2103.0
178.00	700.0
178.90	9136.0
180.00	6399.0
180.90	2969.0
182.00	565.0
183.00	333.0
183.90	750.0
185.00	4383.0
186.00	32816.0
187.00	9282.0
188.00	1001.0
188.90	2157.0
189.80	340.0
191.00	1210.0
192.00	2706.0
193.00	3151.0
194.00	735.0
194.90	352.0
196.00	6542.0
196.80	1007.0
197.90	246528.0
198.90	16191.0
199.90	1130.0
201.40	1002.0
202.90	1477.0
204.00	9455.0
205.00	15299.0
206.00	63704.0
207.00	10208.0
208.00	2371.0
208.90	1008.0
210.10	1144.0
211.00	2912.0
214.90	804.0
215.90	1409.0
216.90	17880.0
218.00	2101.0
219.00	250.0
220.00	274.0
221.00	9459.0
221.70	2281.0
222.90	3700.0
224.00	37496.0
225.00	9310.0
226.00	901.0
227.00	15180.0
227.90	2235.0
228.90	2944.0
229.90	386.0
231.00	1414.0
232.00	202.0
233.00	317.0
234.00	928.0
234.90	1249.0
235.90	784.0
236.90	1454.0
237.80	194.0
239.00	675.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

240.00	449.0
240.90	784.0
242.00	2016.0
243.00	2178.0
244.00	28848.0
245.00	3496.0
245.90	5769.0
246.90	1103.0
247.90	204.0
248.90	960.0
250.90	225.0
252.00	297.0
253.00	582.0
254.00	734.0
255.00	147328.0
256.00	20288.0
256.90	1474.0
258.00	8226.0
258.90	1355.0
260.00	233.0
260.80	235.0
264.00	182.0
264.90	3600.0
265.90	665.0
266.90	238.0
269.90	248.0
270.90	295.0
271.90	260.0
272.90	4325.0
274.00	10948.0
275.00	64144.0
276.00	9173.0
276.90	5868.0
277.90	797.0
279.00	258.0
281.00	470.0
281.90	283.0
282.90	641.0
284.00	463.0
284.90	825.0
285.90	265.0
288.90	198.0
292.00	262.0
293.00	1410.0
293.90	200.0
294.90	388.0
296.00	18960.0
296.90	2505.0
298.00	201.0
300.90	270.0
302.00	418.0
303.00	2261.0
304.00	604.0
308.00	217.0
308.90	254.0
309.90	272.0
312.80	154.0
314.00	923.0
314.90	2193.0
315.90	1159.0
317.00	161.0
320.90	484.0
322.00	264.0
323.00	6033.0
323.90	1251.0
324.90	173.0
326.90	1244.0
327.90	682.0
331.90	535.0
333.00	524.0
334.00	4309.0
335.00	1204.0
340.90	776.0
342.00	252.0
345.90	1429.0
346.80	316.0
352.00	2091.0
352.90	1328.0
354.00	1836.0
355.00	511.0
358.90	196.0
365.00	9010.0
365.90	1489.0
370.00	204.0
371.00	479.0
372.00	3163.0
372.90	707.0
382.90	1056.0
383.80	245.0
390.00	497.0
391.00	240.0
400.80	271.0
402.00	1415.0
403.00	2079.0
404.00	720.0
421.00	1914.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

422.00	1727.0
423.00	10940.0
424.00	2726.0
424.80	280.0
441.00	33808.0
442.00	219008.0
443.00	42424.0
444.00	3634.0
444.90	201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

m/z	Abundance
36.90	276.0
38.00	681.0
39.00	3332.0
39.90	934.0
41.00	512.0
42.90	478.0
44.00	2890.0
45.00	253.0
49.00	557.0
50.00	11496.0
51.00	44192.0
52.00	2228.0
53.00	183.0
54.90	652.0
55.90	1299.0
57.00	3332.0
57.80	183.0
59.00	190.0
60.90	431.0
61.90	626.0
63.00	1924.0
64.00	221.0
65.00	979.0
67.00	296.0
68.00	725.0
69.00	45632.0
70.00	343.0
71.00	167.0
73.00	1168.0
74.00	4462.0
75.00	7227.0
76.00	2593.0
77.00	55256.0
78.00	3886.0
79.00	3711.0
79.90	2764.0
81.00	3801.0
81.90	1194.0
83.00	958.0
83.90	357.0
84.90	696.0
85.90	1214.0
87.00	543.0
87.90	212.0
88.90	151.0
90.90	1001.0
91.90	1003.0
93.00	6355.0
93.90	458.0
95.00	250.0
95.90	573.0
96.90	357.0
97.90	4524.0
98.90	3865.0
99.90	388.0
101.00	2207.0
101.90	158.0
103.00	783.0
103.90	1306.0
104.90	1687.0
106.00	326.0
107.00	16704.0
108.00	2513.0
108.90	551.0
110.00	28896.0
110.90	4526.0
111.90	578.0
112.90	187.0
116.00	783.0
117.00	15004.0
117.90	1156.0
118.90	286.0
120.00	167.0
121.90	1064.0
122.90	1854.0
123.90	982.0
124.90	691.0
126.00	184.0
127.00	59352.0
128.00	4289.0
128.90	22200.0
130.00	1726.0
130.90	530.0
131.80	188.0
132.90	288.0
133.90	558.0
134.90	2013.0
135.90	973.0
137.00	1054.0
138.00	235.0
138.90	180.0
139.90	413.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

140.90	3007.0
141.90	936.0
142.90	518.0
144.10	158.0
145.10	201.0
146.00	581.0
146.90	1778.0
147.90	3938.0
148.90	1108.0
149.80	270.0
151.00	363.0
151.70	230.0
153.00	997.0
153.90	884.0
154.90	1946.0
155.90	2552.0
157.00	386.0
157.90	512.0
158.90	440.0
159.90	865.0
161.00	1760.0
162.00	498.0
162.90	155.0
163.90	151.0
164.90	1130.0
165.90	733.0
166.90	6562.0
167.90	3952.0
168.90	704.0
170.00	197.0
171.00	286.0
171.90	578.0
172.90	834.0
173.90	1497.0
175.00	2642.0
175.90	622.0
176.90	1045.0
177.90	442.0
178.90	5209.0
180.00	3360.0
181.00	1555.0
181.90	274.0
182.90	160.0
183.90	345.0
185.00	2238.0
186.00	18728.0
186.90	4940.0
187.90	467.0
188.90	1098.0
190.00	261.0
190.90	836.0
191.90	1601.0
193.00	2056.0
193.90	549.0
194.90	341.0
196.00	3725.0
196.80	705.0
197.90	134592.0
198.90	8598.0
199.80	757.0
201.40	614.0
202.90	1037.0
204.00	5085.0
205.00	8363.0
206.00	35456.0
207.00	7050.0
207.90	1850.0
208.90	673.0
210.10	694.0
210.90	1520.0
214.80	353.0
215.90	790.0
216.90	9873.0
218.00	1362.0
221.00	5326.0
221.80	1425.0
222.90	2535.0
224.00	20552.0
225.00	5076.0
226.00	566.0
226.90	8249.0
228.00	1131.0
228.90	1757.0
229.90	183.0
230.90	688.0
234.00	529.0
234.90	609.0
235.90	469.0
237.00	901.0
238.90	297.0
240.00	270.0
240.80	419.0
242.00	1205.0
243.00	1048.0
244.00	16816.0
245.00	2170.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118

245.90	3309.0
247.00	579.0
247.90	154.0
248.90	519.0
250.00	152.0
250.80	166.0
252.90	300.0
253.90	571.0
255.00	86904.0
256.00	12666.0
257.00	923.0
257.90	4736.0
258.90	778.0
259.80	155.0
264.90	2167.0
266.00	301.0
270.90	258.0
272.00	196.0
272.90	2771.0
274.00	7391.0
275.00	39432.0
276.00	5022.0
276.90	3558.0
277.90	691.0
281.00	533.0
282.00	268.0
283.00	450.0
283.90	293.0
285.00	644.0
288.90	170.0
291.90	202.0
292.90	781.0
293.90	173.0
294.90	288.0
295.90	11770.0
296.90	1872.0
301.00	231.0
301.90	228.0
303.00	1378.0
303.90	327.0
308.00	282.0
310.00	207.0
314.00	606.0
314.90	1416.0
316.00	684.0
320.90	317.0
322.00	166.0
323.00	3778.0
324.00	625.0
326.90	813.0
327.90	395.0
331.80	312.0
333.00	436.0
334.00	2903.0
334.90	594.0
341.00	553.0
345.90	905.0
346.80	181.0
352.00	1254.0
353.00	886.0
354.00	1157.0
354.90	389.0
364.90	5881.0
365.90	898.0
370.00	184.0
371.00	284.0
372.00	2127.0
372.90	553.0
382.90	731.0
389.80	292.0
391.00	156.0
391.90	168.0
400.90	186.0
402.00	904.0
402.90	1417.0
404.00	369.0
421.00	1209.0
422.00	1089.0
423.00	7645.0
424.00	1615.0
425.00	185.0
441.00	24640.0
442.00	156480.0
443.00	28072.0
444.00	2723.0
444.90	157.0

Instrument :
BNA_M
ClientSampleId :
DFTPP118