

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050415\
 Data File : BM001203.D
 Acq On : 05 May 2015 02:48
 Operator : TP/IZ
 Sample : G2104-10
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EB-2015-04-30

Quant Time: May 05 04:59:03 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	13852	5.00	ng	0.00
6) Naphthalene-d8	10.31	136	55625	5.00	ng	0.00
12) Acenaphthene-d10	14.18	164	32256	5.00	ng	-0.02
18) Phenanthrene-d10	16.94	188	85621	5.00	ng	0.00
24) Chrysene-d12	21.15	240	69667	5.00	ng	0.00
30) Perylene-d12	23.28	264	53647	5.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.11	112	13128	4.87	ng	-0.01
5) Phenol-d6	6.72	99	10488	3.16	ng	0.02
7) Nitrobenzene-d5	8.68	82	25912	8.22	ng	-0.01
13) 2,4,6-Tribromophenol	15.69	330	23891	15.02	ng	0.00
14) 2-Fluorobiphenyl	12.80	172	85998	8.06	ng	-0.01
26) Terphenyl-d14	19.58	244	101602	10.58	ng	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) n-Nitrosodimethylamine	3.27	42	73	0.04	ng	# 1
8) Nitrobenzene	8.35	77	5208	1.38	ng	# 45
9) Naphthalene	10.36	128	313	0.03	ng	# 50
10) Hexachlorobutadiene	10.65	225	2	0.00	ng	# 1
11) 2-Methylnaphthalene	11.98	142	966	0.12	ng	# 84
15) Acenaphthylene	13.89	152	55	0.00	ng	# 67
16) Acenaphthene	14.18	154	158	0.02	ng	# 4
17) Fluorene	15.24	166	238	0.02	ng	# 88
19) Hexachlorobenzene	16.26	284	49	0.01	ng	# 42
20) Pentachlorophenol	16.61	266	23	0.38	ng	# 59
21) Phenanthrene	17.06	178	162	0.01	ng	# 97
22) Anthracene	17.06	178	162	0.01	ng	# 95
23) Fluoranthene	19.01	202	465	0.02	ng	# 99
25) Pyrene	19.38	202	370	0.02	ng	# 96
27) Benzo(a)anthracene	21.14	228	448	0.02	ng	# 79
28) Chrysene	21.18	228	199	0.01	ng	# 60
29) Indeno(1,2,3-cd)pyrene	25.43	276	172	0.01	ng	# 93
31) Benzo(b)fluoranthene	22.65	252	192	0.01	ng	# 1
32) Benzo(k)fluoranthene	22.69	252	154	0.01	ng	# 1
33) Benzo(a)pyrene	23.19	252	119	0.01	ng	# 1
34) Dibenzo(a,h)anthracene	25.44	278	200	0.02	ng	# 1
35) Benzo(g,h,i)perylene	26.08	276	136	0.01	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050415\
Data File : BM001203.D
Acq On : 05 May 2015 02:48
Operator : TP/IZ
Sample : G2104-10
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EB-2015-04-30

Quant Time: May 05 04:59:03 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 04 16:48:07 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.53 min Scan# 314

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion: 152 Resp: 13852

Ion Ratio Lower Upper

152 100

150 163.9 140.1 210.1




115 63.2 57.8 86.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine

Concen: 0.04 ng

RT: 3.27 min Scan# 39

Delta R.T. -0.06 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion: 42 Resp: 73

Ion Ratio Lower Upper

42 100

74 0.0 72.0 108.0#



44 104.1 3.7 5.5#




Abundance Ion 42.00 (41.70 to 42.70): BM0

Ion 74.00 (73.70 to 74.70): BM0

Ion 44.00 (43.70 to 44.70): BM0





#4

2-Fluorophenol

Concen: 4.87 ng

RT: 5.11 min Scan# 158

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion: 112 Resp: 13128

Ion Ratio Lower Upper

112 100

64 64.6 52.6 79.0




63 36.0 29.5 44.3



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): B

Ion 63.00 (62.70 to 63.70): B



#5

Phenol-d6

Concen: 3.16 ng

RT: 6.72 min Scan# 262

Delta R.T. 0.02 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion: 99 Resp: 10488

Ion Ratio Lower Upper

99 100

42 23.4 20.9 31.3



71 34.9 28.5 42.7




Abundance Ion 99.00 (98.70 to 99.70): B

Ion 42.00 (41.70 to 42.70): B

Ion 71.00 (70.70 to 71.70): B









#6
Naphthalene-d8
Concen: 5.00 ng
RT: 10.31 min Scan# 532
Delta R.T. 0.00 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48




Instrument :
BNA_M
ClientSampleId :
EB-2015-04-30


Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.2	12.4
54	8.3	9.0	13.4#



#7
Nitrobenzene-d5
Concen: 8.22 ng
RT: 8.68 min Scan# 401
Delta R.T. -0.01 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48

Tgt Ion	Ratio	Lower	Upper
82	100		
128	34.5	31.9	47.9
54	71.6	46.2	69.2#





#8

Nitrobenzene

Concen: 1.38 ng

RT: 8.35 min Scan# 369

Delta R.T. -0.38 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30





Tgt Ion: 77 Resp: 5208

Ion Ratio Lower Upper

77 100

123 0.0 31.0 46.6#

65 2.1 10.6 15.8#



#9

Naphthalene

Concen: 0.03 ng

RT: 10.36 min Scan# 535

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48




Tgt Ion: 128 Resp: 313


Ion Ratio Lower Upper

128 100

129 32.4 9.0 13.6#

127 30.6 10.0 15.0#









#10
Hexachlorobutadiene
Concen: 0.00 ng
RT: 10.65 min Scan# 552
Delta R.T. 0.00 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48




Instrument :
BNA_M
ClientSampleId :
EB-2015-04-30


Tgt Ion: 225 Resp: 2
Ion Ratio Lower Upper
225 100
223 200.0 50.8 76.2#
227 500.0 51.0 76.4#



#11
2-Methylnaphthalene
Concen: 0.12 ng
RT: 11.98 min Scan# 660
Delta R.T. -0.01 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48

Tgt Ion: 142 Resp: 966
Ion Ratio Lower Upper
142 100
141 93.4 65.1 97.7
115 44.7 25.1 37.7#





#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.18 min Scan# 837

Delta R.T. -0.02 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30





Tgt Ion:164 Resp: 32256

Ion Ratio Lower Upper

164 100

162 122.6 72.7 109.1#

160 63.6 28.7 43.1#



#13

2,4,6-Tribromophenol

Concen: 15.02 ng

RT: 15.69 min Scan# 905

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48




Tgt Ion:330 Resp: 23891


Ion Ratio Lower Upper

330 100

332 93.2 74.5 111.7

141 0.0 21.0 31.4#









#14
2-Fluorobiphenyl
Concen: 8.06 ng
RT: 12.80 min Scan# 739
Delta R.T. -0.01 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48




Instrument :
BNA_M
ClientSampleId :
EB-2015-04-30


Tgt Ion:172 Resp: 85998
Ion Ratio Lower Upper
172 100
171 35.3 25.8 38.6
170 23.7 16.1 24.1



#15
Acenaphthylene
Concen: 0.00 ng
RT: 13.89 min Scan# 824
Delta R.T. -0.02 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48

Tgt Ion:152 Resp: 55
Ion Ratio Lower Upper
152 100
151 0.0 0.0 0.0
153 0.0 10.3 15.5#









#16
Acenaphthene
Concen: 0.02 ng
RT: 14.18 min Scan# 837
Delta R.T. -0.07 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48




Instrument :
BNA_M
ClientSampleId :
EB-2015-04-30


Tgt Ion:154 Resp: 158
Ion Ratio Lower Upper
154 100
153 0.0 90.6 136.0#
152 0.0 45.0 67.4#



#17
Fluorene
Concen: 0.02 ng
RT: 15.24 min Scan# 885
Delta R.T. 0.00 min
Lab File: BM001203.D
Acq: 05 May 2015 02:48

Tgt Ion:166 Resp: 238
Ion Ratio Lower Upper
166 100
165 107.6 81.0 121.6
167 35.3 10.4 15.6#





#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.94 min Scan# 965

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion:188 Resp: 85621

Ion Ratio Lower Upper

188 100

94 7.8 9.3 13.9#




80 7.6 9.8 14.6#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0



#19

Hexachlorobenzene

Concen: 0.01 ng

RT: 16.26 min Scan# 932

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:284 Resp: 49

Ion Ratio Lower Upper

284 100

142 0.0 0.0 0.0



249 0.0 26.5 39.7#




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#20

Pentachlorophenol

Concen: 0.38 ng

RT: 16.61 min Scan# 949

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion:266 Resp: 23

Ion Ratio Lower Upper




266 100

268 100.0 52.6 79.0#

264 94.2 51.0 76.4#



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E



#21

Phenanthrene

Concen: 0.01 ng

RT: 17.06 min Scan# 971

Delta R.T. 0.08 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:178 Resp: 162

Ion Ratio Lower Upper



178 100


176 20.4 15.5 23.3

179 13.6 12.0 18.0



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#22

Anthracene

Concen: 0.01 ng

RT: 17.06 min Scan# 971

Delta R.T. -0.02 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion:178 Resp: 162

Ion Ratio Lower Upper

178 100

176 20.4 14.3 21.5

179 13.6 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E

800

600

400

200

0



17.00

17.06

17.10

17.20

Time-->



#23

Fluoranthene

Concen: 0.02 ng

RT: 19.01 min Scan# 1095

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:202 Resp: 465

Ion Ratio Lower Upper

202 100

101 11.4 9.3 13.9

203 16.8 13.7 20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E

400

300

200

100


0


18.90

19.01

19.10

Time-->





#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.15 min Scan# 1279

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion:240 Resp: 69667

Ion Ratio Lower Upper

240 100

120 4.8 9.2 13.8#




236 23.9 22.4 33.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#25

Pyrene

Concen: 0.02 ng

RT: 19.38 min Scan# 1125

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:202 Resp: 370

Ion Ratio Lower Upper

202 100

200 23.0 16.4 24.6



203 18.4 13.8 20.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#26
 Terphenyl-d14
 Concen: 10.58 ng
 RT: 19.58 min Scan# 1142
 Delta R.T. -0.01 min
 Lab File: BM001203.D
 Acq: 05 May 2015 02:48




Instrument :
 BNA_M
 ClientSampleId :
 EB-2015-04-30


Tgt Ion:244 Resp: 101602
 Ion Ratio Lower Upper
 244 100
 212 10.0 7.1 10.7
 122 17.8 6.2 9.4#



#27
 Benzo(a)anthracene
 Concen: 0.02 ng
 RT: 21.14 min Scan# 1278
 Delta R.T. 0.01 min
 Lab File: BM001203.D
 Acq: 05 May 2015 02:48

Tgt Ion:228 Resp: 448
 Ion Ratio Lower Upper
 228 100
 226 27.9 22.8 34.2
 229 41.6 14.4 21.6#





#28

Chrysene

Concen: 0.01 ng

RT: 21.18 min Scan# 1282

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion:228 Resp: 199

Ion Ratio Lower Upper

228 100

226 45.6 24.4 36.6#




229 45.1 14.3 21.5#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#29

Indeno(1,2,3-cd)pyrene

Concen: 0.01 ng

RT: 25.43 min Scan# 1690

Delta R.T. 0.00 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:276 Resp: 172

Ion Ratio Lower Upper

276 100

138 30.8 24.2 36.4



277 31.4 19.8 29.6#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#30

Perylene-d12

Concen: 5.00 ng

RT: 23.28 min Scan# 1484

Delta R.T. -0.02 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion: 264 Resp: 53647

Ion Ratio Lower Upper

264 100

260 26.6 18.9 28.3




265 20.8 18.1 27.1



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#31

Benzo(b)fluoranthene

Concen: 0.01 ng

RT: 22.65 min Scan# 1423

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion: 252 Resp: 192

Ion Ratio Lower Upper

252 100

253 102.0 19.0 28.4#



125 65.3 9.1 13.7#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#32

Benzo(k)fluoranthene

Concen: 0.01 ng

RT: 22.69 min Scan# 1427

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampled :

EB-2015-04-30


Tgt Ion:252 Resp: 154

Ion Ratio Lower Upper

252 100

253 104.8 19.1 28.7#




125 66.3 9.3 13.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#33

Benzo(a)pyrene

Concen: 0.01 ng

RT: 23.19 min Scan# 1475

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion:252 Resp: 119

Ion Ratio Lower Upper

252 100

253 126.5 18.2 27.2#



125 80.6 10.6 16.0#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#34

Dibenzo(a,h)anthracene

Concen: 0.02 ng

RT: 25.44 min Scan# 1691

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48

Instrument :

BNA_M

ClientSampleId :

EB-2015-04-30


Tgt Ion: 278 Resp: 200

Ion Ratio Lower Upper

278 100

139 84.6 15.5 23.3#




279 100.0 19.7 29.5#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#35

Benzo(g,h,i)perylene

Concen: 0.01 ng

RT: 26.08 min Scan# 1752

Delta R.T. -0.01 min

Lab File: BM001203.D

Acq: 05 May 2015 02:48


Tgt Ion: 276 Resp: 136

Ion Ratio Lower Upper

276 100

277 72.7 19.0 28.4#

138 80.7 21.3 31.9#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

