

Data Path : Z:\HPCHEM1\BNA_M\Data\BM050415\
 Data File : BM001186.D
 Acq On : 04 May 2015 10:52
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 05 02:43:47 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	128	0.00
2	1,4-Dioxane	1.000	0.883	11.7	108	0.00
3	n-Nitrosodimethylamine	1.000	1.016	-1.6	124	0.00
4 S	2-Fluorophenol	1.000	1.043	-4.3	132	-0.01
5 S	Phenol-d6	1.000	1.079	-7.9	137	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	132	0.00
7 S	Nitrobenzene-d5	1.000	1.019	-1.9	135	0.00
8	Nitrobenzene	1.000	0.959	4.1	127	0.00
9	Naphthalene	1.000	1.003	-0.3	130	0.00
10	Hexachlorobutadiene	1.000	0.998	0.2	132	0.00
11	2-Methylnaphthalene	1.000	1.035	-3.5	137	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	136	0.00
13 S	2,4,6-Tribromophenol	1.000	1.006	-0.6	164	0.00
14 S	2-Fluorobiphenyl	1.000	1.024	-2.4	137	-0.01
15	Acenaphthylene	1.000	1.067	-6.7	143	-0.02
16	Acenaphthene	1.000	1.058	-5.8	146	0.00
17	Fluorene	1.000	1.163	-16.3	152	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	144	0.00
19	Hexachlorobenzene	1.000	0.977	2.3	143	0.00
20	Pentachlorophenol	1.000	0.942	5.8	157	0.00
21	Phenanthrene	1.000	1.016	-1.6	147	0.00
22	Anthracene	1.000	1.010	-1.0	142	-0.02
23	Fluoranthene	1.000	1.021	-2.1	147	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	137	0.00
25	Pyrene	1.000	1.072	-7.2	147	0.00
26 S	Terphenyl-d14	1.000	1.059	-5.9	142	0.00
27	Benzo(a)anthracene	1.000	1.051	-5.1	145	0.00
28	Chrysene	1.000	1.014	-1.4	134	0.00
29	Indeno(1,2,3-cd)pyrene	1.000	0.880	12.0	116	0.00
30 I	Perylene-d12	5.000	5.000	0.0	124	-0.01
31	Benzo(b)fluoranthene	1.000	1.004	-0.4	115	0.00
32	Benzo(k)fluoranthene	1.000	1.111	-11.1	146	-0.01
33 C	Benzo(a)pyrene	1.000	1.055	-5.5	130	0.00
34	Dibenzo(a,h)anthracene	1.000	0.949	5.1	115	-0.01
35	Benzo(g,h,i)perylene	1.000	0.933	6.7	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0