

Data Path : Z:\HPCHEM1\BNA_M\Data\BM050415\
 Data File : BM001205.D
 Acq On : 05 May 2015 04:00
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 05 05:05:34 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	129	0.00
2	1,4-Dioxane	1.000	0.872	12.8	107	0.00
3	n-Nitrosodimethylamine	1.000	0.933	6.7	114	0.00
4 S	2-Fluorophenol	1.000	1.094	-9.4	139	-0.01
5 S	Phenol-d6	1.000	1.159	-15.9	148	0.02
6 I	Naphthalene-d8	5.000	5.000	0.0	134	0.00
7 S	Nitrobenzene-d5	1.000	1.064	-6.4	143	-0.01
8	Nitrobenzene	1.000	0.986	1.4	133	-0.01
9	Naphthalene	1.000	1.000	0.0	131	0.00
10	Hexachlorobutadiene	1.000	0.998	0.2	134	0.00
11	2-Methylnaphthalene	1.000	1.042	-4.2	139	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	143	-0.02
13 S	2,4,6-Tribromophenol	1.000	1.106	-10.6	192	0.00
14 S	2-Fluorobiphenyl	1.000	0.996	0.4	139	-0.01
15	Acenaphthylene	1.000	1.080	-8.0	152	-0.02
16	Acenaphthene	1.000	1.060	-6.0	153	0.00
17	Fluorene	1.000	1.130	-13.0	155	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	152	0.00
19	Hexachlorobenzene	1.000	0.952	4.8	148	0.00
20	Pentachlorophenol	1.000	1.130	-13.0	222	0.00
21	Phenanthrene	1.000	0.979	2.1	150	0.00
22	Anthracene	1.000	1.030	-3.0	154	-0.02
23	Fluoranthene	1.000	1.023	-2.3	156	-0.01
24 I	Chrysene-d12	5.000	5.000	0.0	148	0.00
25	Pyrene	1.000	1.050	-5.0	156	-0.01
26 S	Terphenyl-d14	1.000	1.052	-5.2	154	-0.01
27	Benzo(a)anthracene	1.000	1.045	-4.5	156	0.00
28	Chrysene	1.000	1.017	-1.7	146	0.00
29	Indeno(1,2,3-cd)pyrene	1.000	0.890	11.0	128	-0.01
30 I	Perylene-d12	5.000	5.000	0.0	137	-0.01
31	Benzo(b)fluoranthene	1.000	1.104	-10.4	140	-0.01
32	Benzo(k)fluoranthene	1.000	1.003	-0.3	146	-0.01
33 C	Benzo(a)pyrene	1.000	1.058	-5.8	144	-0.01
34	Dibenzo(a,h)anthracene	1.000	0.960	4.0	129	-0.02
35	Benzo(g,h,i)perylene	1.000	0.932	6.8	125	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0