

Data Path : Z:\HPCHEM1\BNA_M\Data\BM050415\
 Data File : BM001205.D
 Acq On : 05 May 2015 04:00
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 05 05:05:34 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 04 16:48:07 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.612	0.519	15.2	107	0.00
3	n-Nitrosodimethylamine	0.612	0.571	6.7	114	0.00
4 S	2-Fluorophenol	0.972	1.064	-9.5	139	-0.01
5 S	Phenol-d6	1.198	1.389	-15.9	148	0.02
6 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
7 S	Nitrobenzene-d5	0.283	0.301	-6.4	143	-0.01
8	Nitrobenzene	0.303	0.318	-5.0	133	-0.01
9	Naphthalene	1.099	1.100	-0.1	131	0.00
10	Hexachlorobutadiene	0.212	0.211	0.5	134	0.00
11	2-Methylnaphthalene	0.754	0.785	-4.1	139	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	143	-0.02
13 S	2,4,6-Tribromophenol	0.188	0.236	-25.5#	192#	0.00
14 S	2-Fluorobiphenyl	1.654	1.646	0.5	139	-0.01
15	Acenaphthylene	2.185	2.360	-8.0	152#	-0.02
16	Acenaphthene	1.510	1.600	-6.0	153#	0.00
17	Fluorene	1.751	1.979	-13.0	155#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	152#	0.00
19	Hexachlorobenzene	0.269	0.257	4.5	148	0.00
20	Pentachlorophenol	0.067	0.093	-38.8#	222#	0.00
21	Phenanthrene	1.287	1.260	2.1	150	0.00
22	Anthracene	1.064	1.096	-3.0	154#	-0.02
23	Fluoranthene	1.303	1.333	-2.3	156#	-0.01
24 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
25	Pyrene	1.562	1.640	-5.0	156#	-0.01
26 S	Terphenyl-d14	0.689	0.725	-5.2	154#	-0.01
27	Benzo(a)anthracene	1.410	1.473	-4.5	156#	0.00
28	Chrysene	1.384	1.407	-1.7	146	0.00
29	Indeno(1,2,3-cd)pyrene	1.262	1.123	11.0	128	-0.01
30 I	Perylene-d12	1.000	1.000	0.0	137	-0.01
31	Benzo(b)fluoranthene	1.540	1.701	-10.5	140	-0.01
32	Benzo(k)fluoranthene	1.442	1.446	-0.3	146	-0.01
33 C	Benzo(a)pyrene	1.299	1.375	-5.9	144	-0.01
34	Dibenzo(a,h)anthracene	1.139	1.094	4.0	129	-0.02
35	Benzo(g,h,i)perylene	1.202	1.121	6.7	125	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0