


Data Path : Z:\HPCHEM1\BNA M\DATA\BM050416\
 Data File : BM005218.D
 Acq On : 04 May 2016 09:57
 Operator : UM/SJ
 Sample : DFTPP63
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP63

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042815.M
 Title : SVOA CALIBRATION
 Last Update : Tue Apr 28 06:36:34 2015



AutoFind: Scans 2535, 2536, 2537; Background Corrected with Scan 2530

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.3	29024	PASS
68	69	0.00	2	1.9	573	PASS
69	198	0.00	100	33.5	30122	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.1	39664	PASS
197	198	0.00	2	0.7	634	PASS
198	198	100	100	100.0	89970	PASS
199	198	5	9	6.7	6003	PASS
275	198	10	60	25.7	23125	PASS
365	198	1	100	3.1	2747	PASS
441	443	0.01	100	79.6	10888	PASS
442	198	50	100	78.2	70349	PASS
443	442	15	24	19.5	13683	PASS

Scan 2530 (17.568 min): BM005218.D
DFTPP63

m/z	Abundance
41.00	338.0
43.00	864.0
44.00	547.0
44.90	487.0
60.00	410.0
207.10	342.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

m/z	Abundance
38.00	328.0
39.10	1939.0
43.00	582.0
44.00	401.0
45.00	439.0
50.10	6778.0
51.10	27440.0
52.10	1424.0
55.00	326.0
56.00	940.0
57.00	1983.0
60.00	336.0
60.90	398.0
62.00	437.0
63.00	1149.0
65.00	616.0
68.10	512.0
69.00	28744.0
73.10	445.0
74.00	2964.0
75.00	4667.0
76.10	1588.0
77.10	34128.0
78.10	2467.0
79.00	2410.0
80.00	1623.0
81.00	2325.0
82.00	598.0
83.00	644.0
85.00	567.0
86.00	709.0
87.00	310.0
91.00	511.0
92.00	627.0
93.00	3697.0
98.00	2976.0
99.00	2447.0
101.00	1382.0
103.00	464.0
104.00	821.0
105.00	856.0
107.00	10763.0
108.00	1682.0
109.00	325.0
110.00	18904.0
111.00	2989.0
112.00	397.0
116.00	586.0
117.00	8755.0
118.00	608.0
122.00	690.0
123.00	1044.0
123.90	475.0
125.00	475.0
127.10	35168.0
128.10	2767.0
129.00	13079.0
130.00	1206.0
134.00	394.0
135.00	1182.0
136.00	466.0
137.00	512.0
141.00	1715.0
142.00	560.0
143.00	355.0
147.00	1027.0
148.00	2115.0
149.00	467.0
153.00	571.0
154.00	431.0
155.00	971.0
156.10	1573.0
158.00	339.0
159.00	306.0
160.00	548.0
161.00	863.0
165.00	598.0
166.00	572.0
167.00	3988.0
168.00	2256.0
169.00	320.0
172.00	316.0
173.00	406.0
174.00	815.0
175.00	1418.0
176.00	409.0
177.00	683.0
179.00	2733.0
180.00	2022.0
181.00	941.0
185.00	1414.0
186.10	10607.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

187.00	2889.0
188.00	333.0
189.00	631.0
192.00	824.0
193.10	1058.0
196.00	2182.0
197.10	499.0
198.00	74688.0
199.00	5212.0
200.00	418.0
201.50	375.0
203.00	500.0
204.10	2690.0
205.10	4499.0
206.10	19968.0
207.10	2911.0
208.00	744.0
211.10	663.0
216.00	462.0
217.00	4873.0
218.00	698.0
221.10	2819.0
221.80	670.0
223.00	1111.0
224.10	10634.0
225.10	2792.0
226.10	338.0
227.00	4287.0
228.00	665.0
229.00	906.0
231.00	353.0
237.00	414.0
242.00	535.0
243.10	581.0
244.10	7984.0
245.10	1151.0
246.00	1657.0
246.90	330.0
254.10	363.0
255.10	39720.0
256.10	6061.0
257.00	491.0
258.00	2305.0
259.00	393.0
265.00	857.0
273.00	1141.0
274.00	2929.0
275.10	17344.0
276.10	2373.0
277.00	1470.0
293.00	359.0
296.00	4779.0
297.00	747.0
303.00	547.0
315.00	456.0
316.00	309.0
323.00	1483.0
334.00	892.0
352.00	431.0
354.10	420.0
365.00	1802.0
372.00	702.0
403.00	402.0
421.00	324.0
422.00	357.0
423.10	2353.0
424.00	657.0
441.10	7012.0
442.10	43736.0
443.10	9120.0
444.00	814.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

m/z	Abundance
38.00	405.0
39.10	2381.0
41.00	316.0
43.00	543.0
44.00	446.0
45.00	407.0
50.10	8350.0
51.10	33352.0
52.00	1748.0
55.00	361.0
56.00	1074.0
57.00	2403.0
60.00	329.0
61.00	477.0
62.00	468.0
63.00	1402.0
65.00	703.0
68.10	613.0
69.00	34440.0
73.00	421.0
74.10	3629.0
75.00	5732.0
76.10	2036.0
77.10	42552.0
78.10	3016.0
79.00	2776.0
80.00	1947.0
81.00	3020.0
82.00	674.0
83.00	689.0
85.00	650.0
86.00	741.0
87.00	446.0
91.00	689.0
92.00	659.0
93.00	4849.0
94.00	348.0
98.00	3872.0
99.00	2979.0
101.00	1673.0
103.00	572.0
104.00	1086.0
105.00	1077.0
106.00	301.0
107.00	13637.0
108.00	1885.0
109.10	426.0
110.00	24904.0
111.00	3683.0
112.00	452.0
116.00	805.0
117.00	11149.0
118.00	853.0
122.00	901.0
123.00	1372.0
124.00	692.0
125.00	644.0
126.20	341.0
127.00	46328.0
128.00	3611.0
129.00	17800.0
130.00	1460.0
131.00	311.0
134.00	560.0
135.00	1553.0
136.10	656.0
137.00	767.0
141.00	2453.0
142.00	719.0
143.00	593.0
146.00	439.0
147.00	1321.0
148.00	2842.0
149.00	633.0
151.10	336.0
153.00	824.0
154.00	591.0
155.00	1323.0
156.00	2159.0
157.00	364.0
158.00	478.0
159.00	353.0
160.00	740.0
161.00	1106.0
165.00	891.0
166.00	766.0
167.10	5340.0
168.10	2782.0
169.00	438.0
171.90	448.0
173.00	555.0
174.00	1094.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

175.10	1999.0
176.10	452.0
177.00	947.0
178.00	338.0
179.00	3985.0
180.00	2662.0
181.00	1209.0
184.00	333.0
185.00	1963.0
186.10	14778.0
187.10	4327.0
188.00	424.0
189.00	851.0
191.00	481.0
192.00	1240.0
193.10	1456.0
196.00	2945.0
197.10	808.0
198.00	104688.0
199.00	7043.0
200.00	566.0
201.50	497.0
203.00	726.0
204.10	3935.0
205.10	6730.0
206.10	28392.0
207.10	3882.0
208.00	1013.0
209.10	306.0
210.30	386.0
211.10	1139.0
215.00	320.0
216.00	678.0
217.00	7206.0
218.00	959.0
221.10	4226.0
221.80	991.0
223.00	1569.0
224.10	15428.0
225.10	4056.0
226.10	489.0
227.00	6625.0
228.00	899.0
229.00	1342.0
231.00	592.0
234.00	431.0
235.00	451.0
236.00	304.0
237.00	556.0
241.10	335.0
242.00	848.0
243.10	830.0
244.10	12030.0
245.10	1540.0
246.00	2321.0
247.00	501.0
249.00	432.0
253.00	307.0
254.20	550.0
255.10	61152.0
256.10	8923.0
257.00	733.0
258.00	3430.0
259.00	612.0
265.00	1245.0
273.00	1864.0
274.10	4891.0
275.10	26832.0
276.00	3704.0
277.00	2514.0
278.00	351.0
285.10	432.0
293.00	577.0
296.00	7435.0
297.00	997.0
303.10	808.0
315.00	740.0
316.00	493.0
323.10	2288.0
324.10	413.0
327.00	415.0
334.10	1503.0
335.00	449.0
346.10	566.0
352.00	639.0
353.00	526.0
354.10	721.0
365.00	3203.0
366.00	441.0
372.10	1049.0
373.00	315.0
383.00	312.0
402.00	525.0
403.00	735.0
421.00	556.0
422.10	587.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

423.10	4324.0
424.10	983.0
441.10	12347.0
442.10	81224.0
443.10	15523.0
444.10	1400.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

m/z	Abundance
38.00	305.0
39.10	1990.0
43.00	719.0
44.00	539.0
45.00	424.0
50.10	6534.0
51.10	26280.0
52.00	1331.0
55.00	337.0
56.00	820.0
57.00	2019.0
60.00	331.0
61.00	405.0
62.00	395.0
63.00	1018.0
65.00	554.0
68.00	594.0
69.00	27184.0
73.00	433.0
74.00	2886.0
75.00	4443.0
76.10	1608.0
77.10	33472.0
78.10	2189.0
79.00	2282.0
80.00	1585.0
81.00	2341.0
82.00	605.0
83.00	601.0
85.00	502.0
86.00	726.0
87.00	383.0
91.00	591.0
92.00	589.0
93.00	3794.0
98.00	2975.0
99.00	2446.0
101.00	1318.0
103.00	435.0
104.00	915.0
105.00	759.0
107.00	10753.0
108.00	1524.0
109.00	365.0
110.00	19752.0
111.00	2947.0
112.00	387.0
116.00	547.0
117.00	9222.0
118.00	644.0
122.00	715.0
123.00	1041.0
124.00	536.0
125.00	509.0
127.10	37496.0
128.00	2700.0
129.00	14074.0
130.00	1233.0
131.00	323.0
134.00	429.0
135.00	1232.0
136.00	504.0
137.00	654.0
141.00	1966.0
142.00	655.0
143.00	461.0
145.90	319.0
147.00	1094.0
148.00	2348.0
149.00	477.0
153.00	669.0
154.00	425.0
155.00	1131.0
156.10	1771.0
157.10	349.0
158.00	386.0
159.00	304.0
160.00	692.0
161.00	925.0
165.00	804.0
166.00	639.0
167.00	4521.0
168.00	2324.0
169.00	359.0
171.90	345.0
173.00	462.0
174.00	903.0
175.00	1830.0
176.10	423.0
177.00	785.0
179.00	3203.0
180.10	2178.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63

181.00	960.0
185.10	1535.0
186.10	12513.0
187.00	3606.0
188.10	368.0
189.00	731.0
191.00	407.0
192.00	1120.0
193.00	1249.0
196.00	2534.0
197.10	597.0
198.00	90536.0
199.00	5755.0
200.00	503.0
201.50	394.0
203.00	601.0
204.10	3170.0
205.10	5779.0
206.10	24760.0
207.10	3350.0
208.00	817.0
210.10	391.0
211.10	896.0
216.00	458.0
217.00	6396.0
218.00	860.0
221.00	3676.0
221.90	803.0
223.00	1420.0
224.10	13799.0
225.10	3369.0
226.10	401.0
227.00	5565.0
228.00	789.0
229.00	1249.0
231.10	469.0
234.00	347.0
235.00	384.0
237.00	474.0
242.00	716.0
243.10	803.0
244.10	10618.0
245.10	1372.0
246.00	2100.0
247.00	394.0
249.00	398.0
254.20	589.0
255.10	54928.0
256.10	8017.0
257.00	567.0
258.00	3054.0
259.00	480.0
265.00	1267.0
273.00	1608.0
274.10	4528.0
275.10	25200.0
276.10	3157.0
277.00	2274.0
278.00	361.0
285.00	366.0
293.00	533.0
296.00	7111.0
297.00	1027.0
303.00	732.0
314.00	342.0
315.00	734.0
316.00	428.0
323.10	2246.0
324.00	330.0
327.00	407.0
334.10	1488.0
335.00	408.0
346.00	519.0
352.00	696.0
353.00	486.0
354.10	823.0
365.00	3236.0
366.00	460.0
372.10	1202.0
383.00	339.0
402.00	557.0
403.00	715.0
421.00	611.0
422.00	645.0
423.10	4482.0
424.00	1065.0
441.10	13307.0
442.10	86088.0
443.10	16408.0
444.10	1528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP63