

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050421\
 Data File : BM029793.D
 Acq On : 04 May 2021 17:23
 Operator : CG/JU
 Sample : PB135972BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK972

Quant Time: May 04 18:01:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 04 14:23:48 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	116246	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	479344	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	277323	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.96	188	526359	20.00	ng/ul	0.00
79) Chrysene-d12	21.15	240	521682	20.00	ng/ul	0.00
88) Perylene-d12	23.31	264	574601	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.08	96	17911	5.70	ng/uL	0.00
4) Pyridine-d5	3.49	84	195461	27.15	ng/ul	0.00
7) Phenol-d5	6.76	99	292885	29.02	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.92	67	190740	30.59	ng/ul	0.00
11) 2-Chlorophenol-d4	7.10	132	245830	30.57	ng/ul	0.00
15) 4-Methylphenol-d8	8.29	113	225855	28.03	ng/ul	0.00
21) Nitrobenzene-d5	8.72	128	106177	29.81	ng/ul	0.00
24) 2-Nitrophenol-d4	9.45	143	102791	26.43	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.98	165	195140	26.83	ng/ul	0.00
31) 4-Chloroaniline-d4	10.50	131	289732	27.34	ng/ul	0.00
46) Dimethylphthalate-d6	13.64	166	632985	30.56	ng/ul	0.00
49) Acenaphthylene-d8	13.91	160	836686	30.89	ng/ul	0.00
54) 4-Nitrophenol-d4	14.46	143	51812	16.26	ng/ul	0.01
60) Fluorene-d10	15.22	176	553720	31.42	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.35	200	61417	17.38	ng/ul	0.00
73) Anthracene-d10	17.06	188	820660	31.52	ng/ul	0.00
81) Pyrene-d10	19.36	212	915832	34.22	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.17	264	1052836	32.63	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050421\
Data File : BM029793.D
Acq On : 04 May 2021 17:23
Operator : CG/JU
Sample : PB135972BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK972

Quant Time: May 04 18:01:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050421.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 04 14:23:48 2021
Response via : Initial Calibration

