

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050515\
 Data File : BM001213.D
 Acq On : 05 May 2015 17:28
 Operator : TP/IZ
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.2

Manual Integrations
 APPROVED

apatel
 5/6/2015 1:53:34 PM

Quant Time: May 06 06:14:04 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 06 02:27:20 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	15048	5.00	ng	0.00
6) Naphthalene-d8	10.30	136	59646	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	37385	5.00	ng	0.00
18) Phenanthrene-d10	16.92	188	71074	5.00	ng	-0.02
24) Chrysene-d12	21.14	240	79289	5.00	ng	0.00
30) Perylene-d12	23.28	264	63608	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.13	112	633	0.21	ng	0.02
5) Phenol-d6	6.72	99	739	0.20	ng	0.00
7) Nitrobenzene-d5	8.69	82	589	0.17	ng	0.01
13) 2,4,6-Tribromophenol	15.70	330	305	0.21	ng	0.00
14) 2-Fluorobiphenyl	12.81	172	2265	0.18	ng	0.00
26) Terphenyl-d14	19.58	244	2175	0.20	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.35	42	328	0.18	ng	# 3
8) Nitrobenzene	8.73	77	607	0.17	ng	96
9) Naphthalene	10.35	128	2631	0.20	ng	95
10) Hexachlorobutadiene	10.64	225	516	0.20	ng	97
11) 2-Methylnaphthalene	11.99	142	1771	0.20	ng	98
15) Acenaphthylene	13.91	152	2820	0.17	ng	100
16) Acenaphthene	14.26	154	1987	0.18	ng	99
17) Fluorene	15.24	166	2386	0.18	ng	97
19) Hexachlorobenzene	16.25	284	1032	0.27	ng	# 70
21) Phenanthrene	16.98	178	3315m	0.17	ng	
22) Anthracene	17.08	178	2988m	0.18	ng	
23) Fluoranthene	19.01	202	4524	0.23	ng	99
25) Pyrene	19.36	202	4762	0.19	ng	99
27) Benzo(a)anthracene	21.13	228	4296m	0.19	ng	
28) Chrysene	21.17	228	4435	0.20	ng	97
29) Indeno(1,2,3-cd)pyrene	25.42	276	3346	0.17	ng	99
31) Benzo(b)fluoranthene	22.65	252	4004	0.21	ng	# 91
32) Benzo(k)fluoranthene	22.69	252	3421	0.18	ng	# 90
33) Benzo(a)pyrene	23.19	252	3159	0.19	ng	# 90
34) Dibenzo(a,h)anthracene	25.43	278	2605	0.18	ng	# 88
35) Benzo(g,h,i)perylene	26.08	276	2777	0.18	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050515\
Data File : BM001213.D
Acq On : 05 May 2015 17:28
Operator : TP/IZ
Sample : SSTDICC0.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.2

Manual Integrations
APPROVED

apatel
5/6/2015 1:53:34 PM

Quant Time: May 06 06:14:04 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 06 02:27:20 2015
Response via : Initial Calibration

