

Data Path : Z:\HPCHEM1\BNA_M\Data\BM050515\
 Data File : BM001221.D
 Acq On : 06 May 2015 10:39
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 06 13:13:27 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 06 12:46:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	1,4-Dioxane	1.000	0.929	7.1	84	0.00
3	n-Nitrosodimethylamine	1.000	0.892	10.8	93	0.00
4 S	2-Fluorophenol	1.000	1.011	-1.1	93	0.00
5 S	Phenol-d6	1.000	0.970	3.0	88	-0.02
6 I	Naphthalene-d8	5.000	5.000	0.0	91	0.00
7 S	Nitrobenzene-d5	1.000	0.989	1.1	87	0.00
8	Nitrobenzene	1.000	0.950	5.0	89	0.00
9	Naphthalene	1.000	1.027	-2.7	91	0.00
10	Hexachlorobutadiene	1.000	1.027	-2.7	94	0.00
11	2-Methylnaphthalene	1.000	1.006	-0.6	89	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	86	0.00
13 S	2,4,6-Tribromophenol	1.000	0.896	10.4	83	0.00
14 S	2-Fluorobiphenyl	1.000	1.046	-4.6	88	-0.01
15	Acenaphthylene	1.000	1.007	-0.7	85	0.00
16	Acenaphthene	1.000	1.031	-3.1	87	0.00
17	Fluorene	1.000	1.004	-0.4	81	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	83	0.00
19	Hexachlorobenzene	1.000	1.103	-10.3	92	0.00
20	Pentachlorophenol	1.000	1.019	-1.9	81	0.00
21	Phenanthrene	1.000	0.987	1.3	80	0.00
22	Anthracene	1.000	0.968	3.2	82	0.00
23	Fluoranthene	1.000	1.073	-7.3	86	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	92	0.00
25	Pyrene	1.000	0.954	4.6	88	0.00
26 S	Terphenyl-d14	1.000	0.950	5.0	86	0.00
27	Benzo(a)anthracene	1.000	0.991	0.9	91	0.00
28	Chrysene	1.000	1.043	-4.3	94	0.00
29	Indeno(1,2,3-cd)pyrene	1.000	1.226	-22.6	110	0.00
30 I	Perylene-d12	5.000	5.000	0.0	103	0.00
31	Benzo(b)fluoranthene	1.000	1.045	-4.5	111	0.01
32	Benzo(k)fluoranthene	1.000	0.939	6.1	90	0.01
33 C	Benzo(a)pyrene	1.000	1.013	-1.3	103	0.00
34	Dibenzo(a,h)anthracene	1.000	1.074	-7.4	109	0.00
35	Benzo(g,h,i)perylene	1.000	1.102	-10.2	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0