

Data Path : Z:\HPCHEM1\BNA_M\Data\BM050515\
 Data File : BM001221.D
 Acq On : 06 May 2015 10:39
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 06 13:13:27 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 06 12:46:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.522	0.530	-1.5	84	0.00
3	n-Nitrosodimethylamine	0.607	0.541	10.9	93	0.00
4 S	2-Fluorophenol	1.045	1.057	-1.1	93	0.00
5 S	Phenol-d6	1.329	1.289	3.0	88	-0.02
6 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
7 S	Nitrobenzene-d5	0.283	0.280	1.1	87	0.00
8	Nitrobenzene	0.300	0.296	1.3	89	0.00
9	Naphthalene	1.089	1.118	-2.7	91	0.00
10	Hexachlorobutadiene	0.211	0.216	-2.4	94	0.00
11	2-Methylnaphthalene	0.765	0.769	-0.5	89	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
13 S	2,4,6-Tribromophenol	0.266	0.249	6.4	83	0.00
14 S	2-Fluorobiphenyl	1.529	1.599	-4.6	88	-0.01
15	Acenaphthylene	2.155	2.171	-0.7	85	0.00
16	Acenaphthene	1.355	1.397	-3.1	87	0.00
17	Fluorene	2.203	2.213	-0.5	81	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
19	Hexachlorobenzene	0.286	0.315	-10.1	92	0.00
20	Pentachlorophenol	0.107	0.087	18.7	81	0.00
21	Phenanthrene	2.093	2.159	-3.2	80	0.00
22	Anthracene	1.713	1.723	-0.6	82	0.00
23	Fluoranthene	1.971	2.116	-7.4	86	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
25	Pyrene	1.598	1.525	4.6	88	0.00
26 S	Terphenyl-d14	0.716	0.680	5.0	86	0.00
27	Benzo(a)anthracene	1.426	1.414	0.8	91	0.00
28	Chrysene	1.376	1.436	-4.4	94	0.00
29	Indeno(1,2,3-cd)pyrene	1.092	1.339	-22.6	110	0.00
30 I	Perylene-d12	1.000	1.000	0.0	103	0.00
31	Benzo(b)fluoranthene	1.599	1.671	-4.5	111	0.01
32	Benzo(k)fluoranthene	1.469	1.380	6.1	90	0.01
33 C	Benzo(a)pyrene	1.320	1.337	-1.3	103	0.00
34	Dibenzo(a,h)anthracene	1.074	1.153	-7.4	109	0.00
35	Benzo(g,h,i)perylene	1.114	1.228	-10.2	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0