

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050516\
 Data File : BM005274.D
 Acq On : 06 May 2016 17:42
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02049

Quant Time: May 06 18:37:04 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 06 04:26:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.776	0.745	4.0	82	0.00
3 S	1,4-Dioxane-d8	0.425	0.409	3.8	85	0.00
4	Benzaldehyde	0.879	1.123	-27.8#	88	0.00
5 S	Phenol-d5	1.814	1.828	-0.8	89	0.00
6	Phenol	1.875	1.931	-3.0	90	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.035	1.059	-2.3	87	0.00
8	Bis(2-Chloroethyl)ether	1.411	1.448	-2.6	88	0.00
9 S	2-Chlorophenol-d4	1.370	1.407	-2.7	90	0.00
10	2-Chlorophenol	1.401	1.438	-2.6	89	0.00
11	2-Methylphenol	1.449	1.475	-1.8	90	0.00
12	2,2'-oxybis(1-Chloropropane	1.915	1.979	-3.3	89	0.00
13 S	4-Methylphenol-d8	1.499	1.560	-4.1	92	0.00
14	Acetophenone	2.221	2.424	-9.1	91	0.00
15 P	N-Nitroso-di-n-propylamine	1.161	1.222	-5.3	91	0.00
16	4-Methylphenol	1.608	1.680	-4.5	90	0.00
17	Hexachloroethane	0.527	0.540	-2.5	91	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
19 S	Nitrobenzene-d5	0.143	0.146	-2.1	92	0.00
20	Nitrobenzene	0.357	0.365	-2.2	92	0.00
21	Isophorone	0.694	0.712	-2.6	92	0.00
22 S	2-Nitrophenol-d4	0.162	0.166	-2.5	91	0.00
23 C	2-Nitrophenol	0.174	0.180	-3.4	93	0.00
24	2,4-Dimethylphenol	0.370	0.385	-4.1	94	0.00
25	Bis(2-Chloroethoxy)methane	0.432	0.429	0.7	90	0.00
26 S	2,4-Dichlorophenol-d3	0.300	0.316	-5.3	94	0.00
27 C	2,4-Dichlorophenol	0.308	0.320	-3.9	93	0.00
28	Naphthalene	1.006	1.031	-2.5	93	0.00
29 S	4-Chloroaniline-d4	0.362	0.430	-18.8	94	0.00
30	4-Chloroaniline	0.369	0.440	-19.2	95	0.00
31 C	Hexachlorobutadiene	0.183	0.190	-3.8	95	0.00
32	Caprolactam	0.115	0.111	3.5	90	0.00
33 C	4-Chloro-3-methylphenol	0.369	0.389	-5.4	94	0.00
34	2-Methylnaphthalene	0.753	0.779	-3.5	93	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
36	1,2,4,5-Tetrachlorobenzene	0.608	0.623	-2.5	95	0.00
37	Hexachlorocyclopentadiene	0.311	0.286	8.0	93	0.00
38 C	2,4,6-Trichlorophenol	0.405	0.405	0.0	93	0.00
39	2,4,5-Trichlorophenol	0.450	0.462	-2.7	96	0.00
40	1,1'-Biphenyl	1.584	1.598	-0.9	94	0.00
41	2-Chloronaphthalene	1.198	1.223	-2.1	95	0.00
42	2-Nitroaniline	0.369	0.376	-1.9	93	0.00
43 S	Dimethylphthalate-d6	1.603	1.609	-0.4	93	0.00
44	Dimethylphthalate	1.602	1.613	-0.7	93	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050516\
 Data File : BM005274.D
 Acq On : 06 May 2016 17:42
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02049

Quant Time: May 06 18:37:04 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 06 04:26:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.316	0.331	-4.7	95	0.00
46 S	Acenaphthylene-d8	1.880	1.937	-3.0	95	0.00
47	Acenaphthylene	1.989	2.061	-3.6	95	0.00
48	3-Nitroaniline	0.337	0.359	-6.5	94	0.00
49 C	Acenaphthene	1.311	1.346	-2.7	97	0.00
50	2,4-Dinitrophenol	0.182	0.152	16.5	94	0.00
51 S	4-Nitrophenol-d4	0.293	0.280	4.4	89	0.00
52	4-Nitrophenol	0.243	0.239	1.6	94	0.00
53	Dibenzofuran	1.902	1.958	-2.9	95	0.00
54	2,4-Dinitrotoluene	0.471	0.496	-5.3	96	0.00
55	2,3,4,6-Tetrachlorophenol	0.382	0.398	-4.2	96	0.00
56	Diethylphthalate	1.618	1.629	-0.7	93	0.00
57 S	Fluorene-d10	1.384	1.413	-2.1	95	0.00
58	Fluorene	1.487	1.554	-4.5	98	0.00
59	4-Chlorophenyl-phenylether	0.737	0.772	-4.7	99	0.00
60	4-Nitroaniline	0.357	0.381	-6.7	98	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.113	0.118	-4.4	99	0.00
63	4,6-Dinitro-2-methylphenol	0.118	0.124	-5.1	97	0.00
64	N-Nitrosodiphenylamine	0.548	0.586	-6.9	97	0.00
65	4-Bromophenyl-phenylether	0.192	0.205	-6.8	96	0.00
66	Hexachlorobenzene	0.215	0.228	-6.0	96	0.00
67	Atrazine	0.200	0.222	-11.0	95	0.00
68 C	Pentachlorophenol	0.120	0.119	0.8	94	0.00
69	Phenanthrene	1.052	1.102	-4.8	94	0.00
70 S	Anthracene-d10	0.884	0.931	-5.3	95	0.00
71	Anthracene	1.048	1.129	-7.7	97	0.00
72	Carbazole	0.922	1.026	-11.3	93	0.00
73	Di-n-butylphthalate	1.125	1.178	-4.7	92	0.00
74 C	Fluoranthene	1.165	1.320	-13.3	92	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
76 S	Pyrene-d10	0.923	0.952	-3.1	92	0.00
77	Pyrene	1.160	1.209	-4.2	93	0.00
78	Butylbenzylphthalate	0.482	0.492	-2.1	91	0.00
79	3,3'-Dichlorobenzidine	0.360	0.390	-8.3	90	0.00
80	Benzo(a)anthracene	1.150	1.175	-2.2	91	0.00
81	Bis(2-ethylhexyl)phthalate	0.669	0.698	-4.3	89	0.00
82	Chrysene	1.091	1.145	-4.9	93	0.00
83 I	Perylene-d12	1.000	1.000	0.0	90	0.00
84	Di-n-octyl phthalate	1.168	1.276	-9.2	90	0.00
85	Benzo(b)fluoranthene	1.169	1.202	-2.8	91	0.00
86	Benzo(k)fluoranthene	1.101	1.177	-6.9	91	0.00
87 S	Benzo(a)pyrene-d12	0.885	0.911	-2.9	90	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM050516\
Data File : BM005274.D
Acq On : 06 May 2016 17:42
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02049

Quant Time: May 06 18:37:04 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 06 04:26:22 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.106	1.148	-3.8	90	0.00
89	Indeno(1,2,3-cd)pyrene	1.206	1.167	3.2	83	-0.01
90	Dibenzo(a,h)anthracene	1.010	0.995	1.5	84	-0.01
91	Benzo(g,h,i)perylene	1.022	0.944	7.6	81	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0