


Hexachlorocyclopentadiene

Response Ratio




$R = 3.67e-002 A^2 + 9.95e-002 A - 4.73e-002$
Coef of Det (r^2) = 0.996 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_M\METHODS\8270-BM050517.M
Calibration Table Last Updated: Fri May 05 14:52:12 2017

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $2.41 \times 10^{-1} \times \text{Amt} - 1.04 \times 10^{-1}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_M\METHODS\8270-BM050517.M
Calibration Table Last Updated: Fri May 05 14:52:12 2017

4,6-Dinitro-2-methylphenol



Resp Ratio = 1.54×10^{-1} * Amt - 4.54×10^{-2}
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_M\METHODS\8270-BM050517.M
Calibration Table Last Updated: Fri May 05 14:52:12 2017



Method Name: Z:\HPCHEM1\BNA_M\METHODS\8270-BM050517.M
Calibration Table Last Updated: Fri May 05 14:52:12 2017