

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050525\
 Data File : BM050104.D
 Acq On : 05 May 2025 18:18
 Operator : RC/JU
 Sample : Q1936-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SMALL-PILE-B

Quant Time: May 06 05:19:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 28 18:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.745	152	282446	20.000	ng	-0.02
21) Naphthalene-d8	10.539	136	1024683	20.000	ng	-0.02
39) Acenaphthene-d10	14.392	164	673586	20.000	ng	-0.02
64) Phenanthrene-d10	17.145	188	1301412	20.000	ng	-0.01
76) Chrysene-d12	21.386	240	1230717	20.000	ng	-0.01
86) Perylene-d12	24.374	264	1268881	20.000	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	1188038	73.655	ng	-0.02
7) Phenol-d6	6.934	99	1515522	74.755	ng	-0.02
23) Nitrobenzene-d5	8.904	82	912091	43.862	ng	-0.02
42) 2,4,6-Tribromophenol	15.892	330	736962	73.993	ng	-0.01
45) 2-Fluorobiphenyl	13.010	172	2485072	43.644	ng	-0.02
79) Terphenyl-d14	19.768	244	3941049	47.385	ng	-0.01


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050525\
Data File : BM050104.D
Acq On : 05 May 2025 18:18
Operator : RC/JU
Sample : Q1936-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SMALL-PILE-B

Quant Time: May 06 05:19:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050525\
Data File : BM050104.D
Acq On : 05 May 2025 18:18
Operator : RC/JU
Sample : Q1936-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SMALL-PILE-B

Quant Time: May 06 05:19:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

