

Data Path : Z:\HPCHEM1\BNA M\Data\BM050615\
 Data File : BM001237.D
 Acq On : 07 May 2015 03:50
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 07 07:29:28 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-ACID-BM050615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 07 07:00:24 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.324	0.332	-2.5	97	0.00
3 S	2-Fluorophenol	0.817	0.846	-3.5	96	0.00
4 S	Phenol-d6	1.104	1.181	-7.0	98	0.00
5	2-Chlorophenol	0.711	0.725	-2.0	96	0.00
6 C	Phenol	1.108	1.082	2.3	91	0.00
7	2-Methylphenol	0.584	0.583	0.2	93	0.00
8	3+4-Methylphenols	0.761	0.819	-7.6	99	-0.03
9 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
10 S	Nitrobenzene-d5	0.285	0.303	-6.3	101	0.00
11 C	2-Nitrophenol	0.112	0.110	1.8	95	0.00
12	2,4-Dimethylphenol	0.241	0.242	-0.4	95	0.00
13 C	2,4-Dichlorophenol	0.246	0.244	0.8	94	0.00
14 C	4-Chloro-3-methylphenol	0.255	0.252	1.2	94	0.00
15	2-Methylnaphthalene	0.714	0.751	-5.2	97	0.00
16 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
17 S	2,4,6-Tribromophenol	0.208	0.197	5.3	94	0.00
18 C	2,4,6-Trichlorophenol	0.366	0.357	2.5	95	0.00
19	2,4,5-Trichlorophenol	0.413	0.397	3.9	93	0.00
20 S	2-Fluorobiphenyl	1.559	1.602	-2.8	98	0.00
21 P	2,4-Dinitrophenol	0.084	0.051	39.3#	92	0.00
22 P	4-Nitrophenol	0.154	0.096	37.7#	91	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
24	4,6-Dinitro-2-methylphenol	0.105	0.073	30.5#	94	0.00
25	Hexachlorobenzene	0.368	0.383	-4.1	99	0.00
26 C	Pentachlorophenol	0.092	0.072	21.7#	91	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28 S	Terphenyl-d14	0.658	0.677	-2.9	100	0.00
29 I	Perylene-d12	1.000	1.000	0.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1