

Data Path : Z:\HPCHEM1\BNA M\DATA\BM050616\
 Data File : BM005298.D
 Acq On : 07 May 2016 08:13
 Operator : UM/SJ
 Sample : H2743-07
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H4014

Quant Time: May 08 00:33:12 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 07 07:05:19 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	61056	20.00	ng/ul	0.00
18) Naphthalene-d8	10.56	136	311020	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.41	164	211299	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.16	188	498552	20.00	ng/ul	0.00
75) Chrysene-d12	21.35	240	595888	20.00	ng/ul	0.00
83) Perylene-d12	23.62	264	547937	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.27	96	1079	0.83	ng/uL	0.00
5) Phenol-d5	6.95	99	26766	4.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.10	67	85199	26.97	ng/ul	0.00
9) 2-Chlorophenol-d4	7.30	132	88615	21.19	ng/ul	0.00
13) 4-Methylphenol-d8	8.47	113	55739	12.18	ng/ul	0.00
19) Nitrobenzene-d5	8.93	128	58710	26.44	ng/ul	0.00
22) 2-Nitrophenol-d4	9.65	143	66809	26.57	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.18	165	110837	23.72	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	0.00
43) Dimethylphthalate-d6	13.82	166	475070	28.05	ng/ul	0.00
46) Acenaphthylene-d8	14.10	160	547447	27.56	ng/ul	0.00
51) 4-Nitrophenol-d4	14.63	143	9288	3.00	ng/ul	0.01
57) Fluorene-d10	15.41	176	450172	30.78	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.53	200	84027	29.96	ng/ul	0.00
70) Anthracene-d10	17.26	188	787851	35.75	ng/ul	0.00
76) Pyrene-d10	19.56	212	932771	33.91	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.48	264	875484	36.10	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM050616\
Data File : BM005298.D
Acq On : 07 May 2016 08:13
Operator : UM/SJ
Sample : H2743-07
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
H4014

Quant Time: May 08 00:33:12 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat May 07 07:05:19 2016
Response via : Initial Calibration

