

Data Path : Z:\HPCHEM1\BNA M\DATA\BM050716\
 Data File : BM005324.D
 Acq On : 08 May 2016 00:38
 Operator : UM/SJ
 Sample : H2853-11
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9WZ7

Quant Time: May 08 02:17:13 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat May 07 07:05:19 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	64672	20.00	ng/ul	0.00
18) Naphthalene-d8	10.56	136	315758	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.41	164	211266	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.16	188	528618	20.00	ng/ul	0.00
75) Chrysene-d12	21.35	240	628356	20.00	ng/ul	0.00
83) Perylene-d12	23.62	264	565595	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.27	96	1733	1.26	ng/uL	0.00
5) Phenol-d5	6.95	99	40052	6.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.10	67	105396	31.50	ng/ul	0.00
9) 2-Chlorophenol-d4	7.30	132	113436	25.61	ng/ul	0.00
13) 4-Methylphenol-d8	8.47	113	68954	14.22	ng/ul	0.00
19) Nitrobenzene-d5	8.93	128	73823	32.75	ng/ul	0.00
22) 2-Nitrophenol-d4	9.65	143	87019	34.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.18	165	146386	30.86	ng/ul	0.00
29) 4-Chloroaniline-d4	10.72	131	6498	1.14	ng/ul	0.02
43) Dimethylphthalate-d6	13.82	166	585189	34.56	ng/ul	0.00
46) Acenaphthylene-d8	14.10	160	682744	34.37	ng/ul	0.00
51) 4-Nitrophenol-d4	14.63	143	16293	5.27	ng/ul	0.00
57) Fluorene-d10	15.40	176	516540	35.32	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.53	200	104806	35.24	ng/ul	0.00
70) Anthracene-d10	17.26	188	846035	36.21	ng/ul	0.00
76) Pyrene-d10	19.56	212	1019990	35.17	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.47	264	949803	37.94	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM050716\
Data File : BM005324.D
Acq On : 08 May 2016 00:38
Operator : UM/SJ
Sample : H2853-11
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9WZ7

Quant Time: May 08 02:17:13 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat May 07 07:05:19 2016
Response via : Initial Calibration

