

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050721\
 Data File : BM029900.D
 Acq On : 08 May 2021 06:35
 Operator : CG/JU
 Sample : SP5557
 Misc : SFAM SURROGATE
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5557

Quant Time: May 08 07:10:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 07 13:53:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	111717	20.00	ng/ul	0.00
20) Naphthalene-d8	10.33	136	519043	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.20	164	333825	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.94	188	685007	20.00	ng/ul	0.00
79) Chrysene-d12	21.13	240	676260	20.00	ng/ul	0.00
88) Perylene-d12	23.29	264	617677	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.06	96	35518	12.70	ng/uL	0.00
4) Pyridine-d5	3.47	84	563764	80.49	ng/ul	-0.01
7) Phenol-d5	6.74	99	844527	81.29	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.90	67	561896	87.22	ng/ul	0.00
11) 2-Chlorophenol-d4	7.09	132	676364	84.35	ng/ul	0.00
15) 4-Methylphenol-d8	8.27	113	677740	78.88	ng/ul	0.00
21) Nitrobenzene-d5	8.71	128	327035	91.43	ng/ul	0.00
24) 2-Nitrophenol-d4	9.43	143	356452	88.29	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.96	165	605639	74.81	ng/ul	0.00
31) 4-Chloroaniline-d4	10.47	131	914674	74.25	ng/ul	0.00
46) Dimethylphthalate-d6	13.63	166	2064857	79.80	ng/ul	0.00
49) Acenaphthylene-d8	13.89	160	2706088	82.53	ng/ul	0.00
54) 4-Nitrophenol-d4	14.43	143	271188	64.35	ng/ul	0.00
60) Fluorene-d10	15.20	176	1810882	82.44	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.34	200	275560	61.93	ng/ul	0.00
73) Anthracene-d10	17.05	188	2798513	81.13	ng/ul	0.00
81) Pyrene-d10	19.34	212	3080818	88.81	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.16	264	2793701	79.79	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050721\
Data File : BM029900.D
Acq On : 08 May 2021 06:35
Operator : CG/JU
Sample : SP5557
Misc : SFAM SURROGATE
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP5557

Quant Time: May 08 07:10:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 07 13:53:15 2021
Response via : Initial Calibration

