

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050721\
 Data File : BM029881.D
 Acq On : 07 May 2021 17:56
 Operator : CG/JU
 Sample : M2129-02MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E0AB9MS

Manual Integrations
 APPROVED

mohammad
 5/10/2021 11:45:14 AM

Quant Time: May 07 18:41:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 07 13:53:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	154253	20.00	ng/ul	0.00
20) Naphthalene-d8	10.33	136	712996	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.20	164	472598	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.95	188	987715	20.00	ng/ul	0.00
79) Chrysene-d12	21.14	240	1069649	20.00	ng/ul	0.00
88) Perylene-d12	23.29	264	1009346	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.07	96	12284	3.18	ng/uL	0.00
4) Pyridine-d5	3.49	84	85351	8.83	ng/ul	0.00
7) Phenol-d5	6.74	99	210816	14.70	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.90	67	143711	16.16	ng/ul	0.00
11) 2-Chlorophenol-d4	7.09	132	186091	16.81	ng/ul	0.00
15) 4-Methylphenol-d8	8.27	113	141584	11.93	ng/ul	0.00
21) Nitrobenzene-d5	8.71	128	92979	18.92	ng/ul	0.00
24) 2-Nitrophenol-d4	9.43	143	102055	18.40	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.96	165	202429	18.20	ng/ul	0.00
31) 4-Chloroaniline-d4	10.48	131	231734	13.69	ng/ul	0.00
46) Dimethylphthalate-d6	13.63	166	784209	21.41	ng/ul	0.00
49) Acenaphthylene-d8	13.89	160	882912	19.02	ng/ul	0.00
54) 4-Nitrophenol-d4	14.43	143	110878	18.58	ng/ul	0.00
60) Fluorene-d10	15.20	176	656885	21.12	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.34	200	148626	23.17	ng/ul	0.00
73) Anthracene-d10	17.05	188	1174231	23.61	ng/ul	0.00
81) Pyrene-d10	19.34	212	1389183	25.32	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.16	264	1363640	23.83	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.10	88	29423	7.039	ng/uL 95
5) Pyridine	3.50	79	150144	15.137	ng/ul 92
6) Benzaldehyde	6.71	77	140291	18.946	ng/ul 98
8) Phenol	6.77	94	301517	20.588	ng/ul 98
10) Bis(2-Chloroethyl)ether	6.99	93	220449	19.037	ng/ul 99
12) 2-Chlorophenol	7.12	128	228070	20.199	ng/ul 100
13) 2-Methylphenol	8.01	108	222136	20.019	ng/ul 98
14) 2,2'-oxybis(1-Chloropropan	8.09	45	347340	19.913	ng/ul 99
16) Acetophenone	8.38	105	375645	19.731	ng/ul 99
17) N-Nitroso-di-n-propylamine	8.37	70	202078	20.746	ng/ul 99
18) 4-Methylphenol	8.33	108	248344	20.188	ng/ul 100
19) Hexachloroethane	8.62	117	96770	20.286	ng/ul 98
22) Nitrobenzene	8.76	77	274477	21.438	ng/ul 98
23) Isophorone	9.27	82	576666	21.294	ng/ul 99
25) 2-Nitrophenol	9.46	139	128864	21.403	ng/ul 99
26) 2,4-Dimethylphenol	9.53	107	285145	20.989	ng/ul 98
27) Bis(2-Chloroethoxy)methane	9.76	93	325019	20.491	ng/ul 98
29) 2,4-Dichlorophenol	9.99	162	225943	20.689	ng/ul 98
30) Naphthalene	10.38	128	796132	20.117	ng/ul 100
32) 4-Chloroaniline	10.50	127	299928	17.204	ng/ul 97
33) Hexachlorobutadiene	10.66	225	140802	19.230	ng/ul 97
34) Caprolactam	11.29	113	104734m	27.023	ng/ul

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050721\
 Data File : BM029881.D
 Acq On : 07 May 2021 17:56
 Operator : CG/JU
 Sample : M2129-02MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E0AB9MS

Manual Integrations
 APPROVED

mohammad
 5/10/2021 11:45:14 AM

Quant Time: May 07 18:41:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 07 13:53:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 4-Chloro-3-methylphenol	11.64	107	293263	24.266	ng/ul	99
36) 2-Methylnaphthalene	12.00	142	586109	20.416	ng/ul	98
37) 1-Methylnaphthalene	12.22	142	574929	20.208	ng/ul	98
39) 1,2,4,5-Tetrachlorobenzene	12.38	216	295862	20.259	ng/ul	99
40) Hexachlorocyclopentadiene	12.35	237	98912	13.039	ng/ul	98
41) 2,4,6-Trichlorophenol	12.63	196	206297	23.165	ng/ul	99
42) 2,4,5-Trichlorophenol	12.70	196	237677	25.217	ng/ul	99
43) 1,1'-Biphenyl	13.03	154	766144	20.854	ng/ul	99
44) 2-Chloronaphthalene	13.07	162	602873	21.383	ng/ul	99
45) 2-Nitroaniline	13.29	65	197538	27.721	ng/ul	99
47) Dimethylphthalate	13.67	163	974252	26.561	ng/ul	99
48) 2,6-Dinitrotoluene	13.79	165	179163	27.209	ng/ul	97
50) Acenaphthylene	13.92	152	950774	21.177	ng/ul	98
51) 3-Nitroaniline	14.13	138	182402	24.659	ng/ul	99
52) Acenaphthene	14.27	153	690445	21.760	ng/ul	99
53) 2,4-Dinitrophenol	14.34	184	95678	21.356	ng/ul	98
55) 4-Nitrophenol	14.45	109	125121	27.275	ng/ul	95
56) Dibenzofuran	14.60	168	984839	22.661	ng/ul	98
57) 2,4-Dinitrotoluene	14.59	165	290613	29.826	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.84	232	240705	26.820	ng/ul	98
59) Diethylphthalate	15.05	149	1029373	26.907	ng/ul	100
61) Fluorene	15.26	166	882557	23.755	ng/ul	100
62) 4-Chlorophenyl-phenylether	15.26	204	421409	23.083	ng/ul	98
63) 4-Nitroaniline	15.30	138	197922m	26.993	ng/ul	
66) 4,6-Dinitro-2-methylphenol	15.35	198	168176	26.635	ng/ul	98
67) N-Nitrosodiphenylamine	15.47	169	831409	25.912	ng/ul	100
68) 4-Bromophenyl-phenylether	16.15	248	276969	25.192	ng/ul	99
69) Hexachlorobenzene	16.26	284	344262	26.369	ng/ul	95
70) Atrazine	16.43	200	322194	27.217	ng/ul	98
71) Pentachlorophenol	16.61	266	187059	25.005	ng/ul	100
72) Phenanthrene	16.99	178	1596417	27.411	ng/ul	100
74) Anthracene	17.09	178	1633579	27.436	ng/ul	99
75) 1,2,3,4-Tetrachlorobenzene	12.99	216	301259	19.282	ng/uL	98
76) Pentachlorobenzene	14.53	250	321779	21.596	ng/uL	99
77) Carbazole	17.36	167	1492627	27.360	ng/ul	99
78) Di-n-butylphthalate	17.93	149	1954531	30.185	ng/ul	100
80) Fluoranthene	19.02	202	2017338	30.208	ng/ul	100
82) Pyrene	19.37	202	2090222	29.742	ng/ul	99
83) Butylbenzylphthalate	20.29	149	859367	34.408	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.06	252	397598	16.339	ng/ul	98
85) Benzo(a)anthracene	21.12	228	2047859	29.344	ng/ul	100
86) Bis(2-ethylhexyl)phthalate	21.06	149	1385668	32.177	ng/ul	99
87) Chrysene	21.17	228	1998947	28.758	ng/ul	99
89) Di-n-octyl phthalate	21.92	149	2349144	30.414	ng/ul	100
90) Benzo(b)fluoranthene	22.65	252	2043179	31.207	ng/ul	100
91) Benzo(k)fluoranthene	22.69	252	1947946	28.300	ng/ul	99
93) Benzo(a)pyrene	23.20	252	1696691	28.534	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.45	276	1900067	26.401	ng/ul	99
95) Dibenzo(a,h)anthracene	25.46	278	1585945	25.997	ng/ul	99
96) Benzo(g,h,i)perylene	26.10	276	1115936	19.138	ng/ul	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050721\
Data File : BM029881.D
Acq On : 07 May 2021 17:56
Operator : CG/JU
Sample : M2129-02MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E0AB9MS


Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM

Quant Time: May 07 18:41:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 07 13:53:15 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.039 ng/uL

RT: 3.10 min Scan# 36

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampleId :

E0AB9MS

Tot Ion:	88	Resp:	29423
Ion	Ratio	Lower	Upper
88	100		
43	38.7	26.2	39.2
58	75.7	59.0	88.4

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#5

Pyridine

Concen: 15.137 ng/uL




RT: 3.50 min Scan# 105


Delta R.T. 0.01 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	79	Resp:	150144
Ion	Ratio	Lower	Upper
79	100		
52	83.9	74.6	111.8
51	39.1	32.6	48.8





#6

Benzaldehyde

Concen: 18.946 ng/ul

RT: 6.71 min Scan# 650

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampled :

E0AB9MS

Tot Ion:	77	Resp:	140291
Ion Ratio	Lower	Upper	
77	100		
105	97.0	77.7	116.5
106	93.9	72.2	108.4

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#8

Phenol

Concen: 20.588 ng/ul




RT: 6.77 min Scan# 660


Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	94	Resp:	301517
Ion Ratio	Lower	Upper	
94	100		
65	30.4	23.3	34.9
66	39.4	30.6	45.8









#10
Bis(2-Chloroethyl)ether
Concen: 19.037 ng/ul
RT: 6.99 min Scan# 698
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
93	100		220449		
63	83.9	66.0	99.0		
95	32.8	25.7	38.5		




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#12
2-Chlorophenol
Concen: 20.199 ng/ul
RT: 7.12 min Scan# 720
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
128	100		228070		
64	49.2	39.3	58.9		
130	32.8	26.6	39.8		









#13
 2-Methylphenol
 Concen: 20.019 ng/ul
 RT: 8.01 min Scan# 871
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Instrument :
 BNA_M
 ClientSampled :
 E0AB9MS

Tot Ion	Ratio	Lower	Upper
108	100		
107	90.0	73.8	110.6




Manual Integrations
 APPROVED


mohammad
 5/10/2021 11:45:14 AM



#14
 2,2'-oxybis(1-Chloropropane)
 Concen: 19.913 ng/ul
 RT: 8.09 min Scan# 884
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
45	100		
77	15.0	12.3	18.5
79	11.8	9.1	13.7









#16
Acetophenone
Concen: 19.731 ng/ul
RT: 8.38 min Scan# 934
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
105	100	375645		
77	79.3	64.2	96.4	
51	29.4	24.3	36.5	




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#17
N-Nitroso-di-n-propylamine
Concen: 20.746 ng/ul
RT: 8.37 min Scan# 932
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
70	100	202078		
42	63.1	49.4	74.0	
101	10.3	7.8	11.6	
130	20.8	16.5	24.7	









#18
4-Methylphenol
Concen: 20.188 ng/ul
RT: 8.33 min Scan# 926
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
108	100		
107	111.8	89.4	134.0




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#19
Hexachloroethane
Concen: 20.286 ng/ul
RT: 8.62 min Scan# 974
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
117	100		
201	96.9	79.4	119.2
199	58.3	48.0	72.0









#22
Nitrobenzene
Concen: 21.438 ng/ul
RT: 8.76 min Scan# 998
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
77	100		
123	42.5	35.0	52.6
65	12.7	10.3	15.5




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#23
Isophorone
Concen: 21.294 ng/ul
RT: 9.27 min Scan# 1086
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
82	100		
95	6.5	5.5	8.3
138	16.8	13.8	20.6









#25
2-Nitrophenol
Concen: 21.403 ng/ul
RT: 9.46 min Scan# 1117
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion:	139	Resp:	128864
Ion Ratio	Lower	Upper	
139	100		
65	55.4	44.6	67.0
109	28.7	24.4	36.6




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#26
2,4-Dimethylphenol
Concen: 20.989 ng/ul
RT: 9.53 min Scan# 1129
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion:	107	Resp:	285145
Ion Ratio	Lower	Upper	
107	100		
121	51.8	40.7	61.1
122	88.3	68.9	103.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.491 ng/ul

RT: 9.76 min Scan# 1169

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampled :

E0AB9MS

Tot Ion:	93	Resp:	325019
Ion	Ratio	Lower	Upper
93	100		
95	32.6	25.3	37.9
123	10.5	8.8	13.2

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#29

2,4-Dichlorophenol

Concen: 20.689 ng/ul




RT: 9.99 min Scan# 1208


Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	162	Resp:	225943
Ion	Ratio	Lower	Upper
162	100		
164	65.6	51.2	76.8
98	36.7	29.7	44.5









#30
Naphthalene
Concen: 20.117 ng/ul
RT: 10.38 min Scan# 1274
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.5	12.7
127	13.2	10.4	15.6




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#32
4-Chloroaniline
Concen: 17.204 ng/ul
RT: 10.50 min Scan# 1295
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.0	26.3	39.5









#33
Hexachlorobutadiene
Concen: 19.230 ng/ul
RT: 10.66 min Scan# 1322
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
225	100	140802		
223	62.2	53.0	79.4	
227	64.7	51.2	76.8	




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#34
Caprolactam
Concen: 27.023 ng/ul m
RT: 11.29 min Scan# 1428
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	104734		
55	180.8	138.9	208.3	
56	128.7	105.4	158.2	









#35
4-Chloro-3-methylphenol
Concen: 24.266 ng/ul
RT: 11.64 min Scan# 1488
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
107	100		
144	26.1	21.3	31.9
142	80.0	64.6	97.0




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#36
2-Methylnaphthalene
Concen: 20.416 ng/ul
RT: 12.00 min Scan# 1550
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.2	72.1	108.1









#37
 1-Methylnaphthalene
 Concen: 20.208 ng/ul
 RT: 12.22 min Scan# 1587
 Delta R.T. -0.01 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Instrument :
 BNA_M
 ClientSampleId :
 E0AB9MS

Tot Ion:	142	Resp:	574929
Ion Ratio	Lower	Upper	
142	100		
141	93.4	72.9	109.3
116	4.0	3.0	4.6




Manual Integrations
 APPROVED


mohammad
 5/10/2021 11:45:14 AM



#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 20.259 ng/ul
 RT: 12.38 min Scan# 1614
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Tgt Ion:	216	Resp:	295862
Ion Ratio	Lower	Upper	
216	100		
214	77.9	61.9	92.9
179	20.5	16.8	25.2
108	18.0	14.1	21.1









#40
Hexachlorocyclopentadiene
Concen: 13.039 ng/ul
RT: 12.35 min Scan# 1609
Delta R.T. -0.01 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
237	100				
235	60.7		50.0	75.0	
272	12.5		10.2	15.2	




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#41
2,4,6-Trichlorophenol
Concen: 23.165 ng/ul
RT: 12.63 min Scan# 1656
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
196	100				
198	97.6		77.0	115.4	
200	29.9		23.7	35.5	









#42
 2,4,5-Trichlorophenol
 Concen: 25.217 ng/ul
 RT: 12.70 min Scan# 1669
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Instrument :
 BNA_M
 ClientSampled :
 E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
196	100	237677		
198	95.6	75.3	112.9	
200	30.5	24.6	36.8	




Manual Integrations
 APPROVED


mohammad
 5/10/2021 11:45:14 AM



#43
 1,1'-Biphenyl
 Concen: 20.854 ng/ul
 RT: 13.03 min Scan# 1725
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	766144		
153	41.2	33.3	49.9	
76	13.2	10.3	15.5	






#44
2-Chloronaphthalene
Concen: 21.383 ng/ul
RT: 13.07 min Scan# 1731
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
162	100		
164	32.6	25.8	38.6
127	39.3	31.0	46.6

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM






Abundance

Ion 162.00 (161.70 to 162.70): E


Ion 164.00 (163.70 to 164.70): E

Ion 127.00 (126.70 to 127.70): E



#45
2-Nitroaniline
Concen: 27.721 ng/ul
RT: 13.29 min Scan# 1769
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
65	100		
92	66.1	53.2	79.8
138	99.9	81.1	121.7






Abundance

Ion 65.00 (64.70 to 65.70): BM0

Ion 92.00 (91.70 to 92.70): BM0

Ion 138.00 (137.70 to 138.70): E









#47
Dimethylphthalate
Concen: 26.561 ng/ul
RT: 13.67 min Scan# 1834
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
163	100	974252		
194	4.0		3.4	5.0
164	10.2		7.8	11.8




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#48
2,6-Dinitrotoluene
Concen: 27.209 ng/ul
RT: 13.79 min Scan# 1854
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	179163		
89	62.1		52.5	78.7
121	23.5		18.8	28.2





#50

Acenaphthylene

Concen: 21.177 ng/ul

RT: 13.92 min Scan# 1876

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampleId :

E0AB9MS

Tot Ion:	152	Resp:	950774
Ion Ratio	Lower	Upper	
152	100		
151	19.7	16.6	24.8
153	13.1	10.6	16.0

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#51

3-Nitroaniline

Concen: 24.659 ng/ul




RT: 14.13 min Scan# 1911


Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	138	Resp:	182402
Ion Ratio	Lower	Upper	
138	100		
108	10.9	8.5	12.7
92	123.9	100.1	150.1





#52

Acenaphthene

Concen: 21.760 ng/ul

RT: 14.27 min Scan# 1935

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampleId :

E0AB9MS

Tot Ion:	153	Resp:	690445
Ion	Ratio	Lower	Upper
153	100		
152	47.0	37.7	56.5
154	88.6	70.2	105.4

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#53

2,4-Dinitrophenol

Concen: 21.356 ng/ul




RT: 14.34 min Scan# 1947


Delta R.T. -0.01 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	184	Resp:	95678
Ion	Ratio	Lower	Upper
184	100		
63	73.5	57.2	85.8
154	55.6	45.8	68.6









#55
4-Nitrophenol
Concen: 27.275 ng/ul
RT: 14.45 min Scan# 1965
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
109	100	125121		
139	134.4	102.1	153.1	
65	147.5	114.6	172.0	




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#56
Dibenzofuran
Concen: 22.661 ng/ul
RT: 14.60 min Scan# 1992
Delta R.T. -0.01 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	984839		
139	38.1	29.8	44.6	





#57

2,4-Dinitrotoluene

Concen: 29.826 ng/ul

RT: 14.59 min Scan# 1989

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M





ClientSampled :

E0AB9MS

Tot Ion:	165	Resp:	290613
Ion Ratio	Lower	Upper	
165	100		
63	50.8	40.7	61.1
182	3.1	2.1	3.1

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM



#58

2,3,4,6-Tetrachlorophenol

Concen: 26.820 ng/ul




RT: 14.84 min Scan# 2032


Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Tgt Ion:	232	Resp:	240705
Ion Ratio	Lower	Upper	
232	100		
131	46.6	38.3	57.5
130	2.9	2.0	3.0
166	31.8	24.6	37.0









#59
Diethylphthalate
Concen: 26.907 ng/ul
RT: 15.05 min Scan# 2067
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampled :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
149	100		
177	21.5	17.2	25.8
150	12.5	10.0	15.0




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#61
Fluorene
Concen: 23.755 ng/ul
RT: 15.26 min Scan# 2103
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.7	78.4	117.6
167	13.4	10.9	16.3









#62
4-Chlorophenyl-phenylether
Concen: 23.083 ng/ul
RT: 15.26 min Scan# 2103
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
204	100		
206	31.8	26.2	39.2
141	57.9	47.2	70.8




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#63
4-Nitroaniline
Concen: 26.993 ng/ul m
RT: 15.30 min Scan# 2110
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
138	100		
92	55.7	44.8	67.2
108	80.7	65.3	97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 26.635 ng/ul
 RT: 15.35 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Instrument :
 BNA_M
 ClientSampleId :
 E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
198	100	168176		
182	4.4		4.3	6.5
77	27.5		23.0	34.4




Manual Integrations
 APPROVED


mohammad
 5/10/2021 11:45:14 AM



#67
 N-Nitrosodiphenylamine
 Concen: 25.912 ng/ul
 RT: 15.47 min Scan# 2140
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	831409		
168	66.3		53.0	79.4
167	35.2		27.8	41.6









#68
4-Bromophenyl-phenylether
Concen: 25.192 ng/ul
RT: 16.15 min Scan# 2255
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampled :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
248	100		
250	98.1	77.5	116.3
141	69.8	55.1	82.7




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#69
Hexachlorobenzene
Concen: 26.369 ng/ul
RT: 16.26 min Scan# 2273
Delta R.T. -0.01 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.2	26.4	39.6
249	31.1	24.5	36.7









#70
Atrazine
Concen: 27.217 ng/ul
RT: 16.43 min Scan# 2303
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
200	100	322194		
173	26.4		20.5	30.7
215	49.0		38.2	57.2




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#71
Pentachlorophenol
Concen: 25.005 ng/ul
RT: 16.61 min Scan# 2333
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	187059		
264	62.0		49.8	74.8
268	64.5		51.9	77.9





#72

Phenanthrene

Concen: 27.411 ng/ul

RT: 16.99 min Scan# 2398

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M

ClientSampleId :

E0AB9MS

Tot Ion:178 Resp: 1596417

Ion Ratio Lower Upper


178 100

179 15.4 12.2 18.4

176 18.7 15.0 22.4

Manual Integrations
APPROVED

mohammad
5/10/2021 11:45:14 AM






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#74

Anthracene

Concen: 27.436 ng/ul

RT: 17.09 min Scan# 2414

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56


Tgt Ion:178 Resp: 1633579

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.1 14.9 22.3






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.282 ng/uL
RT: 12.99 min Scan# 1718
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Resp	Lower	Upper
216	100	301259		
214	78.1	63.9	95.9	
218	48.6	39.7	59.5	




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#76
Pentachlorobenzene
Concen: 21.596 ng/uL
RT: 14.53 min Scan# 1979
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Resp	Lower	Upper
250	100	321779		
248	62.5	51.3	76.9	
252	63.8	51.5	77.3	









#77
 Carbazole
 Concen: 27.360 ng/ul
 RT: 17.36 min Scan# 2461
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Instrument :
 BNA_M
 ClientSampleId :
 E0AB9MS

Tot Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.6	26.4
139	12.5	10.2	15.2




Manual Integrations
 APPROVED


mohammad
 5/10/2021 11:45:14 AM



#78
 Di-n-butylphthalate
 Concen: 30.185 ng/ul
 RT: 17.93 min Scan# 2557
 Delta R.T. -0.00 min
 Lab File: BM029881.D
 Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.0
104	5.3	4.2	6.2









#80
Fluoranthene
Concen: 30.208 ng/ul
RT: 19.02 min Scan# 2742
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
202	100		
101	11.4	9.0	13.4
100	8.8	7.0	10.6




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#82
Pyrene
Concen: 29.742 ng/ul
RT: 19.37 min Scan# 2803
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
202	100		
101	14.2	11.8	17.6
100	11.7	9.2	13.8









#83
Butylbenzylphthalate
Concen: 34.408 ng/ul
RT: 20.29 min Scan# 2958
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
149	100		
91	74.5	60.6	91.0
206	20.0	15.4	23.0




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#84
3,3'-Dichlorobenzidine
Concen: 16.339 ng/ul
RT: 21.06 min Scan# 3090
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.9	51.0	76.6
126	12.0	9.4	14.0









#85
Benzo(a)anthracene
Concen: 29.344 ng/ul
RT: 21.12 min Scan# 3100
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
228	100		
229	19.6	15.9	23.9
226	26.9	21.4	32.0




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#86
Bis(2-ethylhexyl)phthalate
Concen: 32.177 ng/ul
RT: 21.06 min Scan# 3090
Delta R.T. -0.01 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.8	21.1	31.7
279	3.8	3.5	5.3









#87
Chrysene
Concen: 28.758 ng/ul
RT: 21.17 min Scan# 3109
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.6	23.9	35.9
229	19.7	15.6	23.4




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#89
Di-n-octyl phthalate
Concen: 30.414 ng/ul
RT: 21.92 min Scan# 3236
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion:149 Resp: 2349144









#90
Benzo(b)fluoranthene
Concen: 31.207 ng/ul
RT: 22.65 min Scan# 3360
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	11.0	8.6	12.8




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#91
Benzo(k)fluoranthene
Concen: 28.300 ng/ul
RT: 22.69 min Scan# 3367
Delta R.T. -0.01 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	11.1	8.4	12.6









#93
Benzo(a)pyrene
Concen: 28.534 ng/ul
RT: 23.20 min Scan# 3454
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Instrument :
BNA_M
ClientSampleId :
E0AB9MS

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.9	26.9
125	11.8	9.6	14.4




Manual Integrations
APPROVED


mohammad
5/10/2021 11:45:14 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 26.401 ng/ul
RT: 25.45 min Scan# 3836
Delta R.T. -0.00 min
Lab File: BM029881.D
Acq: 07 May 2021 17:56

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.0	21.1	31.7
277	25.3	20.3	30.5





#95

Dibenzo(a,h)anthracene

Concen: 25.997 ng/ul

RT: 25.46 min Scan# 3837

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56

Instrument :

BNA_M

ClientSampleId :

E0AB9MS

Tot Ion:278 Resp: 1585945

Ion Ratio Lower Upper

278 100


139 17.4 13.5 20.3

279 22.8 18.6 28.0

Manual Integrations
APPROVED

mohammad

5/10/2021 11:45:14 AM






Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#96

Benzo(g,h,i)perylene

Concen: 19.138 ng/ul

RT: 26.10 min Scan# 3947

Delta R.T. -0.00 min

Lab File: BM029881.D

Acq: 07 May 2021 17:56


Tgt Ion:276 Resp: 1115936

Ion Ratio Lower Upper

276 100

138 22.7 17.8 26.8

277 23.5 18.5 27.7



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

