

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050819\
 Data File : BM020197.D
 Acq On : 08 May 2019 16:31
 Operator : JU/SJ
 Sample : PB119288BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB119532BL

Quant Time: May 09 02:06:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:58:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	139773	20.00	ng	0.00
21) Naphthalene-d8	10.58	136	488956	20.00	ng	0.00
39) Acenaphthene-d10	14.43	164	298562	20.00	ng	0.00
64) Phenanthrene-d10	17.17	188	662678	20.00	ng	0.00
76) Chrysene-d12	21.36	240	566365	20.00	ng	0.00
87) Perylene-d12	23.65	264	646864	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	1224170	143.16	ng	0.00
7) Phenol-d6	6.96	99	1674173	143.31	ng	0.00
23) Nitrobenzene-d5	8.95	82	1181642	95.41	ng	0.00
42) 2,4,6-Tribromophenol	15.92	330	701159	151.30	ng	0.00
45) 2-Fluorobiphenyl	13.05	172	2391148	98.54	ng	-0.01
79) Terphenyl-d14	19.80	244	3436890	105.84	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050819\
Data File : BM020197.D
Acq On : 08 May 2019 16:31
Operator : JU/SJ
Sample : PB119288BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB119532BL

Quant Time: May 09 02:06:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:58:37 2019
Response via : Initial Calibration

