

Data Path : Z:\HPCHEM1\BNA M\DATA\BM051016\
 Data File : BM005374.D
 Acq On : 10 May 2016 14:30
 Operator : UM/SJ
 Sample : H2888-13
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9WX2

Quant Time: May 11 07:57:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 11 07:57:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	67266	20.00	ng/ul	0.00
18) Naphthalene-d8	10.55	136	313090	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.40	164	208244	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.15	188	510618	20.00	ng/ul	0.00
75) Chrysene-d12	21.34	240	677564	20.00	ng/ul	0.00
83) Perylene-d12	23.61	264	681257	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.26	96	5621	3.93	ng/uL	0.00
5) Phenol-d5	6.93	99	127461	20.89	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	76990	22.12	ng/ul	0.00
9) 2-Chlorophenol-d4	7.29	132	100079	21.72	ng/ul	0.00
13) 4-Methylphenol-d8	8.46	113	107831	21.38	ng/ul	0.00
19) Nitrobenzene-d5	8.92	128	52285	23.39	ng/ul	0.00
22) 2-Nitrophenol-d4	9.63	143	61024	24.11	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.17	165	103156	21.93	ng/ul	0.00
29) 4-Chloroaniline-d4	10.69	131	149923	26.48	ng/ul	0.00
43) Dimethylphthalate-d6	13.81	166	400472	23.99	ng/ul	0.00
46) Acenaphthylene-d8	14.09	160	459815	23.49	ng/ul	0.00
51) 4-Nitrophenol-d4	14.62	143	69840	22.92	ng/ul	0.00
57) Fluorene-d10	15.40	176	345318	23.96	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.53	200	59621	20.76	ng/ul	0.00
70) Anthracene-d10	17.25	188	565472	25.05	ng/ul	0.00
76) Pyrene-d10	19.55	212	709795	22.69	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.47	264	768295	25.48	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	13.86	163	112990	6.77	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM051016\
Data File : BM005374.D
Acq On : 10 May 2016 14:30
Operator : UM/SJ
Sample : H2888-13
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9WX2

Quant Time: May 11 07:57:58 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM050516.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 11 07:57:14 2016
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 6.77 ng/ul

RT: 13.86 min Scan# 1899

Delta R.T. -0.01 min

Lab File: BM005374.D

Acq: 10 May 2016 14:30

Instrument :

BNA_M

ClientSampleId :

D9WX2


Tot Ion:163 Resp: 112990

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.0

164 9.9 7.9 11.9



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

