

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051021\
 Data File : BM029978.D
 Acq On : 12 May 2021 13:08
 Operator : CG/JU
 Sample : PB136093BL
 Misc :
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK093

Quant Time: May 12 13:42:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 11 15:21:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	151508	20.00	ng/ul	-0.01
20) Naphthalene-d8	10.32	136	663800	20.00	ng/ul	-0.01
38) Acenaphthene-d10	14.19	164	435273	20.00	ng/ul	-0.01
64) Phenanthrene-d10	16.94	188	844058	20.00	ng/ul	0.00
79) Chrysene-d12	21.13	240	657653	20.00	ng/ul	0.00
88) Perylene-d12	23.28	264	648142	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.06	96	20145	5.31	ng/uL	0.00
4) Pyridine-d5	3.48	84	220342	23.20	ng/ul	0.00
7) Phenol-d5	6.73	99	343391	24.37	ng/ul	-0.01
9) Bis-(2-Chloroethyl)ether-d	6.89	67	237370	27.17	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.08	132	298781	27.48	ng/ul	0.00
15) 4-Methylphenol-d8	8.26	113	287729	24.69	ng/ul	-0.01
21) Nitrobenzene-d5	8.70	128	136553	29.85	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.42	143	151779	29.40	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.95	165	252055	24.35	ng/ul	0.00
31) 4-Chloroaniline-d4	10.47	131	380474	24.15	ng/ul	0.00
46) Dimethylphthalate-d6	13.61	166	968346	28.70	ng/ul	-0.01
49) Acenaphthylene-d8	13.88	160	1159755	27.12	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.43	143	74176	13.50	ng/ul	0.00
60) Fluorene-d10	15.19	176	813447	28.40	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.33	200	108797	19.84	ng/ul	0.00
73) Anthracene-d10	17.03	188	1225243	28.83	ng/ul	-0.01
81) Pyrene-d10	19.33	212	1242907	36.84	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.14	264	1091222	29.70	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM051021\
Data File : BM029978.D
Acq On : 12 May 2021 13:08
Operator : CG/JU
Sample : PB136093BL
Misc :
ALS Vial : 77 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK093

Quant Time: May 12 13:42:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 11 15:21:38 2021
Response via : Initial Calibration

