

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051021\
 Data File : BM030012.D
 Acq On : 13 May 2021 15:34
 Operator : CG/JU
 Sample : M2336-01
 Misc :
 ALS Vial : 112 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BG6K9

Quant Time: May 13 16:05:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 13 14:02:26 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	144918	20.00	ng/ul	0.00
20) Naphthalene-d8	10.31	136	669201	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.19	164	448375	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.93	188	953562	20.00	ng/ul	0.00
79) Chrysene-d12	21.12	240	1002541	20.00	ng/ul	0.00
88) Perylene-d12	23.27	264	941619	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.07	96	4612	1.27	ng/uL	0.01
4) Pyridine-d5	0.00	84	0d	0.00	ng/ul	
7) Phenol-d5	6.73	99	73251	5.44	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.89	67	201670	24.13	ng/ul	0.00
11) 2-Chlorophenol-d4	7.08	132	222115	21.35	ng/ul	0.00
15) 4-Methylphenol-d8	8.26	113	127443	11.43	ng/ul	0.00
21) Nitrobenzene-d5	8.70	128	124861	27.08	ng/ul	0.00
24) 2-Nitrophenol-d4	9.41	143	141092	27.11	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.95	165	245028	23.48	ng/ul	0.00
31) 4-Chloroaniline-d4	10.43	131	120	0.01	ng/ul	-0.03
46) Dimethylphthalate-d6	13.61	166	1035368	29.79	ng/ul	0.00
49) Acenaphthylene-d8	13.87	160	1182999	26.86	ng/ul	0.00
54) 4-Nitrophenol-d4	14.45	143	18108	3.20	ng/ul	0.03
60) Fluorene-d10	15.19	176	876181	29.70	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.33	200	167228	27.00	ng/ul	0.00
73) Anthracene-d10	17.03	188	1437959	29.95	ng/ul	0.00
81) Pyrene-d10	19.33	212	1771025	34.44	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.14	264	1565579	29.33	ng/ul	0.00
Target Compounds						
47) Dimethylphthalate	13.66	163	41651	1.197	ng/ul	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051021\
Data File : BM030012.D
Acq On : 13 May 2021 15:34
Operator : CG/JU
Sample : M2336-01
Misc :
ALS Vial : 112 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BG6K9

Quant Time: May 13 16:05:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM050721.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 13 14:02:26 2021
Response via : Initial Calibration

