

Method Path : Z:\HPCHEM1\BNA\_M\METHODS\  
 Method File : SOM02.2-EPA-SIM-BM051115.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue May 12 03:03:15 2015  
 Response Via : Initial Calibration

## Calibration Files

0.1 =BM001239.D 0.2 =BM001240.D 0.4 =BM001241.D  
 0.8 =BM001242.D 1.6 =BM001243.D 3.2 =BM001244.D

|          | Compound              | 0.1            | 0.2   | 0.4   | 0.8   | 1.6   | 3.2   | Avg   | %RSD |
|----------|-----------------------|----------------|-------|-------|-------|-------|-------|-------|------|
| -----    |                       |                |       |       |       |       |       |       |      |
| 1) I     | 1,4-Dichlorobenzene-d | -----ISTD----- |       |       |       |       |       |       |      |
| 2) I     | Naphthalene-d8        | -----ISTD----- |       |       |       |       |       |       |      |
| 3)       | Naphthalene           | 1.042          | 1.003 | 1.027 | 1.009 | 0.978 |       | 1.012 | 2.43 |
| 4) SURR2 | Methylnaphthale       | 0.514          | 0.497 | 0.506 | 0.517 | 0.506 |       | 0.508 | 1.53 |
| 5)       | 2-Methylnaphthale     | 0.675          | 0.653 | 0.656 | 0.671 | 0.664 |       | 0.664 | 1.43 |
| 6) I     | Acenaphthene-d10      | -----ISTD----- |       |       |       |       |       |       |      |
| 7)       | Acenaphthylene        | 1.752          | 1.726 | 1.751 | 1.766 | 1.832 |       | 1.766 | 2.26 |
| 8) C     | Acenaphthene          | 1.406          | 1.395 | 1.389 | 1.377 | 1.391 |       | 1.392 | 0.76 |
| 9)       | Fluorene              | 1.587          | 1.503 | 1.551 | 1.537 | 1.584 |       | 1.552 | 2.26 |
| 10) I    | Phenanthrene-d10      | -----ISTD----- |       |       |       |       |       |       |      |
| 11)      | Pentachlorophenol     |                | 0.084 | 0.079 | 0.082 | 0.087 | 0.101 | 0.087 | 9.80 |
| 12)      | Phenanthrene          | 1.168          | 1.106 | 1.124 | 1.155 | 1.134 |       | 1.137 | 2.15 |
| 13)      | Anthracene            | 1.048          | 1.003 | 1.028 | 1.067 | 1.070 |       | 1.043 | 2.71 |
| 14) SURR | Fluoranthene-d10      | 0.898          | 0.862 | 0.867 | 0.881 | 0.859 |       | 0.873 | 1.86 |
| 15) C    | Fluoranthene          | 1.276          | 1.223 | 1.261 | 1.272 | 1.278 |       | 1.262 | 1.80 |
| 16) I    | Chrysene-d12          | -----ISTD----- |       |       |       |       |       |       |      |
| 17)      | Pyrene                | 1.643          | 1.545 | 1.568 | 1.529 | 1.501 |       | 1.557 | 3.45 |
| 18)      | Benzo(a)anthracen     | 1.234          | 1.196 | 1.205 | 1.243 | 1.232 |       | 1.222 | 1.68 |
| 19)      | Chrysene              | 1.447          | 1.354 | 1.442 | 1.380 | 1.343 |       | 1.393 | 3.50 |
| 20) I    | Perylene-d12          | -----ISTD----- |       |       |       |       |       |       |      |
| 21)      | Benzo(b)fluoranth     | 1.378          | 1.316 | 1.440 | 1.438 | 1.636 |       | 1.442 | 8.33 |
| 22)      | Benzo(k)fluoranth     | 1.525          | 1.497 | 1.514 | 1.548 | 1.341 |       | 1.485 | 5.57 |
| 23) C    | Benzo(a)pyrene        | 1.418          | 1.351 | 1.361 | 1.383 | 1.389 |       | 1.381 | 1.90 |
| 24)      | Indeno(1,2,3-cd)p     | 1.450          | 1.499 | 1.547 | 1.555 | 1.561 |       | 1.522 | 3.11 |
| 25)      | Dibenzo(a,h)anthr     | 1.142          | 1.184 | 1.238 | 1.243 | 1.257 |       | 1.213 | 3.98 |
| 26)      | Benzo(g,h,i)peryl     | 1.303          | 1.355 | 1.395 | 1.375 | 1.364 |       | 1.358 | 2.53 |
| -----    |                       |                |       |       |       |       |       |       |      |

(#) = Out of Range