


Data Path : Z:\HPCHEM1\BNA_M\Data\BM051115\
 Data File : BM001238.D
 Acq On : 11 May 2015 09:51
 Operator : TP/IZ
 Sample : DFTPP54
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP54

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM051115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 12 03:03:15 2015



AutoFind: Scans 2492, 2493, 2494; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.2	230749	PASS
68	69	0.00	2	1.7	3491	PASS
69	198	0.00	100	45.1	203477	PASS
70	69	0.00	2	0.5	1042	PASS
127	198	10	80	52.5	236800	PASS
197	198	0.00	2	0.4	1645	PASS
198	198	100	100	100.0	450858	PASS
199	198	5	9	6.8	30842	PASS
275	198	10	60	24.3	109373	PASS
365	198	1	100	2.7	12250	PASS
441	443	0.01	100	78.4	55386	PASS
442	198	50	100	79.6	358805	PASS
443	442	15	24	19.7	70608	PASS

m/z	Abundance
36.00	701.0
37.00	1329.0
38.00	2375.0
39.00	15257.0
40.00	2157.0
41.10	1308.0
42.00	560.0
43.00	1795.0
44.00	8336.0
45.00	1816.0
47.00	3363.0
48.10	1210.0
49.00	15576.0
50.10	50424.0
51.00	192384.0
52.10	10839.0
53.00	448.0
55.00	1185.0
56.00	5804.0
57.00	13065.0
58.00	766.0
60.10	696.0
61.00	2382.0
62.10	2289.0
63.00	6500.0
63.90	852.0
65.00	3015.0
67.10	345.0
68.00	2717.0
69.00	166144.0
70.00	859.0
71.10	404.0
73.00	1438.0
74.00	18896.0
75.00	27200.0
76.00	8066.0
77.10	180928.0
78.10	12698.0
79.00	13093.0
79.90	9606.0
81.00	13739.0
82.00	3238.0
83.10	3168.0
83.90	8432.0
85.00	2209.0
86.00	9938.0
87.00	1549.0
87.90	1110.0
91.00	3186.0
92.00	3578.0
93.00	21248.0
94.00	1514.0
94.80	416.0
95.80	724.0
96.10	722.0
97.10	444.0
98.00	17224.0
99.00	12557.0
100.10	1108.0
101.00	7847.0
102.10	525.0
102.90	2023.0
104.00	3682.0
105.00	4128.0
105.90	1481.0
107.00	55160.0
108.00	8652.0
109.10	1188.0
110.00	92408.0
111.00	14927.0
112.00	1909.0
112.90	452.0
116.00	2567.0
117.00	48808.0
118.00	3594.0
119.00	300.0
119.90	691.0
122.00	2989.0
123.00	6201.0
124.00	2303.0
124.90	2343.0
125.80	328.0
126.10	444.0
127.00	181312.0
128.00	13228.0
129.00	71472.0
130.00	6077.0
131.20	1041.0
132.10	771.0
134.00	1671.0
135.00	6484.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

136.10	2174.0
137.00	2664.0
137.90	530.0
140.00	870.0
141.10	8996.0
141.90	2684.0
143.00	2237.0
144.00	435.0
145.00	493.0
146.00	1977.0
147.00	4296.0
148.00	11228.0
149.00	2477.0
149.90	804.0
151.00	1188.0
152.10	382.0
153.00	3012.0
154.00	1694.0
155.10	5724.0
156.00	7449.0
157.20	1238.0
157.90	1671.0
158.90	1152.0
159.90	2947.0
161.00	4592.0
162.00	1438.0
162.90	491.0
164.00	613.0
165.00	3550.0
166.00	2280.0
167.00	19112.0
168.00	9690.0
169.00	1810.0
170.00	985.0
171.00	761.0
171.90	1445.0
173.00	2121.0
174.00	3780.0
175.00	6586.0
176.00	2002.0
177.10	2953.0
178.00	1667.0
178.90	13092.0
180.00	9759.0
180.90	3850.0
182.10	540.0
182.90	360.0
184.10	1177.0
185.00	5847.0
186.00	44624.0
187.00	12697.0
188.00	1902.0
189.00	3106.0
190.00	399.0
191.00	1792.0
191.90	3683.0
193.00	4284.0
193.90	912.0
195.10	366.0
196.00	9114.0
196.80	1767.0
198.00	310720.0
199.00	21192.0
200.00	1652.0
201.50	1508.0
203.00	1969.0
204.00	11403.0
205.10	19264.0
206.00	80808.0
207.00	10975.0
208.00	2680.0
209.00	1252.0
210.20	1279.0
210.90	3603.0
211.70	1038.0
215.00	1121.0
216.00	2063.0
217.00	22352.0
218.10	2862.0
219.10	379.0
221.10	12228.0
222.00	2666.0
223.00	4791.0
224.00	49968.0
225.10	12235.0
226.10	1802.0
227.00	18344.0
228.00	2980.0
228.90	3870.0
230.00	630.0
231.00	1375.0
231.90	358.0
234.00	1045.0
234.90	1370.0
235.90	757.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

237.00	2834.0
237.90	310.0
239.00	818.0
240.00	749.0
240.90	1092.0
242.00	2914.0
243.10	2216.0
244.00	33056.0
245.10	4766.0
246.10	6797.0
247.00	1999.0
247.80	353.0
249.00	1029.0
251.10	468.0
252.10	304.0
253.00	637.0
254.10	1068.0
255.00	166144.0
256.00	25392.0
256.90	2407.0
258.00	9114.0
258.90	1826.0
261.10	405.0
265.00	3686.0
266.00	604.0
271.00	424.0
271.90	360.0
273.00	5675.0
274.00	11492.0
275.00	73208.0
276.00	9756.0
277.00	6066.0
277.90	1268.0
283.00	803.0
284.10	502.0
285.00	897.0
293.00	1641.0
294.00	369.0
294.90	446.0
296.00	17192.0
297.10	2456.0
301.90	328.0
303.00	1932.0
303.90	688.0
308.10	301.0
314.00	789.0
315.00	2404.0
315.90	1196.0
321.00	448.0
323.10	5914.0
324.00	925.0
326.90	1176.0
328.00	844.0
332.10	341.0
332.90	366.0
334.00	3257.0
335.10	836.0
341.20	634.0
346.00	1287.0
352.00	1524.0
352.90	841.0
354.10	1505.0
354.90	395.0
365.00	7664.0
366.00	922.0
372.10	2195.0
373.00	640.0
383.00	941.0
389.90	470.0
390.90	334.0
402.00	1291.0
403.00	1481.0
404.00	680.0
421.00	1432.0
422.00	1681.0
423.00	9877.0
424.10	4083.0
425.20	545.0
441.00	30352.0
442.00	188992.0
443.00	38544.0
444.00	3656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

m/z	Abundance
36.10	591.0
37.10	1213.0
38.00	4114.0
39.10	21520.0
40.00	2699.0
41.10	1912.0
42.10	745.0
43.00	2146.0
44.00	8066.0
45.00	1534.0
47.00	2948.0
48.00	1410.0
49.00	16656.0
50.00	70552.0
51.10	281216.0
52.00	15492.0
53.10	515.0
55.00	1470.0
56.00	7714.0
57.00	18280.0
58.10	741.0
59.10	381.0
60.00	825.0
61.00	3439.0
62.00	3456.0
63.00	11503.0
64.00	1681.0
65.00	4961.0
66.00	442.0
68.00	4031.0
69.00	242240.0
70.10	1022.0
71.00	433.0
73.10	2381.0
74.00	24296.0
75.00	37984.0
76.10	12908.0
77.10	268608.0
78.10	18944.0
79.00	17176.0
80.00	13654.0
81.00	19944.0
82.00	5880.0
83.10	4867.0
83.90	8087.0
85.00	4206.0
86.00	11533.0
87.00	2603.0
88.00	1892.0
89.00	519.0
91.00	3997.0
92.00	5102.0
93.00	32248.0
93.90	2523.0
95.10	507.0
96.10	1406.0
97.10	697.0
98.00	24032.0
99.00	19152.0
100.00	1729.0
101.00	11547.0
101.80	451.0
103.00	2816.0
104.00	6711.0
105.00	6314.0
106.10	1833.0
107.00	81664.0
108.00	11919.0
109.10	2586.0
110.00	143488.0
111.00	23200.0
111.90	2233.0
112.90	837.0
115.00	401.0
116.00	4725.0
117.00	70840.0
118.00	5236.0
119.00	724.0
120.10	1038.0
121.00	430.0
122.10	5481.0
123.00	8722.0
124.00	4125.0
125.00	3524.0
126.10	1274.0
127.00	275328.0
128.00	22032.0
129.00	106064.0
130.00	8431.0
131.10	1896.0
132.00	1178.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

133.00	651.0
134.00	3236.0
135.00	8729.0
136.00	3548.0
137.00	5303.0
138.00	783.0
139.00	830.0
140.00	1392.0
141.00	15204.0
141.90	4316.0
143.00	3160.0
144.00	727.0
144.90	1078.0
146.00	2283.0
146.90	7114.0
148.00	16960.0
149.00	3191.0
150.10	967.0
151.00	1749.0
152.00	882.0
153.00	4183.0
154.00	2810.0
155.00	7605.0
156.00	11438.0
157.10	2062.0
157.90	2519.0
159.00	2284.0
160.00	4240.0
161.00	6004.0
162.00	1913.0
162.80	554.0
163.80	451.0
164.20	401.0
165.00	5562.0
165.90	4241.0
167.00	29160.0
168.00	14594.0
169.00	2466.0
169.90	926.0
170.90	1249.0
172.00	2820.0
173.00	4013.0
174.10	4788.0
175.00	10411.0
176.00	3164.0
177.00	4397.0
178.00	2092.0
179.00	21192.0
180.00	14611.0
181.00	6569.0
182.20	1028.0
183.10	536.0
184.00	2216.0
185.00	8654.0
186.00	74440.0
187.00	20944.0
188.00	2643.0
189.00	4969.0
189.90	957.0
191.00	1554.0
192.00	6920.0
193.00	7017.0
194.00	1716.0
194.90	1020.0
196.00	14014.0
197.00	2207.0
198.00	528896.0
199.00	37128.0
200.00	3130.0
201.40	2094.0
202.20	480.0
203.00	3534.0
204.00	18160.0
205.00	33304.0
206.00	140160.0
207.00	17968.0
208.00	4571.0
209.00	1928.0
210.30	2917.0
211.00	5476.0
211.80	1312.0
213.00	392.0
214.90	1179.0
216.00	2678.0
217.00	37864.0
218.00	5562.0
218.90	443.0
221.10	23360.0
221.80	4957.0
223.00	7302.0
224.00	75792.0
225.00	19352.0
226.00	2865.0
227.00	30528.0
228.00	4130.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

229.00	6970.0
230.00	1104.0
231.00	2835.0
231.90	446.0
232.90	424.0
233.90	2379.0
235.00	2095.0
235.80	1224.0
237.00	3709.0
237.90	386.0
239.00	1203.0
240.00	1414.0
241.10	1528.0
241.90	3763.0
243.00	4000.0
244.00	56840.0
245.00	6922.0
246.00	10558.0
247.00	2531.0
247.90	462.0
249.00	2164.0
250.00	392.0
251.10	450.0
252.00	520.0
252.90	1215.0
254.00	1712.0
255.00	290368.0
256.00	43320.0
257.00	3968.0
258.00	15774.0
259.00	2713.0
260.10	629.0
261.00	620.0
264.00	549.0
265.00	6587.0
266.10	777.0
270.90	397.0
271.90	1048.0
273.00	7918.0
274.00	20872.0
275.00	122816.0
276.10	15039.0
277.00	9748.0
278.00	2177.0
279.10	318.0
282.00	496.0
282.90	1169.0
284.00	576.0
285.10	1491.0
289.00	456.0
290.00	491.0
290.90	399.0
291.90	428.0
293.00	2179.0
294.10	670.0
295.10	624.0
296.00	31488.0
297.00	4754.0
297.90	316.0
301.00	369.0
302.10	362.0
303.00	3853.0
304.10	1135.0
308.90	348.0
310.00	459.0
314.00	1552.0
315.00	3257.0
316.10	1815.0
317.10	417.0
320.90	937.0
322.10	819.0
323.10	10676.0
324.00	2143.0
326.00	362.0
327.00	2573.0
328.00	1038.0
331.90	732.0
333.00	1081.0
334.10	7191.0
335.00	1909.0
341.00	1049.0
341.90	419.0
346.00	2048.0
346.90	374.0
350.80	332.0
352.10	2990.0
352.90	1652.0
354.00	3069.0
354.90	555.0
358.90	300.0
365.00	13369.0
366.00	1853.0
371.00	765.0
372.00	5426.0
372.90	1314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

382.90	1401.0
384.00	468.0
390.00	885.0
391.10	550.0
401.10	328.0
401.90	2304.0
403.10	2510.0
404.00	1314.0
421.00	3017.0
422.00	2424.0
423.00	18608.0
424.00	5974.0
425.00	929.0
441.00	58264.0
442.00	390400.0
443.00	77720.0
444.00	7057.0
445.10	357.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

m/z	Abundance
35.90	494.0
37.00	1589.0
38.00	3071.0
39.00	16544.0
39.90	2014.0
41.00	1489.0
42.00	658.0
43.10	1690.0
44.00	8681.0
45.00	1911.0
47.00	3445.0
48.00	1462.0
49.00	16297.0
50.00	61024.0
51.10	233408.0
52.10	12234.0
53.10	640.0
55.00	1209.0
56.00	5910.0
57.00	15447.0
58.00	723.0
60.00	764.0
61.00	3039.0
62.00	3418.0
63.00	7929.0
64.10	1158.0
65.00	4722.0
65.90	370.0
68.00	3726.0
69.00	202048.0
70.00	1246.0
71.10	301.0
73.10	2242.0
74.00	20816.0
75.00	32736.0
76.10	9889.0
77.10	227712.0
78.10	14686.0
79.00	15503.0
80.00	13098.0
81.00	17136.0
82.00	5002.0
83.10	3981.0
84.00	8585.0
85.10	3043.0
86.00	9289.0
87.00	2370.0
87.90	1463.0
89.10	392.0
91.00	3651.0
92.00	4254.0
93.00	28344.0
94.00	1428.0
94.80	438.0
95.90	1040.0
96.90	520.0
98.00	21544.0
99.00	17624.0
100.10	1168.0
101.00	9160.0
102.10	570.0
103.00	2922.0
104.00	5863.0
105.00	5881.0
106.10	1924.0
107.00	71704.0
108.00	10234.0
109.00	2583.0
110.00	130728.0
111.00	18000.0
111.90	2413.0
113.00	804.0
115.00	570.0
116.10	3507.0
117.00	65432.0
118.00	4284.0
118.90	420.0
119.90	1046.0
121.20	382.0
122.00	4363.0
123.00	7226.0
124.00	3293.0
125.00	4232.0
126.10	987.0
127.00	253760.0
128.00	19248.0
129.00	97384.0
130.00	8312.0
131.00	1772.0
131.90	1053.0
133.10	477.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

134.00	2937.0
135.00	8039.0
136.00	3520.0
137.00	4049.0
137.90	1167.0
139.00	416.0
140.00	1200.0
141.00	11667.0
142.00	4147.0
142.90	2977.0
144.10	1012.0
145.00	860.0
146.00	2426.0
147.00	6499.0
148.00	16872.0
149.00	3365.0
150.00	995.0
150.90	1361.0
151.70	1147.0
152.20	1067.0
153.00	4348.0
153.90	3212.0
155.00	7897.0
156.00	10472.0
157.00	2141.0
158.00	2189.0
159.00	2043.0
160.00	3820.0
161.00	6707.0
162.00	1953.0
162.90	891.0
163.90	634.0
165.00	5502.0
166.00	4005.0
167.00	25392.0
168.00	12390.0
169.00	2048.0
170.10	683.0
170.90	1175.0
171.90	2091.0
173.00	3672.0
174.00	5623.0
175.00	10396.0
176.00	3201.0
177.00	4895.0
177.80	1200.0
179.00	20560.0
180.00	14066.0
181.00	5283.0
181.90	974.0
182.20	1000.0
182.90	815.0
184.00	1987.0
185.00	9376.0
186.00	71328.0
187.00	18632.0
188.00	2524.0
189.00	4858.0
190.00	608.0
191.00	2109.0
192.00	6691.0
193.00	6454.0
194.10	1268.0
195.10	1338.0
196.00	14784.0
196.80	3168.0
198.00	512960.0
199.00	34208.0
200.00	2099.0
201.30	2162.0
203.00	3989.0
204.00	19080.0
205.00	33040.0
206.00	138048.0
207.00	18312.0
208.00	4598.0
209.10	1500.0
210.10	2570.0
210.90	4490.0
213.00	599.0
214.90	1500.0
216.10	2589.0
217.00	37168.0
218.00	4630.0
219.00	463.0
221.00	22104.0
221.90	4608.0
223.00	8142.0
224.00	74600.0
225.00	19184.0
226.10	2484.0
227.00	32440.0
228.00	5187.0
229.00	6794.0
230.10	778.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

231.10	2950.0
231.90	568.0
234.10	2157.0
234.90	2396.0
235.90	1552.0
237.00	2708.0
238.00	544.0
239.00	1391.0
239.90	1054.0
240.90	1943.0
242.00	4288.0
243.10	4456.0
244.00	60160.0
245.10	7778.0
246.00	11194.0
247.00	2622.0
248.00	664.0
249.00	2322.0
250.00	496.0
251.00	663.0
252.10	761.0
253.00	1418.0
254.00	1748.0
255.00	302336.0
256.00	43016.0
257.00	3392.0
258.00	16872.0
259.00	2569.0
260.10	550.0
260.90	553.0
264.00	453.0
265.00	6647.0
266.10	1198.0
270.80	365.0
271.80	864.0
273.00	8538.0
274.00	22424.0
275.00	132096.0
276.00	17240.0
277.00	10972.0
277.90	1954.0
279.20	317.0
282.80	864.0
284.00	1079.0
285.00	1907.0
286.00	475.0
288.80	688.0
289.90	403.0
291.10	301.0
292.10	366.0
293.00	2577.0
293.90	733.0
295.10	752.0
296.00	34040.0
297.00	5189.0
298.00	361.0
301.00	490.0
302.10	492.0
303.10	4289.0
304.00	1006.0
308.00	378.0
309.00	333.0
310.10	456.0
312.90	337.0
314.00	1624.0
315.00	4115.0
316.00	2169.0
317.10	491.0
321.10	743.0
321.80	324.0
323.10	10258.0
324.10	2299.0
325.10	326.0
326.90	2317.0
328.00	1193.0
331.90	799.0
333.00	1399.0
334.10	7178.0
335.00	1795.0
341.00	1312.0
342.00	666.0
345.90	2643.0
346.90	620.0
350.90	312.0
352.10	2718.0
353.00	2705.0
354.00	4595.0
354.90	716.0
359.00	363.0
364.90	15718.0
366.00	2438.0
370.00	334.0
371.00	786.0
372.10	5978.0
373.00	1509.0

Instrument :

BNA_M

ClientSampleId :

DFTPP54

383.00	1654.0
384.10	650.0
390.00	834.0
390.90	733.0
401.00	473.0
402.00	3207.0
403.00	3866.0
404.00	1266.0
421.00	3066.0
422.00	2846.0
423.00	24776.0
424.00	6821.0
425.00	1107.0
441.00	77544.0
442.00	497024.0
443.00	95560.0
444.10	9323.0
444.90	697.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54

m/z	Abundance
36.00	559.0
37.00	466.0
40.00	1338.0
41.00	896.0
42.00	1097.0
43.10	1770.0
44.00	8175.0
44.90	1219.0
47.00	3103.0
48.00	1476.0
49.00	14514.0
50.10	681.0
51.00	4920.0
55.00	568.0
55.90	367.0
57.10	586.0
60.00	552.0
72.80	344.0
83.10	419.0
83.90	7885.0
86.00	5518.0
86.90	343.0
87.80	1017.0
167.00	535.0
207.00	397.0

Instrument :
BNA_M
ClientSampleId :
DFTPP54