

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051122\
 Data File : BM034985.D
 Acq On : 11 May 2022 15:06
 Operator : CG/JU
 Sample : N2603-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW484

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/12/2022
 Supervised By :Yogesh Patel 05/17/2022

Quant Time: May 12 03:27:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 16:40:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	195603	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	855965	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	539965	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1017224	20.000	ng/u1	0.00
79) Chrysene-d12	21.468	240	738702	20.000	ng/u1	0.00
88) Perylene-d12	23.850	264	721616	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	18691	3.998	ng/uL	0.00
4) Pyridine-d5	3.775	84	228655	17.743	ng/u1	0.00
7) Phenol-d5	7.081	99	343301	20.288	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.245	67	238776	22.481	ng/u1	0.00
11) 2-Chlorophenol-d4	7.451	132	300280	22.582	ng/u1	0.00
15) 4-Methylphenol-d8	8.622	113	303916	22.029	ng/u1	0.00
21) Nitrobenzene-d5	9.075	128	146621	22.349	ng/u1	0.00
24) 2-Nitrophenol-d4	9.798	143	158232	22.362	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.339	165	282662	21.345	ng/u1	0.00
31) 4-Chloroaniline-d4	10.851	131	294046	15.595	ng/u1	0.00
46) Dimethylphthalate-d6	13.963	166	1057627	25.972	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	1201222	23.247	ng/u1	0.00
54) 4-Nitrophenol-d4	14.733	143	89768m	12.901	ng/u1	0.00
60) Fluorene-d10	15.545	176	879092	25.588	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	78532	12.251	ng/u1	0.00
73) Anthracene-d10	17.392	188	1244434	25.145	ng/u1	0.00
81) Pyrene-d10	19.680	212	1211428	26.263	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.697	264	993052	26.331	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051122\
Data File : BM034985.D
Acq On : 11 May 2022 15:06
Operator : CG/JU
Sample : N2603-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW484

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/12/2022
Supervised By :Yogesh Patel 05/17/2022

Quant Time: May 12 03:27:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM051022.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 10 16:40:09 2022
Response via : Initial Calibration

