

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051122\
 Data File : BM035013.D
 Acq On : 12 May 2022 08:32
 Operator : CG/JU
 Sample : PB144679BS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS679

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/16/2022
 Supervised By :Yogesh Patel 05/17/2022

Quant Time: May 13 04:38:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 16:40:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	200170	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	915076	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	594620	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1192706	20.000	ng/u1	0.00
79) Chrysene-d12	21.468	240	903563	20.000	ng/u1	0.00
88) Perylene-d12	23.850	264	852819	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	27401	5.727	ng/uL	0.00
4) Pyridine-d5	3.775	84	399761	30.312	ng/u1	0.00
7) Phenol-d5	7.081	99	550126	31.768	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.245	67	336321	30.943	ng/u1	0.00
11) 2-Chlorophenol-d4	7.451	132	462115	33.959	ng/u1	0.00
15) 4-Methylphenol-d8	8.628	113	480346	34.023	ng/u1	0.00
21) Nitrobenzene-d5	9.075	128	233774	33.332	ng/u1	0.00
24) 2-Nitrophenol-d4	9.798	143	263345	34.812	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.339	165	473793	33.467	ng/u1	0.00
31) 4-Chloroaniline-d4	10.857	131	628710	31.190	ng/u1	0.00
46) Dimethylphthalate-d6	13.963	166	1485056	33.116	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	1760884	30.946	ng/u1	0.00
54) 4-Nitrophenol-d4	14.745	143	279438m	36.469	ng/u1	0.00
60) Fluorene-d10	15.539	176	1244182	32.886	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	248407	33.051	ng/u1	0.00
73) Anthracene-d10	17.392	188	1861633	32.081	ng/u1	0.00
81) Pyrene-d10	19.680	212	1882705	33.369	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.703	264	1524471	34.203	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	58645	12.310	ng/uL	97
5) Pyridine	3.793	79	443954	31.896	ng/u1	99
6) Benzaldehyde	7.057	77	362161	37.474	ng/u1	99
8) Phenol	7.110	94	604121	35.047	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.340	93	459576	33.684	ng/u1	98
12) 2-Chlorophenol	7.481	128	495654	35.715	ng/u1	98
13) 2-Methylphenol	8.363	108	468415	35.852	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.451	45	697813	31.283	ng/u1	100
16) Acetophenone	8.740	105	760181	34.688	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.734	70	414164	35.433	ng/u1	96
18) 4-Methylphenol	8.692	108	509318	35.680	ng/u1	98
19) Hexachloroethane	9.004	117	201302	34.193	ng/u1	96
22) Nitrobenzene	9.122	77	589034	33.845	ng/u1	99
23) Isophorone	9.651	82	1126091	35.215	ng/u1	98
25) 2-Nitrophenol	9.834	139	299684	36.870	ng/u1	93
26) 2,4-Dimethylphenol	9.898	107	573953	33.764	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.134	93	643502	32.464	ng/u1	99
29) 2,4-Dichlorophenol	10.369	162	494399	35.896	ng/u1	98
30) Naphthalene	10.775	128	1593538	33.147	ng/u1	100
32) 4-Chloroaniline	10.881	127	661610	32.797	ng/u1	99
33) Hexachlorobutadiene	11.069	225	299887	33.221	ng/u1	98
34) Caprolactam	11.651	113	183214m	44.487	ng/u1	
35) 4-Chloro-3-methylphenol	12.004	107	555653	37.530	ng/u1	98
36) 2-Methylnaphthalene	12.380	142	1136035	34.177	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051122\
 Data File : BM035013.D
 Acq On : 12 May 2022 08:32
 Operator : CG/JU
 Sample : PB144679BS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS679

Quant Time: May 13 04:38:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 16:40:09 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 05/16/2022
 Supervised By :Yogesh Patel 05/17/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	1167863	34.746	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.751	216	576906	32.988	ng/ul	99
40) Hexachlorocyclopentadiene	12.733	237	324962	26.792	ng/ul	97
41) 2,4,6-Trichlorophenol	12.986	196	409973	36.823	ng/ul	97
42) 2,4,5-Trichlorophenol	13.057	196	440588	36.939	ng/ul	99
43) 1,1'-Biphenyl	13.392	154	1549765	32.806	ng/ul	99
44) 2-Chloronaphthalene	13.427	162	1183571	32.899	ng/ul	98
45) 2-Nitroaniline	13.627	65	396470	38.819	ng/ul	93
47) Dimethylphthalate	14.010	163	1558492	35.175	ng/ul	99
48) 2,6-Dinitrotoluene	14.127	165	334134	39.706	ng/ul	94
50) Acenaphthylene	14.274	152	1960184	34.289	ng/ul	98
51) 3-Nitroaniline	14.451	138	326597	42.384	ng/ul#	92
52) Acenaphthene	14.616	153	1291111	33.791	ng/ul	99
53) 2,4-Dinitrophenol	14.657	184	190643	38.835	ng/ul	93
55) 4-Nitrophenol	14.757	109	265690	39.432	ng/ul	94
56) Dibenzofuran	14.951	168	1784513	33.911	ng/ul	95
57) 2,4-Dinitrotoluene	14.910	165	476647	40.155	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.174	232	362500	38.762	ng/ul	97
59) Diethylphthalate	15.374	149	1616768	35.626	ng/ul	99
61) Fluorene	15.598	166	1501876	34.888	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.592	204	726369	34.738	ng/ul	99
63) 4-Nitroaniline	15.610	138	344055	50.165	ng/ul#	84
66) 4,6-Dinitro-2-methylph...	15.674	198	260710	35.034	ng/ul#	96
67) N-Nitrosodiphenylamine	15.804	169	1285935	33.489	ng/ul	99
68) 4-Bromophenyl-phenylether	16.486	248	424506	34.227	ng/ul	99
69) Hexachlorobenzene	16.604	284	479476	34.166	ng/ul	97
70) Atrazine	16.757	200	480041	35.170	ng/ul	98
71) Pentachlorophenol	16.945	266	303638	34.856	ng/ul	97
72) Phenanthrene	17.333	178	2271450	33.858	ng/ul	99
74) Anthracene	17.427	178	2322866	34.234	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.357	216	587786	30.797	ng/ul	98
76) Pentachlorobenzene	14.874	250	557897	31.911	ng/ul	100
77) Carbazole	17.692	167	2080151	34.871	ng/ul	99
78) Di-n-butylphthalate	18.262	149	2591022	35.026	ng/ul	99
80) Fluoranthene	19.345	202	2446352	36.068	ng/ul	97
82) Pyrene	19.709	202	2442219	35.260	ng/ul	100
83) Butylbenzylphthalate	20.609	149	1040084	37.537	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.386	252	652006	38.240	ng/ul	99
85) Benzo(a)anthracene	21.456	228	2087526	35.559	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.392	149	1417100	34.515	ng/ul	100
87) Chrysene	21.503	228	1997681	34.944	ng/ul	98
89) Di-n-octyl phthalate	22.309	149	2174828	30.633	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	2002357	34.954	ng/ul	99
91) Benzo(k)fluoranthene	23.180	252	1930339	35.388	ng/ul	100
93) Benzo(a)pyrene	23.750	252	1971577	36.102	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.327	276	2283878	36.558	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.338	278	1942522	36.269	ng/ul#	96
96) Benzo(g,h,i)perylene	27.085	276	1909207	36.792	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051122\
 Data File : BM035013.D
 Acq On : 12 May 2022 08:32
 Operator : CG/JU
 Sample : PB144679BS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SLCS679

Quant Time: May 13 04:38:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 16:40:09 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
 Supervised By :Yogesh Patel 05/17/2022





#2
1,4-Dioxane
Concen: 12.310 ng/uL
RT: 3.399 min Scan# 51
Delta R.T. -0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion: 88 Resp: 5864
Ion Ratio Lower Upper
88 100
43 38.3 27.6 41.4
58 74.9 59.5 89.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#5
Pyridine
Concen: 31.896 ng/uL
RT: 3.793 min Scan# 120
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion: 79 Resp: 443954
Ion Ratio Lower Upper
79 100
52 79.9 64.3 96.5
51 41.0 32.4 48.6





#6
Benzaldehyde
Concen: 37.474 ng/ul
RT: 7.057 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion: 77 Resp: 36216
Ion Ratio Lower Upper
77 100
105 95.9 77.4 116.0
106 89.9 73.4 110.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#8
Phenol
Concen: 35.047 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion: 94 Resp: 604121
Ion Ratio Lower Upper
94 100
65 32.3 23.8 35.6
66 43.5 31.3 46.9





#10
Bis(2-Chloroethyl)ether
Concen: 33.684 ng/ul
RT: 7.340 min Scan# 723
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion: 93 Resp: 459570
Ion Ratio Lower Upper
93 100
63 80.0 62.5 93.7
95 32.5 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#12
2-Chlorophenol
Concen: 35.715 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:128 Resp: 495654
Ion Ratio Lower Upper
128 100
64 48.3 37.1 55.7
130 32.4 26.2 39.4





#13
2-Methylphenol
Concen: 35.852 ng/ul
RT: 8.363 min Scan# 897
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32



Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:108 Resp: 468415
Ion Ratio Lower Upper
108 100
107 89.8 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.283 ng/ul
RT: 8.451 min Scan# 912
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32



Tgt Ion: 45 Resp: 697813
Ion Ratio Lower Upper
45 100
77 13.0 10.6 15.8
79 10.1 8.1 12.1









#16
Acetophenone
Concen: 34.688 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32

Instrument :
BNA_M
ClientSampleId :
SLCS679

Tgt Ion: 105 Resp: 76018
Ion Ratio Lower Upper
105 100
77 79.2 62.4 93.6
51 32.4 23.8 35.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#17
N-Nitroso-di-n-propylamine
Concen: 35.433 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion: 70 Resp: 414164
Ion Ratio Lower Upper
70 100
42 69.9 53.3 79.9
101 9.8 8.2 12.4
130 21.1 18.3 27.5





#18
4-Methylphenol
Concen: 35.680 ng/ul
RT: 8.692 min Scan# 953
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:108 Resp: 509318
Ion Ratio Lower Upper
108 100
107 114.7 89.7 134.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#19
Hexachloroethane
Concen: 34.193 ng/ul
RT: 9.004 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:117 Resp: 201302
Ion Ratio Lower Upper
117 100
201 97.1 81.3 121.9
199 63.4 53.4 80.0





#22
Nitrobenzene
Concen: 33.845 ng/u1
RT: 9.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion: 77 Resp: 589034
Ion Ratio Lower Upper
77 100
123 44.3 36.1 54.1
65 13.8 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#23
Isophorone
Concen: 35.215 ng/u1
RT: 9.651 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion: 82 Resp: 1126091
Ion Ratio Lower Upper
82 100
95 7.9 6.2 9.4
138 17.8 15.3 22.9





#25
2-Nitrophenol
Concen: 36.870 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:139 Resp: 299684
Ion Ratio Lower Upper
139 100
65 53.2 38.9 58.3
109 33.5 24.1 36.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#26
2,4-Dimethylphenol
Concen: 33.764 ng/ul
RT: 9.898 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:107 Resp: 573953
Ion Ratio Lower Upper
107 100
121 49.8 40.3 60.5
122 83.3 66.6 100.0





#27
Bis(2-Chloroethoxy)methane
Concen: 32.464 ng/ul
RT: 10.134 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion: 93 Resp: 643502
Ion Ratio Lower Upper
93 100
95 32.7 25.8 38.6
123 11.7 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#29
2,4-Dichlorophenol
Concen: 35.896 ng/ul
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:162 Resp: 494399
Ion Ratio Lower Upper
162 100
164 66.1 53.0 79.4
98 37.3 27.4 41.0





#30
Naphthalene
Concen: 33.147 ng/u1
RT: 10.775 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:128 Resp: 1593538
Ion Ratio Lower Upper
128 100
129 11.1 9.1 13.7
127 13.2 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#32
4-Chloroaniline
Concen: 32.797 ng/u1
RT: 10.881 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:127 Resp: 661610
Ion Ratio Lower Upper
127 100
129 32.5 25.4 38.2





#33
Hexachlorobutadiene
Concen: 33.221 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:225 Resp: 29988
Ion Ratio Lower Upper
225 100
223 61.4 49.8 74.8
227 64.7 49.6 74.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#34
Caprolactam
Concen: 44.487 ng/ul m
RT: 11.651 min Scan# 1456
Delta R.T. 0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:113 Resp: 183214
Ion Ratio Lower Upper
113 100
55 174.9 130.6 195.8
56 136.8 104.3 156.5





#35
4-Chloro-3-methylphenol
Concen: 37.530 ng/ul
RT: 12.004 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:107 Resp: 55565
Ion Ratio Lower Upper
107 100
144 26.2 21.5 32.3
142 82.0 66.8 100.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#36
2-Methylnaphthalene
Concen: 34.177 ng/ul
RT: 12.380 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:142 Resp: 1136035
Ion Ratio Lower Upper
142 100
141 90.1 72.2 108.2





#37
1-Methylnaphthalene
Concen: 34.746 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:142 Resp: 116786
Ion Ratio Lower Upper
142 100
141 91.8 73.0 109.4
116 4.3 3.2 4.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.988 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:216 Resp: 576906
Ion Ratio Lower Upper
216 100
214 79.6 64.5 96.7
179 23.4 17.9 26.9
108 18.9 15.0 22.4





#40
Hexachlorocyclopentadiene
Concen: 26.792 ng/ul
RT: 12.733 min Scan# 1640
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:237 Resp: 324962
Ion Ratio Lower Upper
237 100
235 63.8 49.2 73.8
272 12.9 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#41
2,4,6-Trichlorophenol
Concen: 36.823 ng/ul
RT: 12.986 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:196 Resp: 409973
Ion Ratio Lower Upper
196 100
198 95.5 78.5 117.7
200 29.8 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 36.939 ng/u1
RT: 13.057 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:196 Resp: 44058
Ion Ratio Lower Upper
196 100
198 97.7 77.7 116.5
200 31.7 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#43
1,1'-Biphenyl
Concen: 32.806 ng/u1
RT: 13.392 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:154 Resp: 1549765
Ion Ratio Lower Upper
154 100
153 40.4 32.2 48.2
76 11.5 8.6 12.8





#44
2-Chloronaphthalene
Concen: 32.899 ng/ul
RT: 13.427 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:162 Resp: 118357
Ion Ratio Lower Upper
162 100
164 32.9 25.7 38.5
127 39.3 29.8 44.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#45
2-Nitroaniline
Concen: 38.819 ng/ul
RT: 13.627 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion: 65 Resp: 396470
Ion Ratio Lower Upper
65 100
92 65.0 55.9 83.9
138 100.5 87.3 130.9





#47
Dimethylphthalate
Concen: 35.175 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:163 Resp: 155849
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.0
164 10.3 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#48
2,6-Dinitrotoluene
Concen: 39.706 ng/ul
RT: 14.127 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:165 Resp: 334134
Ion Ratio Lower Upper
165 100
89 59.2 43.1 64.7
121 21.5 16.6 24.8





#50
Acenaphthylene
Concen: 34.289 ng/ul
RT: 14.274 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:152 Resp: 1960184
Ion Ratio Lower Upper
152 100
151 20.0 15.3 22.9
153 13.2 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#51
3-Nitroaniline
Concen: 42.384 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:138 Resp: 326597
Ion Ratio Lower Upper
138 100
108 10.1 10.2 15.4#
92 115.2 85.4 128.0





#52
Acenaphthene
Concen: 33.791 ng/u1
RT: 14.616 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:153 Resp: 129111
Ion Ratio Lower Upper
153 100
152 47.0 37.5 56.3
154 87.3 71.0 106.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#53
2,4-Dinitrophenol
Concen: 38.835 ng/u1
RT: 14.657 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:184 Resp: 190643
Ion Ratio Lower Upper
184 100
63 73.1 51.8 77.6
154 70.1 54.3 81.5





#55
4-Nitrophenol
Concen: 39.432 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:109 Resp: 265690
Ion Ratio Lower Upper
109 100
139 118.2 104.6 156.8
65 126.8 100.6 150.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#56
Dibenzofuran
Concen: 33.911 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:168 Resp: 1784513
Ion Ratio Lower Upper
168 100
139 40.2 30.0 45.0





#57
2,4-Dinitrotoluene
Concen: 40.155 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:165 Resp: 47664
Ion Ratio Lower Upper
165 100
63 48.9 34.2 51.4
182 2.8 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 38.762 ng/ul
RT: 15.174 min Scan# 2055
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:232 Resp: 362500
Ion Ratio Lower Upper
232 100
131 53.0 39.6 59.4
130 2.8 2.1 3.1
166 34.8 27.7 41.5





#59
Diethylphthalate
Concen: 35.626 ng/ul
RT: 15.374 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:149 Resp: 161676
Ion Ratio Lower Upper
149 100
177 20.9 17.2 25.8
150 12.7 9.8 14.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#61
Fluorene
Concen: 34.888 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:166 Resp: 1501876
Ion Ratio Lower Upper
166 100
165 98.9 77.9 116.9
167 13.8 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 34.738 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:204 Resp: 726369
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.4
141 62.7 49.5 74.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#63
4-Nitroaniline
Concen: 50.165 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:138 Resp: 344055
Ion Ratio Lower Upper
138 100
92 51.8 37.2 55.8
108 90.7 58.4 87.6#





#66
4,6-Dinitro-2-methylphenol
Concen: 35.034 ng/ul
RT: 15.674 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:198 Resp: 260710
Ion Ratio Lower Upper
198 100
182 4.2 4.6 6.8
77 29.6 22.2 33.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#67
N-Nitrosodiphenylamine
Concen: 33.489 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32



Tgt Ion:169 Resp: 1285935
Ion Ratio Lower Upper
169 100
168 68.4 54.8 82.2
167 37.0 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 34.227 ng/ul
RT: 16.486 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32

Instrument :
BNA_M
ClientSampleId :
SLCS679

Tgt Ion:248 Resp: 424500
Ion Ratio Lower Upper
248 100
250 97.3 77.9 116.9
141 77.5 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#69
Hexachlorobenzene
Concen: 34.166 ng/ul
RT: 16.604 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:284 Resp: 479476
Ion Ratio Lower Upper
284 100
142 42.3 31.2 46.8
249 32.2 25.9 38.9





#70
Atrazine
Concen: 35.170 ng/u1
RT: 16.757 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:200 Resp: 48004
Ion Ratio Lower Upper
200 100
173 25.8 21.3 31.9
215 45.1 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#71
Pentachlorophenol
Concen: 34.856 ng/u1
RT: 16.945 min Scan# 2356
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:266 Resp: 303638
Ion Ratio Lower Upper
266 100
264 62.8 48.4 72.6
268 65.6 50.6 76.0





#72
Phenanthrene
Concen: 33.858 ng/u1
RT: 17.333 min Scan# 2422
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:178 Resp: 2271450
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 19.3 15.0 22.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#74
Anthracene
Concen: 34.234 ng/u1
RT: 17.427 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:178 Resp: 2322866
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 19.1 14.6 22.0





#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.797 ng/uL
RT: 13.357 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679




Tgt Ion:216 Resp: 587780
Ion Ratio Lower Upper
216 100
214 77.6 62.6 94.0
218 48.4 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022




#76
Pentachlorobenzene
Concen: 31.911 ng/uL
RT: 14.874 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Tgt Ion:250 Resp: 557897
Ion Ratio Lower Upper
250 100
248 62.5 49.9 74.9
252 64.7 51.5 77.3





#77
 Carbazole
 Concen: 34.871 ng/ul
 RT: 17.692 min Scan# 2483
 Delta R.T. 0.000 min
 Lab File: BM035013.D
 Acq: 12 May 2022 08:32




Instrument :
 BNA_M
 ClientSampleId :
 SLCS679




Tgt Ion:167 Resp: 208015
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.1 25.7
 139 13.3 10.5 15.7

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/16/2022
 Supervised By :Yogesh Patel 05/17/2022




#78
 Di-n-butylphthalate
 Concen: 35.026 ng/ul
 RT: 18.262 min Scan# 2580
 Delta R.T. -0.006 min
 Lab File: BM035013.D
 Acq: 12 May 2022 08:32




Tgt Ion:149 Resp: 2591022
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.6 4.1 6.1





#80
Fluoranthene
Concen: 36.068 ng/u1
RT: 19.345 min Scan# 2764
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:202 Resp: 2446352
Ion Ratio Lower Upper
202 100
101 10.5 7.4 11.2
100 8.1 5.8 8.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#82
Pyrene
Concen: 35.260 ng/u1
RT: 19.709 min Scan# 2826
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:202 Resp: 2442219
Ion Ratio Lower Upper
202 100
101 11.7 9.4 14.0
100 9.7 7.8 11.6





#83
Butylbenzylphthalate
Concen: 37.537 ng/ul
RT: 20.609 min Scan# 2979
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:149 Resp: 1040084
Ion Ratio Lower Upper
149 100
91 67.8 54.3 81.5
206 19.3 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#84
3,3'-Dichlorobenzidine
Concen: 38.240 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:252 Resp: 652006
Ion Ratio Lower Upper
252 100
254 64.2 51.0 76.4
126 10.8 9.0 13.6





#85
Benzo(a)anthracene
Concen: 35.559 ng/ul
RT: 21.456 min Scan# 3122
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:228 Resp: 2087520
Ion Ratio Lower Upper
228 100
229 19.6 15.4 23.0
226 27.6 21.7 32.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.515 ng/ul
RT: 21.392 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:149 Resp: 1417100
Ion Ratio Lower Upper
149 100
167 26.6 21.4 32.0
279 4.0 3.4 5.2





#87
Chrysene
Concen: 34.944 ng/ul
RT: 21.503 min Scan# 3131
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:228 Resp: 1997681
Ion Ratio Lower Upper
228 100
226 30.8 23.7 35.5
229 19.9 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#89
Di-n-octyl phthalate
Concen: 30.633 ng/ul
RT: 22.309 min Scan# 3268
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:149 Resp: 2174828





#90
Benzo(b)fluoranthene
Concen: 34.954 ng/u1
RT: 23.127 min Scan# 3407
Delta R.T. -0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:252 Resp: 200235
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 9.8 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#91
Benzo(k)fluoranthene
Concen: 35.388 ng/u1
RT: 23.180 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:252 Resp: 1930339
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 9.5 7.8 11.6





#93
Benzo(a)pyrene
Concen: 36.102 ng/ul
RT: 23.750 min Scan# 3513
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32



Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:252 Resp: 1971577
Ion Ratio Lower Upper
252 100
253 21.8 16.6 25.0
125 10.8 6.2 9.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 36.558 ng/ul
RT: 26.327 min Scan# 3951
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32


Tgt Ion:276 Resp: 2283878
Ion Ratio Lower Upper
276 100
138 23.9 15.3 22.9
277 24.7 19.6 29.4





#95
Dibenzo(a,h)anthracene
Concen: 36.269 ng/ul
RT: 26.338 min Scan# 3953
Delta R.T. 0.000 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32




Instrument :
BNA_M
ClientSampleId :
SLCS679



Tgt Ion:278 Resp: 194252
Ion Ratio Lower Upper
278 100
139 15.9 9.7 14.5
279 23.9 18.8 28.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/16/2022
Supervised By :Yogesh Patel 05/17/2022



#96
Benzo(g,h,i)perylene
Concen: 36.792 ng/ul
RT: 27.085 min Scan# 4080
Delta R.T. 0.006 min
Lab File: BM035013.D
Acq: 12 May 2022 08:32

Tgt Ion:276 Resp: 1909207
Ion Ratio Lower Upper
276 100
138 22.0 13.0 19.6
277 23.3 18.6 28.0

