

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051216\
 Data File : BM005411.D
 Acq On : 12 May 2016 09:36
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02064

Quant Time: May 12 10:06:38 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 12 02:20:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
2	1,4-Dioxane	0.776	0.771	0.6	59	0.00
3 S	1,4-Dioxane-d8	0.425	0.425	0.0	62	0.00
4	Benzaldehyde	0.879	1.109	-26.2#	61	0.00
5 S	Phenol-d5	1.814	1.832	-1.0	63	0.00
6	Phenol	1.875	1.932	-3.0	63	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.035	1.048	-1.3	61	0.00
8	Bis(2-Chloroethyl)ether	1.411	1.427	-1.1	61	0.00
9 S	2-Chlorophenol-d4	1.370	1.411	-3.0	63	0.00
10	2-Chlorophenol	1.401	1.433	-2.3	62	0.00
11	2-Methylphenol	1.449	1.462	-0.9	63	0.00
12	2,2'-oxybis(1-Chloropropane	1.915	1.988	-3.8	63	0.00
13 S	4-Methylphenol-d8	1.499	1.529	-2.0	64	0.00
14	Acetophenone	2.221	2.430	-9.4	64	0.00
15 P	N-Nitroso-di-n-propylamine	1.161	1.244	-7.1	65	0.00
16	4-Methylphenol	1.608	1.662	-3.4	63	0.00
17	Hexachloroethane	0.527	0.547	-3.8	65	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
19 S	Nitrobenzene-d5	0.143	0.147	-2.8	64	0.00
20	Nitrobenzene	0.357	0.363	-1.7	63	0.00
21	Isophorone	0.694	0.718	-3.5	64	0.00
22 S	2-Nitrophenol-d4	0.162	0.169	-4.3	64	0.00
23 C	2-Nitrophenol	0.174	0.184	-5.7	65	0.00
24	2,4-Dimethylphenol	0.370	0.391	-5.7	66	0.00
25	Bis(2-Chloroethoxy)methane	0.432	0.432	0.0	63	0.00
26 S	2,4-Dichlorophenol-d3	0.300	0.327	-9.0	67	0.00
27 C	2,4-Dichlorophenol	0.308	0.330	-7.1	66	0.00
28	Naphthalene	1.006	1.028	-2.2	64	0.00
29 S	4-Chloroaniline-d4	0.362	0.419	-15.7	63	0.00
30	4-Chloroaniline	0.369	0.427	-15.7	64	0.00
31 C	Hexachlorobutadiene	0.183	0.196	-7.1	68	0.00
32	Caprolactam	0.115	0.113	1.7	64	0.00
33 C	4-Chloro-3-methylphenol	0.369	0.406	-10.0	68	0.00
34	2-Methylnaphthalene	0.753	0.788	-4.6	65	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
36	1,2,4,5-Tetrachlorobenzene	0.608	0.646	-6.3	67	0.00
37	Hexachlorocyclopentadiene	0.311	0.218	29.9#	48	0.00
38 C	2,4,6-Trichlorophenol	0.405	0.434	-7.2	68	0.00
39	2,4,5-Trichlorophenol	0.450	0.491	-9.1	70	0.00
40	1,1'-Biphenyl	1.584	1.638	-3.4	66	0.00
41	2-Chloronaphthalene	1.198	1.237	-3.3	66	0.00
42	2-Nitroaniline	0.369	0.396	-7.3	67	0.00
43 S	Dimethylphthalate-d6	1.603	1.661	-3.6	65	0.00
44	Dimethylphthalate	1.602	1.656	-3.4	65	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051216\
 Data File : BM005411.D
 Acq On : 12 May 2016 09:36
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02064

Quant Time: May 12 10:06:38 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 12 02:20:05 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.316	0.343	-8.5	67	0.00
46 S	Acenaphthylene-d8	1.880	1.980	-5.3	66	0.00
47	Acenaphthylene	1.989	2.077	-4.4	66	0.00
48	3-Nitroaniline	0.337	0.356	-5.6	63	0.00
49 C	Acenaphthene	1.311	1.355	-3.4	66	0.00
50	2,4-Dinitrophenol	0.182	0.117	35.7#	50	0.00
51 S	4-Nitrophenol-d4	0.293	0.253	13.7	55	0.00
52	4-Nitrophenol	0.243	0.222	8.6	59	0.00
53	Dibenzofuran	1.902	1.977	-3.9	66	0.00
54	2,4-Dinitrotoluene	0.471	0.505	-7.2	67	0.00
55	2,3,4,6-Tetrachlorophenol	0.382	0.410	-7.3	68	0.00
56	Diethylphthalate	1.618	1.696	-4.8	66	0.00
57 S	Fluorene-d10	1.384	1.451	-4.8	67	0.00
58	Fluorene	1.487	1.583	-6.5	68	0.00
59	4-Chlorophenyl-phenylether	0.737	0.794	-7.7	69	0.00
60	4-Nitroaniline	0.357	0.365	-2.2	64	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	64	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.113	0.098	13.3	57	0.00
63	4,6-Dinitro-2-methylphenol	0.118	0.101	14.4	55	0.00
64	N-Nitrosodiphenylamine	0.548	0.589	-7.5	67	0.00
65	4-Bromophenyl-phenylether	0.192	0.213	-10.9	69	0.00
66	Hexachlorobenzene	0.215	0.237	-10.2	68	0.00
67	Atrazine	0.200	0.229	-14.5	68	0.00
68 C	Pentachlorophenol	0.120	0.108	10.0	59	0.00
69	Phenanthrene	1.052	1.108	-5.3	65	0.00
70 S	Anthracene-d10	0.884	0.938	-6.1	66	0.00
71	Anthracene	1.048	1.121	-7.0	67	0.00
72	Carbazole	0.922	1.006	-9.1	63	0.00
73	Di-n-butylphthalate	1.125	1.242	-10.4	67	0.00
74 C	Fluoranthene	1.165	1.289	-10.6	62	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
76 S	Pyrene-d10	0.923	1.137	-23.2#	61	0.00
77	Pyrene	1.160	1.429	-23.2#	61	0.00
78	Butylbenzylphthalate	0.482	0.603	-25.1#	61	0.00
79	3,3'-Dichlorobenzidine	0.360	0.354	1.7	45	0.00
80	Benzo(a)anthracene	1.150	1.206	-4.9	52	0.00
81	Bis(2-ethylhexyl)phthalate	0.669	0.818	-22.3#	57	0.00
82	Chrysene	1.091	1.131	-3.7	51	0.00
83 I	Perylene-d12	1.000	1.000	0.0	39	0.00
84	Di-n-octyl phthalate	1.168	1.591	-36.2#	49	0.00
85	Benzo(b)fluoranthene	1.169	1.258	-7.6	42	0.00
86	Benzo(k)fluoranthene	1.101	1.220	-10.8	41	0.00
87 S	Benzo(a)pyrene-d12	0.885	0.937	-5.9	40	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051216\
Data File : BM005411.D
Acq On : 12 May 2016 09:36
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02064

Quant Time: May 12 10:06:38 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM050516.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 12 02:20:05 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.106	1.177	-6.4	40	0.00
89	Indeno(1,2,3-cd)pyrene	1.206	1.257	-4.2	39	0.00
90	Dibenzo(a,h)anthracene	1.010	1.065	-5.4	39	0.00
91	Benzo(g,h,i)perylene	1.022	1.055	-3.2	39	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0