


Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM051223\  
 Data File : BM039912.D  
 Acq On : 12 May 2023 10:59  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM051223.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Sat May 13 02:16:09 2023



AutoFind: Scans 2518, 2519, 2520; Background Corrected with Scan 2512

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 35.4         | 159729     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 3442       | PASS                |
| 69             | 198             | 0.00            | 100             | 46.0         | 207573     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 1092       | PASS                |
| 127            | 198             | 10              | 80              | 57.5         | 259136     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.3          | 1341       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 450816     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 30368      | PASS                |
| 275            | 198             | 10              | 60              | 26.6         | 119824     | PASS                |
| 365            | 198             | 1               | 100             | 4.0          | 18232      | PASS                |
| 441            | 198             | 0.01            | 100             | 13.5         | 60995      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 369600     | PASS                |
| 443            | 442             | 15              | 24              | 19.4         | 71541      | PASS                |

# DDT Breakdown




| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 5/12/2023     | BNA_M            | <u>BM039912.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 677345           | 20.883             |
| DDD           | 3765             | 20.495             |
| DDE           | 127              | 19.93              |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 3892          | 681237           | 0.57               |
|               |                  |                    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM051223\  
Data File : BM039912.D  
Acq On : 12 May 2023 10:59  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP

Quant Time: May 13 04:18:11 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM051223.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat May 13 02:16:09 2023  
Response via : Initial Calibration



TIC: BM039912.D\data.ms

(70) Pentachlorophenol (C)

17.054min (-0.006) 0.00 ng




response 443911

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70  | 64.14  |
| 264.00 | 63.70  | 62.78  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM051223\  
 Data File : BM039912.D  
 Acq On : 12 May 2023 10:59  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Quant Time: May 13 04:18:11 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM051223.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat May 13 02:16:09 2023  
 Response via : Initial Calibration



TIC: BM039912.D\data.ms

**(77) Benzidine**

**19.648min (-0.000) 1072219.19 ng**

**response 2307146**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.10  | 14.29  |
| 183.00 | 11.90  | 11.48  |
| 0.00   | 0.00   | 0.00   |