

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051315\
 Data File : BM001267.D
 Acq On : 13 May 2015 00:54
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM051315

Quant Time: May 13 05:49:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.582	0.659	-13.2	98	0.00
3	n-Nitrosodimethylamine	0.566	0.568	-0.4	107	0.00
4 S	2-Fluorophenol	0.974	0.923	5.2	93	0.00
5 S	Phenol-d6	1.099	1.007	8.4	92	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
7 S	Nitrobenzene-d5	0.263	0.242	8.0	90	0.00
8	Nitrobenzene	0.276	0.257	6.9	93	0.00
9	Naphthalene	1.064	1.032	3.0	92	0.00
10	Hexachlorobutadiene	0.208	0.198	4.8	93	0.00
11	2-Methylnaphthalene	0.692	0.652	5.8	92	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
13 S	2,4,6-Tribromophenol	0.149	0.110	26.2#	80	0.00
14 S	2-Fluorobiphenyl	1.552	1.502	3.2	93	0.00
15	Acenaphthylene	2.058	1.907	7.3	91	0.00
16	Acenaphthene	1.318	1.247	5.4	92	0.00
17	Fluorene	1.194	1.025	14.2	88	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
19	Hexachlorobenzene	0.322	0.272	15.5	96	0.00
20	Pentachlorophenol	0.067	0.042	37.3#	93	0.00
21	Phenanthrene	0.760	0.642	15.5	97	0.00
22	Anthracene	0.741	0.614	17.1	95	0.00
23	Fluoranthene	1.193	0.940	21.2	92	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	92	-0.01
25	Pyrene	1.545	1.429	7.5	93	0.00
26 S	Terphenyl-d14	0.672	0.622	7.4	91	-0.01
27	Benzo(a)anthracene	1.357	1.224	9.8	94	0.00
28	Chrysene	1.376	1.322	3.9	94	0.00
29	Indeno(1,2,3-cd)pyrene	1.398	1.366	2.3	96	0.00
30 I	Perylene-d12	1.000	1.000	0.0	93	-0.01
31	Benzo(b)fluoranthene	1.465	1.408	3.9	98	0.00
32	Benzo(k)fluoranthene	1.394	1.255	10.0	90	0.00
33 C	Benzo(a)pyrene	1.269	1.153	9.1	93	0.00
34	Dibenzo(a,h)anthracene	1.193	1.148	3.8	96	-0.01
35	Benzo(g,h,i)perylene	1.309	1.274	2.7	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0