

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM051315\
 Data File : BM001274.D
 Acq On : 13 May 2015 05:04
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 13 06:46:32 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIMPAH-BM051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 05:24:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.582	0.644	-10.7	97	0.00
3	n-Nitrosodimethylamine	0.566	0.500	11.7	96	0.00
4 S	2-Fluorophenol	0.974	0.923	5.2	95	0.00
5 S	Phenol-d6	1.099	1.029	6.4	96	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.263	0.247	6.1	94	0.00
8	Nitrobenzene	0.276	0.254	8.0	94	0.00
9	Naphthalene	1.064	1.035	2.7	95	0.00
10	Hexachlorobutadiene	0.208	0.198	4.8	95	0.00
11	2-Methylnaphthalene	0.692	0.655	5.3	95	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
13 S	2,4,6-Tribromophenol	0.149	0.114	23.5	86	0.00
14 S	2-Fluorobiphenyl	1.552	1.487	4.2	94	0.00
15	Acenaphthylene	2.058	1.919	6.8	93	0.00
16	Acenaphthene	1.318	1.244	5.6	94	0.00
17	Fluorene	1.194	0.993	16.8	87	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
19	Hexachlorobenzene	0.322	0.254	21.1	96	0.00
20	Pentachlorophenol	0.067	0.050	25.4#	117	0.00
21	Phenanthrene	0.760	0.622	18.2	100	0.00
22	Anthracene	0.741	0.610	17.7	101	0.00
23	Fluoranthene	1.193	0.897	24.8	94	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	97	-0.01
25	Pyrene	1.545	1.406	9.0	97	0.00
26 S	Terphenyl-d14	0.672	0.608	9.5	94	-0.01
27	Benzo(a)anthracene	1.357	1.206	11.1	98	0.00
28	Chrysene	1.376	1.302	5.4	98	-0.01
29	Indeno(1,2,3-cd)pyrene	1.398	1.387	0.8	103	0.00
30 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
31	Benzo(b)fluoranthene	1.465	1.421	3.0	106	0.00
32	Benzo(k)fluoranthene	1.394	1.194	14.3	91	0.00
33 C	Benzo(a)pyrene	1.269	1.153	9.1	100	0.00
34	Dibenzo(a,h)anthracene	1.193	1.154	3.3	103	-0.01
35	Benzo(g,h,i)perylene	1.309	1.279	2.3	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0