

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
 Data File : BM035063.D
 Acq On : 14 May 2022 13:41
 Operator : CG/JU
 Sample : PB144572BL
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK572

Quant Time: May 15 01:30:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 12 18:15:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.917	152	3865	0.400	ng/ul	0.00
4) Naphthalene-d8	10.715	136	12021	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.548	164	8883	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.293	188	19601	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.473	240	15513	0.400	ng/ul	0.00
23) Perylene-d12	23.855	264	13363	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.356	96	23721	5.123	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.305	152	6336	0.330	ng/ul	0.00
18) Fluoranthene-d10	19.318	212	16047	0.302	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
Data File : BM035063.D
Acq On : 14 May 2022 13:41
Operator : CG/JU
Sample : PB144572BL
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK572

Quant Time: May 15 01:30:59 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 12 18:15:39 2022
Response via : Initial Calibration

