

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
 Data File : BM035078.D
 Acq On : 14 May 2022 23:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4097

Quant Time: May 15 01:32:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 12 18:15:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.921	152	3115	0.400	ng/ul	0.00
4) Naphthalene-d8	10.715	136	9382	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.553	164	6116	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.293	188	13892	0.400	ng/ul #	0.00
17) Chrysene-d12	21.473	240	12678	0.400	ng/ul	0.00
23) Perylene-d12	23.861	264	11794	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.361	96	1359	0.364	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.305	152	6340	0.424	ng/ul	0.00
18) Fluoranthene-d10	19.318	212	16448	0.378	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.394	88	1264	0.337	ng/ul#	73
5) Naphthalene	10.765	128	11797	0.381	ng/ul#	89
7) 2-Methylnaphthalene	12.382	142	7980	0.392	ng/ul	100
8) 1-Methylnaphthalene	12.596	142	7841	0.414	ng/ul	99
10) Acenaphthylene	14.270	152	14162	0.391	ng/ul#	90
11) Acenaphthene	14.613	153	9952	0.381	ng/ul	96
12) Fluorene	15.598	166	11690	0.382	ng/ul#	97
14) Pentachlorophenol	16.947	266	2231	0.425	ng/ul	99
15) Phenanthrene	17.335	178	19884	0.376	ng/ul	94
16) Anthracene	17.428	178	18697	0.388	ng/ul#	91
19) Fluoranthene	19.350	202	24200	0.340	ng/ul	95
20) Pyrene	19.713	202	24975	0.345	ng/ul	98
21) Benzo(a)anthracene	21.458	228	22794	0.389	ng/ul	97
22) Chrysene	21.511	228	23113	0.379	ng/ul	98
24) Benzo(b)fluoranthene	23.136	252	22815	0.369	ng/ul	91
25) Benzo(k)fluoranthene	23.182	252	21911	0.371	ng/ul#	90
26) Benzo(a)pyrene	23.752	252	21165	0.376	ng/ul#	83
27) Indeno(1,2,3-cd)pyrene	26.328	276	22415	0.403	ng/ul#	84
28) Dibenzo(a,h)anthracene	26.352	278	18149	0.410	ng/ul#	90
29) Benzo(g,h,i)perylene	27.090	276	18820	0.393	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM051322\
Data File : BM035078.D
Acq On : 14 May 2022 23:18
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4097

Quant Time: May 15 01:32:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM051222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 12 18:15:39 2022
Response via : Initial Calibration

